

Report of Analysis

Client:	Coppola Services	Date Collected:	06/17/25
Project:	Millville Sewage Treatment Plant - Field Sampling	Date Received:	06/17/25
Client Sample ID:	PIBLK-PD088990.D	SDG No.:	Q2409
Lab Sample ID:	I.BLK-PD088990.D	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:		uL	
Extraction Type:		Test:	TCLP PesticideGroup1
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088990.D	1		06/17/25	PD061825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	18.6		30 (57) - 150 (171)	93%	SPK: 20
877-09-8	Tetrachloro-m-xylene	17.7		30 (61) - 150 (148)	88%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit