

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2409</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Coppola Services</u>	Datafile :	<u>PD089279.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168676BS (Column 1)	gamma-BHC (Lindane)	0.5	0.57	ug/L	113				40 (82)	140 (129)	
	Heptachlor	0.5	0.56	ug/L	111				40 (79)	140 (127)	
	Heptachlor epoxide	0.5	0.56	ug/L	111				40 (81)	140 (124)	
	Endrin	0.5	0.52	ug/L	104				40 (81)	140 (128)	
	Methoxychlor	0.5	0.51	ug/L	103				40 (78)	140 (108)	
PB168676BS (Column 2)	gamma-BHC (Lindane)	0.5	0.56	ug/L	112				40 (82)	140 (129)	
	Heptachlor	0.5	0.54	ug/L	108				40 (79)	140 (127)	
	Heptachlor epoxide	0.5	0.56	ug/L	113				40 (81)	140 (124)	
	Endrin	0.5	0.51	ug/L	103				40 (81)	140 (128)	
	Methoxychlor	0.5	0.53	ug/L	105				40 (78)	140 (108)	