

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: COPP02
 ProjectID: Millville Sewage Treatment Plant - Field Sampling
 Lab Code: CHEM Case No.: Q2409 SAS No.: Q2409 SDG No.: Q2409

Calibration Sequence : FF061325		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	202708062	119240	FF015964.D
850	99337935	116868	FF015965.D
340	42126327	123901	FF015966.D
170	20873459	122785	FF015967.D
85	11262958	132505	FF015968.D
AVG RF : 123060		% RSD : 4.857	AVG RT : 15.0944