

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061925\
 Data File : BF142793.D
 Acq On : 19 Jun 2025 19:40
 Operator : RC/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Quant Time: Jun 20 04:40:40 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 20 04:36:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	69874	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	262911	20.000	ng	0.00
39) Acenaphthene-d10	9.922	164	139156	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	229780	20.000	ng	0.00
76) Chrysene-d12	14.057	240	117825	20.000	ng	0.00
86) Perylene-d12	15.545	264	139593	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.492	112	441114	107.674	ng	0.00
7) Phenol-d6	6.510	99	516909	107.451	ng	0.00
23) Nitrobenzene-d5	7.451	82	506846	105.669	ng	0.00
42) 2,4,6-Tribromophenol	10.716	330	152358	100.743	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1043477	99.755	ng	0.00
79) Terphenyl-d14	12.998	244	904305	106.820	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.628	88	88835	54.606	ng	99
3) Pyridine	3.398	79	227576	55.874	ng	99
4) n-Nitrosodimethylamine	3.351	42	116636	55.913	ng	98
6) Aniline	6.545	93	346211	53.700	ng	99
8) 2-Chlorophenol	6.663	128	238691	53.507	ng	99
9) Benzaldehyde	6.434	77	145420	49.583	ng	99
10) Phenol	6.528	94	288461	53.781	ng	95
11) bis(2-Chloroethyl)ether	6.616	93	206661	51.933	ng	100
12) 1,3-Dichlorobenzene	6.822	146	262098	51.598	ng	99
13) 1,4-Dichlorobenzene	6.898	146	266098	51.685	ng	99
14) 1,2-Dichlorobenzene	7.051	146	253411	51.487	ng	99
15) Benzyl Alcohol	7.022	79	193284	53.677	ng	97
16) 2,2'-oxybis(1-Chloropr...	7.157	45	328251	52.354	ng	99
17) 2-Methylphenol	7.133	107	182167	53.415	ng	99
18) Hexachloroethane	7.392	117	94097	51.126	ng	99
19) n-Nitroso-di-n-propyla...	7.298	70	147905	48.753	ng	98
20) 3+4-Methylphenols	7.286	107	229169	53.390	ng	94
22) Acetophenone	7.292	105	303087	52.102	ng	98
24) Nitrobenzene	7.469	77	229219	53.773	ng	98
25) Isophorone	7.704	82	404886	50.334	ng	100
26) 2-Nitrophenol	7.781	139	127773	53.795	ng	99
27) 2,4-Dimethylphenol	7.816	122	216204	53.466	ng	99
28) bis(2-Chloroethoxy)met...	7.916	93	256119	51.148	ng	100
29) 2,4-Dichlorophenol	8.022	162	195589	52.478	ng	100
30) 1,2,4-Trichlorobenzene	8.104	180	213118	51.763	ng	100
31) Naphthalene	8.186	128	668713	51.717	ng	100
32) Benzoic acid	7.933	122	119196	53.128	ng	97
33) 4-Chloroaniline	8.239	127	276441	53.347	ng	99
34) Hexachlorobutadiene	8.304	225	131697	50.481	ng	99
35) Caprolactam	8.610	113	53024	52.873	ng	98
36) 4-Chloro-3-methylphenol	8.716	107	195744	50.843	ng	99
37) 2-Methylnaphthalene	8.880	142	414854	50.859	ng	99
38) 1-Methylnaphthalene	8.980	142	430144	50.792	ng	100
40) 1,2,4,5-Tetrachloroben...	9.045	216	207997	51.779	ng	100
41) Hexachlorocyclopentadiene	9.033	237	127033	50.012	ng	99
43) 2,4,6-Trichlorophenol	9.157	196	141932	54.597	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061925\
 Data File : BF142793.D
 Acq On : 19 Jun 2025 19:40
 Operator : RC/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Quant Time: Jun 20 04:40:40 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 20 04:36:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.198	196	144717	51.550	ng	100
46) 1,1'-Biphenyl	9.345	154	559765	51.807	ng	99
47) 2-Chloronaphthalene	9.369	162	420603	52.754	ng	100
48) 2-Nitroaniline	9.463	65	124478	54.955	ng	99
49) Acenaphthylene	9.786	152	694098	51.783	ng	100
50) Dimethylphthalate	9.645	163	464372	50.308	ng	100
51) 2,6-Dinitrotoluene	9.710	165	102361	51.156	ng	98
52) Acenaphthene	9.957	154	427135	51.498	ng	100
53) 3-Nitroaniline	9.874	138	117448	54.038	ng	98
54) 2,4-Dinitrophenol	9.986	184	60308	55.696	ng	92
55) Dibenzofuran	10.127	168	609507	51.506	ng	100
56) 4-Nitrophenol	10.033	139	82688	56.047	ng	98
57) 2,4-Dinitrotoluene	10.110	165	141691	53.554	ng	98
58) Fluorene	10.474	166	473968	50.747	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.245	232	120008	50.988	ng	97
60) Diethylphthalate	10.345	149	458546	50.010	ng	100
61) 4-Chlorophenyl-phenyle...	10.463	204	222613	49.076	ng	98
62) 4-Nitroaniline	10.492	138	109973	56.363	ng	97
63) Azobenzene	10.622	77	409164	51.063	ng	100
65) 4,6-Dinitro-2-methylph...	10.522	198	76696	53.713	ng	97
66) n-Nitrosodiphenylamine	10.580	169	411917	52.171	ng	99
67) 4-Bromophenyl-phenylether	10.957	248	135699	50.425	ng	98
68) Hexachlorobenzene	11.021	284	150669	50.432	ng	99
69) Atrazine	11.110	200	111915	53.385	ng	99
70) Pentachlorophenol	11.216	266	81577	52.574	ng	98
71) Phenanthrene	11.439	178	644668	52.403	ng	99
72) Anthracene	11.492	178	654942	51.433	ng	99
73) Carbazole	11.645	167	575784	53.996	ng	100
74) Di-n-butylphthalate	11.968	149	616452	52.211	ng	100
75) Fluoranthene	12.627	202	608190	52.032	ng	100
77) Benzidine	12.745	184	242794	58.208	ng	100
78) Pyrene	12.857	202	597913	55.205	ng	99
80) Butylbenzylphthalate	13.474	149	180565	56.257	ng	97
81) Benzo(a)anthracene	14.045	228	413349	53.431	ng	100
82) 3,3'-Dichlorobenzidine	14.010	252	140152	55.949	ng	99
83) Chrysene	14.080	228	381352	52.950	ng	100
84) Bis(2-ethylhexyl)phtha...	14.027	149	253542	53.016	ng	100
85) Di-n-octyl phthalate	14.645	149	474143	52.029	ng	100
87) Indeno(1,2,3-cd)pyrene	17.062	276	575271	55.741	ng	100
88) Benzo(b)fluoranthene	15.109	252	401915	49.437	ng	100
89) Benzo(k)fluoranthene	15.139	252	416169	52.440	ng	100
90) Benzo(a)pyrene	15.486	252	419030	53.860	ng	99
91) Dibenzo(a,h)anthracene	17.080	278	466485	55.594	ng	99
92) Benzo(g,h,i)perylene	17.521	276	464951	55.540	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 20 04:40:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061925.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 20 04:36:38 2025
Response via : Initial Calibration

