

Report of Analysis

Client:	Coppola Services		Date Collected:	
Project:	Millville Sewage Treatment Plant - Field Sampling		Date Received:	
Client Sample ID:	PB168614BL		SDG No.:	Q2409
Lab Sample ID:	PB168614BL		Matrix:	TCLP
Analytical Method:	8270E		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142860.D	1	06/25/25 11:10	06/26/25 11:07	PB168614

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	1.30	U	1.30	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.53	U	0.53	5.00	ug/L
95-48-7	2-Methylphenol	1.10	U	1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	1.10	U	1.10	10.0	ug/L
67-72-1	Hexachloroethane	0.65	U	0.65	5.00	ug/L
98-95-3	Nitrobenzene	0.76	U	0.76	5.00	ug/L
87-68-3	Hexachlorobutadiene	0.54	U	0.54	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	0.51	U	0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	0.62	U	0.62	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	1.20	U	1.20	5.00	ug/L
118-74-1	Hexachlorobenzene	0.52	U	0.52	5.00	ug/L
87-86-5	Pentachlorophenol	1.60	U	1.60	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	119		15 (23) - 110 (138)	79%	SPK: 150
13127-88-3	Phenol-d6	118		15 (10) - 110 (134)	79%	SPK: 150
4165-60-0	Nitrobenzene-d5	75.0		30 (67) - 130 (132)	75%	SPK: 100
321-60-8	2-Fluorobiphenyl	72.7		30 (52) - 130 (132)	73%	SPK: 100
118-79-6	2,4,6-Tribromophenol	125		15 (44) - 110 (137)	84%	SPK: 150
1718-51-0	Terphenyl-d14	70.5		30 (42) - 130 (152)	70%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	88200	6.875			
1146-65-2	Naphthalene-d8	340000	8.157			
15067-26-2	Acenaphthene-d10	183000	9.916			
1517-22-2	Phenanthrene-d10	326000	11.404			
1719-03-5	Chrysene-d12	216000	14.051			
1520-96-3	Perylene-d12	170000	15.545			

Report of Analysis

Client:	Coppola Services		Date Collected:	
Project:	Millville Sewage Treatment Plant - Field Sampling		Date Received:	
Client Sample ID:	PB168614BL		SDG No.:	Q2409
Lab Sample ID:	PB168614BL		Matrix:	TCLP
Analytical Method:	8270E		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142860.D	1	06/25/25 11:10	06/26/25 11:07	PB168614

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products