

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2409 SAS No.: Q2409 SDG NO.: Q2409

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/26/2025

Lab File ID: BF142857.D Time Analyzed: 09:39

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	77664	6.875	293795	8.16	161192	9.92
UPPER LIMIT	155328	7.375	587590	8.663	322384	10.422
LOWER LIMIT	38832	6.375	146898	7.663	80596	9.422
EPA SAMPLE NO.						
01 PB168614BL	88228	6.88	339552	8.16	183081	9.92
02 PB168614BS	87019	6.88	334350	8.16	184301	9.92
03 COP-SOIL-PILE	80928	6.88	314887	8.16	165118	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.