

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: COPP02

Lab Code: CHEM Case No.: Q2409 SAS No.: Q2409 SDG No.: Q2409

Instrument ID: MSVOA_N Calibration Date/Time: 06/26/2025 09:50

Lab File ID: VN087181.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.664	0.690		3.92	20
1,1-Dichloroethene	0.557	0.472		-15.26	20
2-Butanone	0.577	0.656		13.69	20
Carbon Tetrachloride	0.433	0.532		22.86	20
Chloroform	1.118	1.334		19.32	20
Benzene	1.444	1.725		19.46	20
1,2-Dichloroethane	0.438	0.565		29	20
Trichloroethene	0.342	0.400		16.96	20
Tetrachloroethene	0.316	0.341		7.91	20
Chlorobenzene	1.103	1.276	0.3	15.68	20
1,2-Dichloroethane-d4	0.669	0.704		5.23	20
Dibromofluoromethane	0.296	0.323		9.12	20
Toluene-d8	1.173	1.166		-0.6	20
4-Bromofluorobenzene	0.436	0.448		2.75	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.