

SOP ID : M1311-TCLP-16
SDG No : N/A
Welgh By : JP
Balance ID : WC SC-7
pH Meter ID : WC PH METER-1
Extraction By : JP
Filter By : JP
Pippete ID : WC
Tumbler ID : ZHE-1
TCLP Filter ID : 50223706

Start Prep Date : 06/24/2025 **Time :** 15:30
End Prep Date : 06/25/2025 **Time :** 08:38
Combination Ratio : 20
ZHE Cleaning Batch : ^{JP} N/A
Initial Room Temperature: 23 °C
Final Room Temperature: 22 °C
TCLP Technician Signature : *JP*
Supervisor By : *12*

Standard Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
TCLP-FLUID-1	N/A	WP112795
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
40ml VOA Vials	430992	N/A

Extraction Conformance/Non-Conformance Comments:

Tumbler ZHE-1 checked, 30 rpm. ALL ZHE samples are extracted and given as vial A & B. Leak checked after 10 minutes of tumbling.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
06/25/25 10:30	<i>JP</i> / <i>ECP Room</i>	<i>S.Y</i> / <i>VOC Lab</i>
	Preparation Group	Analysis Group

Sample ID	ClientID	ZHE Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB168572TB	LEB572	10	N/A	500	N/A	N/A	N/A	4.93	N/A	ZHE-1
Q2386-02	SU-1-062025	01	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q2388-04	TP-12	02	25.03	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q2389-04	MH-J-I	03	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q2391-01	AUD-1623	04	25.01	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q2394-04	MH-K/L	05	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q2399-04	TP-13	06	25.03	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q2399-08	EP-13	07	25.01	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q2405-04	MH-M/N	08	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q2409-01	COP-SOIL-PILE	09	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-1

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
PB168572TB	LEB572	N/A	N/A	N/A	N/A	N/A	N/A
Q2386-02	SU-1-062025	N/A	N/A	N/A	N/A	100	N/A
Q2388-04	TP-12	N/A	N/A	N/A	N/A	100	N/A
Q2389-04	MH-J-I	N/A	N/A	N/A	N/A	100	N/A
Q2391-01	AUD-1623	N/A	N/A	N/A	N/A	100	N/A
Q2394-04	MH-K/L	N/A	N/A	N/A	N/A	100	N/A
Q2399-04	TP-13	N/A	N/A	N/A	N/A	100	N/A
Q2399-08	EP-13	N/A	N/A	N/A	N/A	100	N/A
Q2405-04	MH-M/N	N/A	N/A	N/A	N/A	100	N/A
Q2409-01	COP-SOIL-PILE	N/A	N/A	N/A	N/A	100	N/A