

Surrogate Summary

SDG No.: Q2409

Client: Coppola Services

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2409-01	COP-SOIL-PILE	1,2-Dichloroethane-d4	50	58.5	117	70 (74)	130 (125)
		Dibromofluoromethane	50	51.1	102	70 (75)	130 (124)
		Toluene-d8	50	51.7	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	47.4	95	70 (77)	130 (121)
VN0626WBL01	VN0626WBL01	1,2-Dichloroethane-d4	50	56.6	113	70 (74)	130 (125)
		Dibromofluoromethane	50	57.1	114	70 (75)	130 (124)
		Toluene-d8	50	55.7	111	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.5	101	70 (77)	130 (121)
VN0626WBS01	VN0626WBS01	1,2-Dichloroethane-d4	50	57.9	116	70 (74)	130 (125)
		Dibromofluoromethane	50	60.7	121	70 (75)	130 (124)
		Toluene-d8	50	55.5	111	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.4	109	70 (77)	130 (121)
VN0626WBSD0	VN0626WBSD01	1,2-Dichloroethane-d4	50	48.7	97	70 (74)	130 (125)
		Dibromofluoromethane	50	48.4	97	70 (75)	130 (124)
		Toluene-d8	50	44.7	89	70 (86)	130 (113)
		4-Bromofluorobenzene	50	45.2	90	70 (77)	130 (121)

