

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2409

Client: Coppola Services

Analytical Method: SW8260D

Datafile : VN087184.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0626WBS01	Vinyl chloride	20	16.8	ug/L	84			70 (65)	130 (117)	
	1,1-Dichloroethene	20	19.7	ug/L	99			70 (74)	130 (110)	
	2-Butanone	100	98.3	ug/L	98			40 (65)	160 (122)	
	Carbon Tetrachloride	20	21.5	ug/L	108			70 (77)	130 (113)	
	Chloroform	20	21.2	ug/L	106			70 (79)	130 (113)	
	Benzene	20	21.3	ug/L	106			70 (82)	130 (109)	
	1,2-Dichloroethane	20	23.3	ug/L	117			70 (80)	130 (115)	
	Trichloroethene	20	20.4	ug/L	102			70 (77)	130 (113)	
	Tetrachloroethene	20	18.4	ug/L	92			70 (67)	130 (123)	
	Chlorobenzene	20	20.0	ug/L	100			70 (82)	130 (109)	