

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2409

Client: Coppola Services

Analytical Method: SW8260D

Datafile : VN087187.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0626WBSD01	Vinyl chloride	20	14.8	ug/L	74	13		70 (65)	130 (117)	20 (19)
	1,1-Dichloroethene	20	15.4	ug/L	77	25	*	70 (74)	130 (110)	20 (20)
	2-Butanone	100	86.9	ug/L	87	12		40 (65)	160 (122)	20 (26)
	Carbon Tetrachloride	20	18.0	ug/L	90	18		70 (77)	130 (113)	20 (15)
	Chloroform	20	18.1	ug/L	91	15		70 (79)	130 (113)	20 (20)
	Benzene	20	17.4	ug/L	87	20		70 (82)	130 (109)	20 (15)
	1,2-Dichloroethane	20	18.8	ug/L	94	22	*	70 (80)	130 (115)	20 (20)
	Trichloroethene	20	17.0	ug/L	85	18		70 (77)	130 (113)	20 (15)
	Tetrachloroethene	20	15.9	ug/L	79	15		70 (67)	130 (123)	20 (15)
	Chlorobenzene	20	17.2	ug/L	86	15		70 (82)	130 (109)	20 (15)