



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: CDM Smith SDG No.: Q2413
Contract: CAMP02 Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	105	70 - 130	CV	06/26/2025	14:44	lb136298
CRI01	Aluminum	92.6	100	93	65 - 135	P	06/27/2025	12:58	LB136319
	Antimony	56.3	50.0	113	65 - 135	P	06/27/2025	12:58	LB136319
	Arsenic	16.9	20.0	84	65 - 135	P	06/27/2025	12:58	LB136319
	Barium	106	100	106	65 - 135	P	06/27/2025	12:58	LB136319
	Beryllium	6.39	6.0	106	65 - 135	P	06/27/2025	12:58	LB136319
	Cadmium	5.61	6.0	94	65 - 135	P	06/27/2025	12:58	LB136319
	Calcium	2090	2000	104	65 - 135	P	06/27/2025	12:58	LB136319
	Chromium	12.6	10.0	126	65 - 135	P	06/27/2025	12:58	LB136319
	Cobalt	30.4	30.0	101	65 - 135	P	06/27/2025	12:58	LB136319
	Copper	23.3	20.0	116	65 - 135	P	06/27/2025	12:58	LB136319
	Iron	109	100	108	65 - 135	P	06/27/2025	12:58	LB136319
	Lead	12.2	12.0	101	65 - 135	P	06/27/2025	12:58	LB136319
	Magnesium	2330	2000	117	65 - 135	P	06/27/2025	12:58	LB136319
	Manganese	22.7	20.0	114	65 - 135	P	06/27/2025	12:58	LB136319
	Nickel	40.0	40.0	100	65 - 135	P	06/27/2025	12:58	LB136319
	Potassium	1640	2000	82	65 - 135	P	06/27/2025	12:58	LB136319
	Selenium	20.1	20.0	100	65 - 135	P	06/27/2025	12:58	LB136319
	Silver	11.2	10.0	112	65 - 135	P	06/27/2025	12:58	LB136319
	Sodium	1900	2000	95	65 - 135	P	06/27/2025	12:58	LB136319
	Thallium	40.8	40.0	102	65 - 135	P	06/27/2025	12:58	LB136319
	Vanadium	39.3	40.0	98	65 - 135	P	06/27/2025	12:58	LB136319
	Zinc	34.2	40.0	86	65 - 135	P	06/27/2025	12:58	LB136319