

DIESEL RANGE ORGANICS CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: CAMP02
ProjectID: South River WM Replacement
Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG No.: Q2413
DataFile: FF016056.D Analyst Name: YP\AJ Analyst Date: 07-01-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
500	66132155	132264	134918	1.967