

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF063025\
 Data File : FF016056.D
 Signal(s) : FID2B.ch
 Acq On : 01 Jul 2025 02:08
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Jul 01 03:18:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF061325.M
 Quant Title :
 QLast Update : Fri Jun 13 15:23:30 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.096	5855996	48.499 ug/ml
Target Compounds			
2) N-DECANE	4.621	6563067	50.091 ug/ml
3) N-DODECANE	6.796	6685594	50.208 ug/ml
4) N-TETRADECANE	8.629	6683179	49.866 ug/ml
5) N-HEXADECANE	10.237	6694046	49.565 ug/ml
6) N-OCTADECANE	11.683	6807988	49.108 ug/ml
7) N-EICOSANE	12.995	6807379	48.750 ug/ml
8) N-DOCOSANE	14.196	6599785	48.575 ug/ml
10) N-TETRACOSANE	15.301	6554051	48.359 ug/ml
11) N-HEXACOSANE	16.325	6443992	48.005 ug/ml
12) N-OCTACOSANE	17.277	6293074	47.668 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF063025\
Data File : FF016056.D
Signal(s) : FID2B.ch
Acq On : 01 Jul 2025 02:08
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Jul 01 03:18:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF061325.M
Quant Title :
QLast Update : Fri Jun 13 15:23:30 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

