

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL01 Date Analyzed: 06/27/2025

Lab Sample No.: PSTDCCC750 Data File : PS030838.D Time Analyzed: 17:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.657	9.568	9.768	752.430	712.500	5.6
2,4,5-TP(Silvex)	9.363	9.275	9.475	711.660	712.500	-0.1
2,4-D	8.476	8.386	8.586	681.510	705.000	-3.3
2,4-DB	10.234	10.147	10.347	812.930	712.500	14.1
2,4-DCAA	7.339	7.249	7.449	717.100	750.000	-4.4
DICAMBA	7.531	7.439	7.639	689.260	705.000	-2.2
DICHLORPROP	8.243	8.153	8.353	647.060	705.000	-8.2
Dinoseb	11.453	11.366	11.566	712.890	705.000	1.1

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL01 Date Analyzed: 06/27/2025

Lab Sample No.: PSTDCCC750 Data File : PS030838.D Time Analyzed: 17:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.356	10.259	10.459	699.200	712.500	-1.9
2,4,5-TP(Silvex)	9.929	9.833	10.033	691.030	712.500	-3.0
2,4-D	9.024	8.927	9.127	659.170	705.000	-6.5
2,4-DB	10.923	10.827	11.027	674.590	712.500	-5.3
2,4-DCAA	7.767	7.670	7.870	708.520	750.000	-5.5
DICAMBA	7.970	7.872	8.072	688.710	705.000	-2.3
DICHLORPROP	8.688	8.591	8.791	652.280	705.000	-7.5
Dinoseb	11.308	11.211	11.411	679.930	705.000	-3.6