

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL02 Date Analyzed: 06/27/2025

Lab Sample No.: PSTDCCC750 Data File : PS030850.D Time Analyzed: 21:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.656	9.568	9.768	730.330	712.500	2.5
2,4,5-TP(Silvex)	9.361	9.275	9.475	683.420	712.500	-4.1
2,4-D	8.474	8.386	8.586	659.940	705.000	-6.4
2,4-DB	10.233	10.147	10.347	787.750	712.500	10.6
2,4-DCAA	7.339	7.249	7.449	703.400	750.000	-6.2
DICAMBA	7.530	7.439	7.639	652.000	705.000	-7.5
DICHLORPROP	8.243	8.153	8.353	622.390	705.000	-11.7
Dinoseb	11.452	11.366	11.566	680.110	705.000	-3.5

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL02 Date Analyzed: 06/27/2025

Lab Sample No.: PSTDCCC750 Data File : PS030850.D Time Analyzed: 21:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.355	10.259	10.459	673.490	712.500	-5.5
2,4,5-TP(Silvex)	9.929	9.833	10.033	664.000	712.500	-6.8
2,4-D	9.024	8.927	9.127	634.280	705.000	-10.0
2,4-DB	10.922	10.827	11.027	645.730	712.500	-9.4
2,4-DCAA	7.767	7.670	7.870	682.230	750.000	-9.0
DICAMBA	7.969	7.872	8.072	662.330	705.000	-6.1
DICHLORPROP	8.688	8.591	8.791	622.740	705.000	-11.7
Dinoseb	11.308	11.211	11.411	650.390	705.000	-7.7