

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL03 Date Analyzed: 06/28/2025

Lab Sample No.: PSTDCCC750 Data File : PS030859.D Time Analyzed: 01:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.655	9.568	9.768	761.830	712.500	6.9
2,4,5-TP(Silvex)	9.362	9.275	9.475	703.060	712.500	-1.3
2,4-D	8.474	8.386	8.586	682.180	705.000	-3.2
2,4-DB	10.232	10.147	10.347	803.530	712.500	12.8
2,4-DCAA	7.339	7.249	7.449	690.510	750.000	-7.9
DICAMBA	7.529	7.439	7.639	681.160	705.000	-3.4
DICHLORPROP	8.243	8.153	8.353	635.220	705.000	-9.9
Dinoseb	11.449	11.366	11.566	680.910	705.000	-3.4

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL03 Date Analyzed: 06/28/2025

Lab Sample No.: PSTDCCC750 Data File : PS030859.D Time Analyzed: 01:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.355	10.259	10.459	687.080	712.500	-3.6
2,4,5-TP(Silvex)	9.929	9.833	10.033	677.480	712.500	-4.9
2,4-D	9.022	8.927	9.127	664.420	705.000	-5.8
2,4-DB	10.922	10.827	11.027	667.890	712.500	-6.3
2,4-DCAA	7.766	7.670	7.870	694.470	750.000	-7.4
DICAMBA	7.969	7.872	8.072	673.860	705.000	-4.4
DICHLORPROP	8.688	8.591	8.791	634.380	705.000	-10.0
Dinoseb	11.307	11.211	11.411	645.120	705.000	-8.5