

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL04 Date Analyzed: 06/30/2025

Lab Sample No.: PSTDCCC750 Data File : PS030862.D Time Analyzed: 13:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.660	9.568	9.768	755.930	712.500	6.1
2,4,5-TP(Silvex)	9.365	9.275	9.475	706.160	712.500	-0.9
2,4-D	8.477	8.386	8.586	768.500	705.000	9.0
2,4-DB	10.237	10.147	10.347	779.940	712.500	9.5
2,4-DCAA	7.342	7.249	7.449	707.790	750.000	-5.6
DICAMBA	7.532	7.439	7.639	677.030	705.000	-4.0
DICHLORPROP	8.245	8.153	8.353	742.000	705.000	5.2
Dinoseb	11.454	11.366	11.566	709.220	705.000	0.6

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL04 Date Analyzed: 06/30/2025

Lab Sample No.: PSTDCCC750 Data File : PS030862.D Time Analyzed: 13:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.352	10.259	10.459	662.340	712.500	-7.0
2,4,5-TP(Silvex)	9.925	9.833	10.033	654.920	712.500	-8.1
2,4-D	9.020	8.927	9.127	622.260	705.000	-11.7
2,4-DB	10.919	10.827	11.027	642.880	712.500	-9.8
2,4-DCAA	7.763	7.670	7.870	687.130	750.000	-8.4
DICAMBA	7.966	7.872	8.072	641.910	705.000	-8.9
DICHLORPROP	8.684	8.591	8.791	619.130	705.000	-12.2
Dinoseb	11.304	11.211	11.411	636.090	705.000	-9.8