

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL05 Date Analyzed: 06/30/2025

Lab Sample No.: PSTDCCC750 Data File : PS030866.D Time Analyzed: 17:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.656	9.568	9.768	763.330	712.500	7.1
2,4,5-TP(Silvex)	9.363	9.275	9.475	714.320	712.500	0.3
2,4-D	8.475	8.386	8.586	687.580	705.000	-2.5
2,4-DB	10.233	10.147	10.347	797.880	712.500	12.0
2,4-DCAA	7.340	7.249	7.449	704.330	750.000	-6.1
DICAMBA	7.530	7.439	7.639	684.800	705.000	-2.9
DICHLORPROP	8.243	8.153	8.353	644.430	705.000	-8.6
Dinoseb	11.452	11.366	11.566	711.590	705.000	0.9

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL05 Date Analyzed: 06/30/2025

Lab Sample No.: PSTDCCC750 Data File : PS030866.D Time Analyzed: 17:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.357	10.259	10.459	679.450	712.500	-4.6
2,4,5-TP(Silvex)	9.930	9.833	10.033	670.610	712.500	-5.9
2,4-D	9.025	8.927	9.127	630.620	705.000	-10.6
2,4-DB	10.924	10.827	11.027	664.790	712.500	-6.7
2,4-DCAA	7.768	7.670	7.870	687.550	750.000	-8.3
DICAMBA	7.970	7.872	8.072	661.910	705.000	-6.1
DICHLORPROP	8.689	8.591	8.791	635.760	705.000	-9.8
Dinoseb	11.310	11.211	11.411	654.820	705.000	-7.1