

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS061825

Review By	Abdul	Review On	6/19/2025 8:25:06 AM
Supervise By	mohammad	Supervise On	6/20/2025 8:27:44 AM
SubDirectory	PS061825	HP Acquire Method	HP Processing Method PS061825 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC	PP24559		
Internal Standard/PEM			
ICV/I.BLK	PP24562		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS030737.D	18 Jun 2025 10:37		AR\AJ	Ok
2	I.BLK	I.BLK	PS030738.D	18 Jun 2025 11:01		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS030739.D	18 Jun 2025 11:25		AR\AJ	Ok,M
4	HSTDICC500	HSTDICC500	PS030740.D	18 Jun 2025 11:49		AR\AJ	Ok
5	HSTDICC750	HSTDICC750	PS030741.D	18 Jun 2025 12:13		AR\AJ	Ok,M
6	HSTDICC1000	HSTDICC1000	PS030742.D	18 Jun 2025 12:37		AR\AJ	Ok,M
7	HSTDICC1500	HSTDICC1500	PS030743.D	18 Jun 2025 13:29		AR\AJ	Ok,M
8	HSTDICV750	ICVPS061825	PS030744.D	18 Jun 2025 14:18		AR\AJ	Ok,M
9	I.BLK	I.BLK	PS030745.D	18 Jun 2025 14:44		AR\AJ	Ok
10	HSTDCCC750	HSTDCCC750	PS030746.D	18 Jun 2025 15:08		AR\AJ	Ok,M
11	Q2312-01MS	TP-1MS	PS030747.D	18 Jun 2025 17:07	some compounds fail for recovery	AR\AJ	Ok,M
12	Q2312-01MSD	TP-1MSD	PS030748.D	18 Jun 2025 17:31	some compounds fail for recovery & RPD	AR\AJ	Ok,M
13	I.BLK	I.BLK	PS030749.D	18 Jun 2025 17:55		AR\AJ	Ok
14	HSTDCCC750	HSTDCCC750	PS030750.D	18 Jun 2025 18:19		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS062725

Review By	Abdul	Review On	6/30/2025 8:00:21 AM
Supervise By	mohammad	Supervise On	7/1/2025 3:04:37 AM
SubDirectory	PS062725	HP Acquire Method	HP Processing Method PS061825 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24562
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS030827.D	27 Jun 2025 12:15		AR\AJ	Ok
2	I.BLK	I.BLK	PS030828.D	27 Jun 2025 12:39		AR\AJ	Ok
3	PSTDCCC750	PSTDCCC750	PS030829.D	27 Jun 2025 13:03		AR\AJ	Ok,M
4	PB168638BL	PB168638BL	PS030830.D	27 Jun 2025 13:51		AR\AJ	Ok
5	PB168638BS	PB168638BS	PS030831.D	27 Jun 2025 14:15		AR\AJ	Ok,M
6	PB168638BSD	PB168638BSD	PS030832.D	27 Jun 2025 14:39		AR\AJ	Ok,M
7	Q2418-01	VAC-TRUCK-4074	PS030833.D	27 Jun 2025 15:03		AR\AJ	Ok,M
8	Q2394-01	MH-K/L	PS030834.D	27 Jun 2025 15:27		AR\AJ	Ok
9	Q2399-01	TP-13	PS030835.D	27 Jun 2025 15:52		AR\AJ	Ok
10	Q2399-05	EP-7	PS030836.D	27 Jun 2025 16:16		AR\AJ	Ok
11	I.BLK	I.BLK	PS030837.D	27 Jun 2025 16:40		AR\AJ	Ok
12	PSTDCCC750	PSTDCCC750	PS030838.D	27 Jun 2025 17:04		AR\AJ	Ok,M
13	PB168634BL	PB168634BL	PS030839.D	27 Jun 2025 17:28	Surrogate high in 2nd column	AR\AJ	Not Ok
14	PB168634BS	PB168634BS	PS030840.D	27 Jun 2025 17:52		AR\AJ	Ok,M
15	Q2405-01	MH-M/N	PS030841.D	27 Jun 2025 18:16		AR\AJ	Ok,M
16	Q2405-01MS	MH-M/NMS	PS030842.D	27 Jun 2025 18:40	some compound recovery fail	AR\AJ	Ok,M
17	Q2405-01MSD	MH-M/NMSD	PS030843.D	27 Jun 2025 19:04	some compound recovery fail , RPD Fail	AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS062725

Review By	Abdul	Review On	6/30/2025 8:00:21 AM
Supervise By	mohammad	Supervise On	7/1/2025 3:04:37 AM
SubDirectory	PS062725	HP Acquire Method	HP Processing Method PS061825 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562		

18	Q2413-01	TP-35	PS030844.D	27 Jun 2025 19:29		AR\AJ	Ok
19	Q2413-02	TP-34	PS030845.D	27 Jun 2025 19:53		AR\AJ	Ok,M
20	Q2413-03	TP-77	PS030846.D	27 Jun 2025 20:17		AR\AJ	Ok
21	Q2413-04	TP-74	PS030847.D	27 Jun 2025 20:41		AR\AJ	Ok,M
22	Q2413-05	TP-73	PS030848.D	27 Jun 2025 21:05		AR\AJ	Ok
23	I.BLK	I.BLK	PS030849.D	27 Jun 2025 21:29		AR\AJ	Ok
24	PSTDCCC750	PSTDCCC750	PS030850.D	27 Jun 2025 21:53		AR\AJ	Ok,M
25	Q2413-06	TP-72	PS030851.D	27 Jun 2025 22:41		AR\AJ	Ok,M
26	Q2414-01	WC-1	PS030852.D	27 Jun 2025 23:06		AR\AJ	Ok
27	Q2415-01	WC-1	PS030853.D	27 Jun 2025 23:30		AR\AJ	Ok,M
28	Q2416-01	MH-G/H	PS030854.D	27 Jun 2025 23:54	F Flag in surrogate in 2nd column , surrogate low in 1st column	AR\AJ	ReRun
29	Q2420-01	72-11933	PS030855.D	28 Jun 2025 00:18		AR\AJ	Ok
30	Q2429-01	TP-4	PS030856.D	28 Jun 2025 00:42		AR\AJ	Ok,M
31	Q2430-01	MH-E/F	PS030857.D	28 Jun 2025 01:06		AR\AJ	Ok,M
32	I.BLK	I.BLK	PS030858.D	28 Jun 2025 01:31		AR\AJ	Ok
33	PSTDCCC750	PSTDCCC750	PS030859.D	28 Jun 2025 01:55		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS063025

Review By	Abdul	Review On	7/1/2025 8:28:34 AM
Supervise By	mohammad	Supervise On	7/1/2025 8:56:00 AM
SubDirectory	PS063025	HP Acquire Method	HP Processing Method PS061825 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24562
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS030860.D	30 Jun 2025 09:05		AR\AJ	Ok
2	I.BLK	I.BLK	PS030861.D	30 Jun 2025 09:29		AR\AJ	Ok
3	PSTDCCC750	PSTDCCC750	PS030862.D	30 Jun 2025 13:00		AR\AJ	Ok,M
4	PB168634BL	PB168634BL	PS030863.D	30 Jun 2025 16:33		AR\AJ	Ok
5	Q2416-01RE	MH-G/HRE	PS030864.D	30 Jun 2025 17:01	F Flag in surrogate 2nd column , Surrogate low in 1st column confirm with original run	AR\AJ	Confirms
6	I.BLK	I.BLK	PS030865.D	30 Jun 2025 17:25		AR\AJ	Ok
7	PSTDCCC750	PSTDCCC750	PS030866.D	30 Jun 2025 17:49		AR\AJ	Ok,M

M : Manual Integration