

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 15:26 Initial Calibration Time(s): 10:04 20:10

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1254-1       | (1) | 6.49       | 6.49      | 6.39      | 6.59  | 0.00       |
| Aroclor-1254-2       | (2) | 6.70       | 6.71      | 6.61      | 6.81  | 0.01       |
| Aroclor-1254-3       | (3) | 7.07       | 7.07      | 6.97      | 7.17  | 0.00       |
| Aroclor-1254-4       | (4) | 7.35       | 7.35      | 7.25      | 7.45  | 0.00       |
| Aroclor-1254-5       | (5) | 7.77       | 7.77      | 7.67      | 7.87  | 0.01       |
| Tetrachloro-m-xylene |     | 4.49       | 4.49      | 4.39      | 4.59  | 0.00       |
| Decachlorobiphenyl   |     | 10.18      | 10.19     | 10.09     | 10.29 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 15:26 Initial Calibration Time(s): 10:04 20:10

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1254-1       | (1) | 5.66       | 5.67      | 5.57      | 5.77 | 0.01       |
| Aroclor-1254-2       | (2) | 5.81       | 5.82      | 5.72      | 5.92 | 0.01       |
| Aroclor-1254-3       | (3) | 6.22       | 6.22      | 6.12      | 6.32 | 0.00       |
| Aroclor-1254-4       | (4) | 6.44       | 6.45      | 6.35      | 6.55 | 0.01       |
| Aroclor-1254-5       | (5) | 6.86       | 6.86      | 6.76      | 6.96 | 0.00       |
| Tetrachloro-m-xylene |     | 3.79       | 3.79      | 3.69      | 3.89 | 0.00       |
| Decachlorobiphenyl   |     | 8.79       | 8.80      | 8.70      | 8.90 | 0.01       |