

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 20:19 Initial Calibration Time(s): 10:04 20:10

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1254-1	(1)	6.49	6.49	6.39	6.59	0.00
Aroclor-1254-2	(2)	6.70	6.71	6.61	6.81	0.01
Aroclor-1254-3	(3)	7.07	7.07	6.97	7.17	0.00
Aroclor-1254-4	(4)	7.35	7.35	7.25	7.45	0.00
Aroclor-1254-5	(5)	7.77	7.77	7.67	7.87	0.01
Tetrachloro-m-xylene		4.49	4.49	4.39	4.59	0.00
Decachlorobiphenyl		10.18	10.19	10.09	10.29	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 20:19 Initial Calibration Time(s): 10:04 20:10

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.66	5.67	5.57	5.77	0.01
Aroclor-1254-2 (2)	5.81	5.82	5.72	5.92	0.01
Aroclor-1254-3 (3)	6.21	6.22	6.12	6.32	0.01
Aroclor-1254-4 (4)	6.44	6.45	6.35	6.55	0.01
Aroclor-1254-5 (5)	6.86	6.86	6.76	6.96	0.00
Tetrachloro-m-xylene	3.79	3.79	3.69	3.89	0.00
Decachlorobiphenyl	8.79	8.80	8.70	8.90	0.01