

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL02 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PP073294.D Time Analyzed: 14:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.643	5.546	5.746	477.730	500.000	-4.5
Aroclor-1016-2	5.665	5.567	5.767	514.610	500.000	2.9
Aroclor-1016-3	5.727	5.629	5.829	507.950	500.000	1.6
Aroclor-1016-4	5.824	5.727	5.927	534.040	500.000	6.8
Aroclor-1016-5	6.116	6.019	6.219	509.130	500.000	1.8
Aroclor-1260-1	7.233	7.136	7.336	554.370	500.000	10.9
Aroclor-1260-2	7.486	7.389	7.589	530.980	500.000	6.2
Aroclor-1260-3	7.844	7.747	7.947	525.830	500.000	5.2
Aroclor-1260-4	8.068	7.971	8.171	527.090	500.000	5.4
Aroclor-1260-5	8.386	8.288	8.488	505.830	500.000	1.2
Decachlorobiphenyl	10.182	10.087	10.287	51.630	50.000	3.3
Tetrachloro-m-xylene	4.492	4.394	4.594	49.270	50.000	-1.5

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL02 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PP073294.D Time Analyzed: 14:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.865	4.768	4.968	475.950	500.000	-4.8
Aroclor-1016-2	4.882	4.786	4.986	486.130	500.000	-2.8
Aroclor-1016-3	5.059	4.962	5.162	484.990	500.000	-3.0
Aroclor-1016-4	5.100	5.004	5.204	478.600	500.000	-4.3
Aroclor-1016-5	5.314	5.218	5.418	545.720	500.000	9.1
Aroclor-1260-1	6.344	6.249	6.449	518.030	500.000	3.6
Aroclor-1260-2	6.534	6.438	6.638	531.220	500.000	6.2
Aroclor-1260-3	6.685	6.589	6.789	499.440	500.000	-0.1
Aroclor-1260-4	7.155	7.060	7.260	498.560	500.000	-0.3
Aroclor-1260-5	7.396	7.302	7.502	503.650	500.000	0.7
Decachlorobiphenyl	8.790	8.697	8.897	52.420	50.000	4.8
Tetrachloro-m-xylene	3.786	3.688	3.888	49.210	50.000	-1.6