

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PP073309.D Time Analyzed: 19:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.640	5.546	5.746	451.620	500.000	-9.7
Aroclor-1016-2	5.662	5.567	5.767	494.360	500.000	-1.1
Aroclor-1016-3	5.724	5.629	5.829	490.840	500.000	-1.8
Aroclor-1016-4	5.821	5.727	5.927	522.060	500.000	4.4
Aroclor-1016-5	6.113	6.019	6.219	481.760	500.000	-3.6
Aroclor-1260-1	7.230	7.136	7.336	513.110	500.000	2.6
Aroclor-1260-2	7.483	7.389	7.589	481.840	500.000	-3.6
Aroclor-1260-3	7.842	7.747	7.947	473.110	500.000	-5.4
Aroclor-1260-4	8.065	7.971	8.171	460.670	500.000	-7.9
Aroclor-1260-5	8.383	8.288	8.488	459.970	500.000	-8.0
Decachlorobiphenyl	10.178	10.087	10.287	47.810	50.000	-4.4
Tetrachloro-m-xylene	4.490	4.394	4.594	47.870	50.000	-4.3

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PP073309.D Time Analyzed: 19:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.884	4.768	4.968	487.530	500.000	-2.5
Aroclor-1016-2	4.902	4.786	4.986	475.770	500.000	-4.8
Aroclor-1016-3	5.079	4.962	5.162	480.790	500.000	-3.8
Aroclor-1016-4	5.121	5.004	5.204	474.850	500.000	-5.0
Aroclor-1016-5	5.334	5.218	5.418	533.340	500.000	6.7
Aroclor-1260-1	6.365	6.249	6.449	483.000	500.000	-3.4
Aroclor-1260-2	6.554	6.438	6.638	508.980	500.000	1.8
Aroclor-1260-3	6.706	6.589	6.789	485.230	500.000	-3.0
Aroclor-1260-4	7.175	7.060	7.260	463.340	500.000	-7.3
Aroclor-1260-5	7.417	7.302	7.502	476.860	500.000	-4.6
Decachlorobiphenyl	8.811	8.697	8.897	51.100	50.000	2.2
Tetrachloro-m-xylene	3.806	3.688	3.888	49.010	50.000	-2.0