

Surrogate Summary

SDG No.: Q2413

Client: CDM Smith

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP072990.D	PIBLK-PP072990.D	Tetrachloro-m-xylene	1	20	17.3	86		60	140
		Decachlorobiphenyl	1	20	18.3	91		60	140
		Tetrachloro-m-xylene	2	20	18.0	90		60	140
		Decachlorobiphenyl	2	20	17.2	86		60	140
I.BLK-PP073283.D	PIBLK-PP073283.D	Tetrachloro-m-xylene	1	20	18.9	94		60	140
		Decachlorobiphenyl	1	20	18.6	93		60	140
		Tetrachloro-m-xylene	2	20	18.8	94		60	140
		Decachlorobiphenyl	2	20	20.9	105		60	140
PB168622BL	PB168622BL	Tetrachloro-m-xylene	1	20	19.3	96		32	144
		Decachlorobiphenyl	1	20	20.0	100		32	175
		Tetrachloro-m-xylene	2	20	18.8	94		32	144
		Decachlorobiphenyl	2	20	21.1	106		32	175
PB168622BS	PB168622BS	Tetrachloro-m-xylene	1	20	20.2	101		32	144
		Decachlorobiphenyl	1	20	20.0	100		32	175
		Tetrachloro-m-xylene	2	20	18.6	93		32	144
		Decachlorobiphenyl	2	20	23.0	115		32	175
Q2409-02MS	COP-SOIL-PILEMS	Tetrachloro-m-xylene	1	20	20.0	100		32	144
		Decachlorobiphenyl	1	20	22.5	112		32	175
		Tetrachloro-m-xylene	2	20	18.9	94		32	144
		Decachlorobiphenyl	2	20	22.7	113		32	175
Q2409-02MSD	COP-SOIL-PILEMSD	Tetrachloro-m-xylene	1	20	19.3	97		32	144
		Decachlorobiphenyl	1	20	23.0	115		32	175
		Tetrachloro-m-xylene	2	20	19.2	96		32	144
		Decachlorobiphenyl	2	20	22.7	114		32	175
I.BLK-PP073298.D	PIBLK-PP073298.D	Tetrachloro-m-xylene	1	20	18.3	91		60	140
		Decachlorobiphenyl	1	20	18.9	95		60	140
		Tetrachloro-m-xylene	2	20	18.7	93		60	140
		Decachlorobiphenyl	2	20	20.5	103		60	140
Q2413-01	TP-35	Tetrachloro-m-xylene	1	20	21.8	109		32	144
		Decachlorobiphenyl	1	20	19.4	97		32	175
		Tetrachloro-m-xylene	2	20	21.9	110		32	144
		Decachlorobiphenyl	2	20	21.7	109		32	175
Q2413-02	TP-34	Tetrachloro-m-xylene	1	20	20.7	104		32	144
		Decachlorobiphenyl	1	20	19.9	100		32	175
		Tetrachloro-m-xylene	2	20	21.7	108		32	144
		Decachlorobiphenyl	2	20	20.8	104		32	175
Q2413-03	TP-77	Tetrachloro-m-xylene	1	20	20.7	103		32	144
		Decachlorobiphenyl	1	20	18.3	92		32	175
		Tetrachloro-m-xylene	2	20	20.3	101		32	144
		Decachlorobiphenyl	2	20	19.5	98		32	175
Q2413-04	TP-74	Tetrachloro-m-xylene	1	20	20.0	100		32	144

Surrogate Summary

SDG No.: Q2413

Client: CDM Smith

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q2413-04	TP-74	Decachlorobiphenyl	1	20	16.0	80		32	175
		Tetrachloro-m-xylene	2	20	20.5	102		32	144
Q2413-05	TP-73	Decachlorobiphenyl	2	20	17.1	85		32	175
		Tetrachloro-m-xylene	1	20	22.1	110		32	144
		Decachlorobiphenyl	1	20	18.7	94		32	175
		Tetrachloro-m-xylene	2	20	22.7	114		32	144
		Decachlorobiphenyl	2	20	20.2	101		32	175
		Tetrachloro-m-xylene	1	20	21.0	105		32	144
Q2413-06	TP-72	Decachlorobiphenyl	1	20	16.6	83		32	175
		Tetrachloro-m-xylene	2	20	21.7	109		32	144
		Decachlorobiphenyl	2	20	18.8	94		32	175
		Tetrachloro-m-xylene	1	20	17.4	87		60	140
I.BLK-PP073313.D	PIBLK-PP073313.D	Decachlorobiphenyl	1	20	16.9	85		60	140
		Tetrachloro-m-xylene	2	20	18.9	95		60	140
		Decachlorobiphenyl	2	20	20.5	102		60	140
		Tetrachloro-m-xylene	1	20	17.4	87		60	140