

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL01 Date Analyzed: 06/26/2025

Lab Sample No.: PSTDCCC050 Data File : PD089159.D Time Analyzed: 11:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.705	6.605	6.805	53.110	50.000	6.2
4,4'-DDE	6.196	6.096	6.296	52.180	50.000	4.4
4,4'-DDT	7.021	6.921	7.121	48.890	50.000	-2.2
Aldrin	5.271	5.171	5.371	54.570	50.000	9.1
alpha-BHC	4.000	3.899	4.099	56.190	50.000	12.4
alpha-Chlordane	6.027	5.927	6.127	53.680	50.000	7.4
beta-BHC	4.516	4.415	4.615	53.430	50.000	6.9
Decachlorobiphenyl	9.072	8.972	9.172	49.280	50.000	-1.4
delta-BHC	4.764	4.664	4.864	58.080	50.000	16.2
Dieldrin	6.347	6.247	6.447	53.280	50.000	6.6
Endosulfan I	6.074	5.975	6.175	53.510	50.000	7.0
Endosulfan II	6.786	6.686	6.886	52.520	50.000	5.0
Endosulfan sulfate	7.149	7.049	7.249	50.240	50.000	0.5
Endrin	6.574	6.475	6.675	51.670	50.000	3.3
Endrin aldehyde	6.915	6.815	7.015	50.270	50.000	0.5
Endrin ketone	7.630	7.530	7.730	50.680	50.000	1.4
gamma-BHC (Lindane)	4.331	4.230	4.430	55.550	50.000	11.1
gamma-Chlordane	5.946	5.846	6.046	54.360	50.000	8.7
Heptachlor	4.929	4.829	5.029	54.610	50.000	9.2
Heptachlor epoxide	5.691	5.591	5.791	53.310	50.000	6.6
Methoxychlor	7.493	7.393	7.593	46.320	50.000	-7.4
Tetrachloro-m-xylene	3.551	3.450	3.650	53.870	50.000	7.7

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL01 Date Analyzed: 06/26/2025

Lab Sample No.: PSTDCCC050 Data File : PD089159.D Time Analyzed: 11:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.929	5.830	6.030	51.260	50.000	2.5
4,4'-DDE	5.375	5.275	5.475	51.130	50.000	2.3
4,4'-DDT	6.183	6.084	6.284	48.440	50.000	-3.1
Aldrin	4.369	4.269	4.469	52.050	50.000	4.1
alpha-BHC	3.393	3.293	3.493	52.750	50.000	5.5
alpha-Chlordane	5.189	5.091	5.291	50.880	50.000	1.8
beta-BHC	4.025	3.926	4.126	49.610	50.000	-0.8
Decachlorobiphenyl	8.071	7.972	8.172	47.520	50.000	-5.0
delta-BHC	4.262	4.162	4.362	52.030	50.000	4.1
Dieldrin	5.512	5.413	5.613	51.120	50.000	2.2
Endosulfan I	5.246	5.147	5.347	46.580	50.000	-6.8
Endosulfan II	6.079	5.981	6.181	50.470	50.000	0.9
Endosulfan sulfate	6.481	6.383	6.583	48.980	50.000	-2.0
Endrin	5.789	5.689	5.889	51.790	50.000	3.6
Endrin aldehyde	6.258	6.159	6.359	48.930	50.000	-2.1
Endrin ketone	6.990	6.892	7.092	48.810	50.000	-2.4
gamma-BHC (Lindane)	3.730	3.630	3.830	52.520	50.000	5.0
gamma-Chlordane	5.125	5.026	5.226	51.120	50.000	2.2
Heptachlor	4.083	3.983	4.183	50.700	50.000	1.4
Heptachlor epoxide	4.872	4.773	4.973	51.480	50.000	3.0
Methoxychlor	6.753	6.655	6.855	45.550	50.000	-8.9
Tetrachloro-m-xylene	2.881	2.780	2.980	52.560	50.000	5.1