

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL02 Date Analyzed: 06/26/2025

Lab Sample No.: PSTDCCC050 Data File : PD089172.D Time Analyzed: 17:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.709	6.605	6.805	52.210	50.000	4.4
4,4'-DDE	6.200	6.096	6.296	53.450	50.000	6.9
4,4'-DDT	7.025	6.921	7.121	46.980	50.000	-6.0
Aldrin	5.275	5.171	5.371	55.770	50.000	11.5
alpha-BHC	4.004	3.899	4.099	57.640	50.000	15.3
alpha-Chlordane	6.031	5.927	6.127	53.780	50.000	7.6
beta-BHC	4.520	4.415	4.615	54.870	50.000	9.7
Decachlorobiphenyl	9.077	8.972	9.172	49.220	50.000	-1.6
delta-BHC	4.769	4.664	4.864	59.310	50.000	18.6
Dieldrin	6.351	6.247	6.447	53.140	50.000	6.3
Endosulfan I	6.078	5.975	6.175	53.750	50.000	7.5
Endosulfan II	6.790	6.686	6.886	50.330	50.000	0.7
Endosulfan sulfate	7.153	7.049	7.249	49.420	50.000	-1.2
Endrin	6.578	6.475	6.675	48.980	50.000	-2.0
Endrin aldehyde	6.919	6.815	7.015	49.300	50.000	-1.4
Endrin ketone	7.634	7.530	7.730	49.390	50.000	-1.2
gamma-BHC (Lindane)	4.335	4.230	4.430	56.910	50.000	13.8
gamma-Chlordane	5.950	5.846	6.046	54.970	50.000	9.9
Heptachlor	4.934	4.829	5.029	54.900	50.000	9.8
Heptachlor epoxide	5.695	5.591	5.791	53.490	50.000	7.0
Methoxychlor	7.497	7.393	7.593	42.870	50.000	-14.3
Tetrachloro-m-xylene	3.555	3.450	3.650	57.010	50.000	14.0

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL02 Date Analyzed: 06/26/2025

Lab Sample No.: PSTDCCC050 Data File : PD089172.D Time Analyzed: 17:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.930	5.830	6.030	49.850	50.000	-0.3
4,4'-DDE	5.376	5.275	5.475	51.580	50.000	3.2
4,4'-DDT	6.184	6.084	6.284	46.860	50.000	-6.3
Aldrin	4.369	4.269	4.469	53.840	50.000	7.7
alpha-BHC	3.393	3.293	3.493	54.360	50.000	8.7
alpha-Chlordane	5.191	5.091	5.291	51.880	50.000	3.8
beta-BHC	4.026	3.926	4.126	52.990	50.000	6.0
Decachlorobiphenyl	8.073	7.972	8.172	44.560	50.000	-10.9
delta-BHC	4.262	4.162	4.362	53.600	50.000	7.2
Dieldrin	5.513	5.413	5.613	51.790	50.000	3.6
Endosulfan I	5.247	5.147	5.347	52.450	50.000	4.9
Endosulfan II	6.081	5.981	6.181	50.470	50.000	0.9
Endosulfan sulfate	6.483	6.383	6.583	48.080	50.000	-3.8
Endrin	5.789	5.689	5.889	48.870	50.000	-2.3
Endrin aldehyde	6.260	6.159	6.359	48.550	50.000	-2.9
Endrin ketone	6.992	6.892	7.092	46.790	50.000	-6.4
gamma-BHC (Lindane)	3.730	3.630	3.830	54.110	50.000	8.2
gamma-Chlordane	5.126	5.026	5.226	52.120	50.000	4.2
Heptachlor	4.083	3.983	4.183	51.710	50.000	3.4
Heptachlor epoxide	4.873	4.773	4.973	54.670	50.000	9.3
Methoxychlor	6.755	6.655	6.855	43.150	50.000	-13.7
Tetrachloro-m-xylene	2.881	2.780	2.980	56.090	50.000	12.2