

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/27/2025

Lab Sample No.: PSTDCCC050 Data File : PD089176.D Time Analyzed: 03:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.704	6.605	6.805	55.140	50.000	10.3
4,4'-DDE	6.195	6.096	6.296	55.960	50.000	11.9
4,4'-DDT	7.021	6.921	7.121	49.570	50.000	-0.9
Aldrin	5.271	5.171	5.371	58.050	50.000	16.1
alpha-BHC	4.000	3.899	4.099	59.200	50.000	18.4
alpha-Chlordane	6.027	5.927	6.127	55.950	50.000	11.9
beta-BHC	4.516	4.415	4.615	56.150	50.000	12.3
Decachlorobiphenyl	9.072	8.972	9.172	51.290	50.000	2.6
delta-BHC	4.764	4.664	4.864	61.300	50.000	22.6
Dieldrin	6.347	6.247	6.447	55.610	50.000	11.2
Endosulfan I	6.074	5.975	6.175	55.920	50.000	11.8
Endosulfan II	6.786	6.686	6.886	53.380	50.000	6.8
Endosulfan sulfate	7.149	7.049	7.249	52.100	50.000	4.2
Endrin	6.574	6.475	6.675	51.810	50.000	3.6
Endrin aldehyde	6.914	6.815	7.015	52.050	50.000	4.1
Endrin ketone	7.629	7.530	7.730	51.950	50.000	3.9
gamma-BHC (Lindane)	4.331	4.230	4.430	58.540	50.000	17.1
gamma-Chlordane	5.946	5.846	6.046	57.530	50.000	15.1
Heptachlor	4.929	4.829	5.029	56.570	50.000	13.1
Heptachlor epoxide	5.690	5.591	5.791	56.070	50.000	12.1
Methoxychlor	7.493	7.393	7.593	44.260	50.000	-11.5
Tetrachloro-m-xylene	3.551	3.450	3.650	58.530	50.000	17.1

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/27/2025

Lab Sample No.: PSTDCCC050 Data File : PD089176.D Time Analyzed: 03:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.928	5.830	6.030	51.620	50.000	3.2
4,4'-DDE	5.374	5.275	5.475	53.890	50.000	7.8
4,4'-DDT	6.182	6.084	6.284	48.080	50.000	-3.8
Aldrin	4.368	4.269	4.469	55.330	50.000	10.7
alpha-BHC	3.393	3.293	3.493	55.980	50.000	12.0
alpha-Chlordane	5.189	5.091	5.291	53.700	50.000	7.4
beta-BHC	4.025	3.926	4.126	54.810	50.000	9.6
Decachlorobiphenyl	8.071	7.972	8.172	46.750	50.000	-6.5
delta-BHC	4.262	4.162	4.362	55.070	50.000	10.1
Dieldrin	5.512	5.413	5.613	53.990	50.000	8.0
Endosulfan I	5.246	5.147	5.347	54.380	50.000	8.8
Endosulfan II	6.079	5.981	6.181	52.150	50.000	4.3
Endosulfan sulfate	6.481	6.383	6.583	50.440	50.000	0.9
Endrin	5.788	5.689	5.889	50.380	50.000	0.8
Endrin aldehyde	6.258	6.159	6.359	50.110	50.000	0.2
Endrin ketone	6.990	6.892	7.092	50.220	50.000	0.4
gamma-BHC (Lindane)	3.729	3.630	3.830	55.710	50.000	11.4
gamma-Chlordane	5.125	5.026	5.226	53.800	50.000	7.6
Heptachlor	4.082	3.983	4.183	53.010	50.000	6.0
Heptachlor epoxide	4.872	4.773	4.973	55.860	50.000	11.7
Methoxychlor	6.752	6.655	6.855	44.470	50.000	-11.1
Tetrachloro-m-xylene	2.881	2.780	2.980	58.020	50.000	16.0