

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL04 Date Analyzed: 06/27/2025

Lab Sample No.: PSTDCCC050 Data File : PD089188.D Time Analyzed: 12:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.706	6.605	6.805	52.920	50.000	5.8
4,4'-DDE	6.197	6.096	6.296	52.760	50.000	5.5
4,4'-DDT	7.022	6.921	7.121	41.870	50.000	-16.3
Aldrin	5.273	5.171	5.371	56.700	50.000	13.4
alpha-BHC	4.001	3.899	4.099	58.280	50.000	16.6
alpha-Chlordane	6.028	5.927	6.127	54.740	50.000	9.5
beta-BHC	4.518	4.415	4.615	55.030	50.000	10.1
Decachlorobiphenyl	9.074	8.972	9.172	47.460	50.000	-5.1
delta-BHC	4.766	4.664	4.864	59.810	50.000	19.6
Dieldrin	6.348	6.247	6.447	53.510	50.000	7.0
Endosulfan I	6.076	5.975	6.175	53.660	50.000	7.3
Endosulfan II	6.787	6.686	6.886	51.560	50.000	3.1
Endosulfan sulfate	7.151	7.049	7.249	48.780	50.000	-2.4
Endrin	6.575	6.475	6.675	49.740	50.000	-0.5
Endrin aldehyde	6.916	6.815	7.015	47.780	50.000	-4.4
Endrin ketone	7.631	7.530	7.730	48.700	50.000	-2.6
gamma-BHC (Lindane)	4.332	4.230	4.430	57.160	50.000	14.3
gamma-Chlordane	5.947	5.846	6.046	56.600	50.000	13.2
Heptachlor	4.931	4.829	5.029	51.810	50.000	3.6
Heptachlor epoxide	5.692	5.591	5.791	53.630	50.000	7.3
Methoxychlor	7.495	7.393	7.593	41.710	50.000	-16.6
Tetrachloro-m-xylene	3.553	3.450	3.650	57.800	50.000	15.6

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL04 Date Analyzed: 06/27/2025

Lab Sample No.: PSTDCCC050 Data File : PD089188.D Time Analyzed: 12:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.928	5.830	6.030	50.730	50.000	1.5
4,4'-DDE	5.374	5.275	5.475	52.450	50.000	4.9
4,4'-DDT	6.182	6.084	6.284	42.840	50.000	-14.3
Aldrin	4.368	4.269	4.469	54.590	50.000	9.2
alpha-BHC	3.393	3.293	3.493	55.380	50.000	10.8
alpha-Chlordane	5.189	5.091	5.291	53.310	50.000	6.6
beta-BHC	4.025	3.926	4.126	53.250	50.000	6.5
Decachlorobiphenyl	8.072	7.972	8.172	44.640	50.000	-10.7
delta-BHC	4.261	4.162	4.362	54.340	50.000	8.7
Dieldrin	5.512	5.413	5.613	52.640	50.000	5.3
Endosulfan I	5.246	5.147	5.347	52.920	50.000	5.8
Endosulfan II	6.079	5.981	6.181	51.330	50.000	2.7
Endosulfan sulfate	6.481	6.383	6.583	48.770	50.000	-2.5
Endrin	5.788	5.689	5.889	49.850	50.000	-0.3
Endrin aldehyde	6.257	6.159	6.359	48.640	50.000	-2.7
Endrin ketone	6.990	6.892	7.092	47.270	50.000	-5.5
gamma-BHC (Lindane)	3.729	3.630	3.830	54.810	50.000	9.6
gamma-Chlordane	5.125	5.026	5.226	54.490	50.000	9.0
Heptachlor	4.082	3.983	4.183	49.790	50.000	-0.4
Heptachlor epoxide	4.871	4.773	4.973	54.520	50.000	9.0
Methoxychlor	6.753	6.655	6.855	43.610	50.000	-12.8
Tetrachloro-m-xylene	2.881	2.780	2.980	57.130	50.000	14.3