

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL05 Date Analyzed: 06/27/2025

Lab Sample No.: PSTDCCC050 Data File : PD089192.D Time Analyzed: 18:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.704	6.605	6.805	57.030	50.000	14.1
4,4'-DDE	6.195	6.096	6.296	55.130	50.000	10.3
4,4'-DDT	7.020	6.921	7.121	43.230	50.000	-13.5
Aldrin	5.270	5.171	5.371	57.560	50.000	15.1
alpha-BHC	4.000	3.899	4.099	58.380	50.000	16.8
alpha-Chlordane	6.026	5.927	6.127	56.760	50.000	13.5
beta-BHC	4.515	4.415	4.615	55.200	50.000	10.4
Decachlorobiphenyl	9.071	8.972	9.172	50.650	50.000	1.3
delta-BHC	4.763	4.664	4.864	59.910	50.000	19.8
Dieldrin	6.346	6.247	6.447	56.130	50.000	12.3
Endosulfan I	6.073	5.975	6.175	55.650	50.000	11.3
Endosulfan II	6.785	6.686	6.886	54.300	50.000	8.6
Endosulfan sulfate	7.148	7.049	7.249	52.510	50.000	5.0
Endrin	6.573	6.475	6.675	51.200	50.000	2.4
Endrin aldehyde	6.914	6.815	7.015	50.910	50.000	1.8
Endrin ketone	7.629	7.530	7.730	52.730	50.000	5.5
gamma-BHC (Lindane)	4.330	4.230	4.430	57.270	50.000	14.5
gamma-Chlordane	5.945	5.846	6.046	58.450	50.000	16.9
Heptachlor	4.929	4.829	5.029	52.650	50.000	5.3
Heptachlor epoxide	5.690	5.591	5.791	55.210	50.000	10.4
Methoxychlor	7.492	7.393	7.593	44.350	50.000	-11.3
Tetrachloro-m-xylene	3.550	3.450	3.650	58.070	50.000	16.1

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL05 Date Analyzed: 06/27/2025

Lab Sample No.: PSTDCCC050 Data File : PD089192.D Time Analyzed: 18:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.928	5.830	6.030	53.960	50.000	7.9
4,4'-DDE	5.374	5.275	5.475	54.320	50.000	8.6
4,4'-DDT	6.182	6.084	6.284	43.220	50.000	-13.6
Aldrin	4.367	4.269	4.469	55.730	50.000	11.5
alpha-BHC	3.393	3.293	3.493	56.430	50.000	12.9
alpha-Chlordane	5.189	5.091	5.291	54.770	50.000	9.5
beta-BHC	4.024	3.926	4.126	54.310	50.000	8.6
Decachlorobiphenyl	8.070	7.972	8.172	50.160	50.000	0.3
delta-BHC	4.261	4.162	4.362	55.240	50.000	10.5
Dieldrin	5.511	5.413	5.613	54.310	50.000	8.6
Endosulfan I	5.246	5.147	5.347	54.500	50.000	9.0
Endosulfan II	6.079	5.981	6.181	53.660	50.000	7.3
Endosulfan sulfate	6.481	6.383	6.583	51.810	50.000	3.6
Endrin	5.788	5.689	5.889	51.230	50.000	2.5
Endrin aldehyde	6.257	6.159	6.359	50.400	50.000	0.8
Endrin ketone	6.990	6.892	7.092	51.820	50.000	3.6
gamma-BHC (Lindane)	3.729	3.630	3.830	55.860	50.000	11.7
gamma-Chlordane	5.125	5.026	5.226	55.840	50.000	11.7
Heptachlor	4.082	3.983	4.183	50.960	50.000	1.9
Heptachlor epoxide	4.872	4.773	4.973	55.580	50.000	11.2
Methoxychlor	6.753	6.655	6.855	46.290	50.000	-7.4
Tetrachloro-m-xylene	2.881	2.780	2.980	58.440	50.000	16.9