

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG No.: Q2413
Instrument ID: BNA_F Calibration Date(s): 06/19/2025 06/19/2025
Calibration Time(s): 17:07 20:40

LAB FILE ID:		RRF2.5 = BF142788.D			RRF005 = BF142789.D		RRF010 = BF142790.D	
		RRF020 = BF142791.D			RRF040 = BF142792.D		RRF050 = BF142793.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.238	1.230	1.230	1.149	1.263	1.201	4.3
Benzaldehyde			0.988	0.943	0.757	0.832	0.841	15.0
Phenol-d6		1.512	1.481	1.423	1.340	1.480	1.417	5.5
Phenol		1.648	1.653	1.591	1.516	1.651	1.575	5.3
bis(2-Chloroethyl) ether		1.167	1.133	1.134	1.072	1.183	1.126	3.7
2-Chlorophenol		1.362	1.303	1.292	1.245	1.366	1.296	4.1
2-Methylphenol		1.015	0.997	0.980	0.938	1.043	0.982	4.1
2,2-oxybis(1-Chloropropane)		1.975	1.869	1.815	1.723	1.879	1.809	6.1
Acetophenone		0.464	0.451	0.459	0.418	0.461	0.441	5.3
3+4-Methylphenols			1.292	1.262	1.176	1.312	1.224	6.4
n-Nitroso-di-n-propylamine	0.831	0.879	0.840	0.833	0.787	0.847	0.824	4.3
Nitrobenzene-d5		0.369	0.362	0.373	0.354	0.386	0.365	3.5
Hexachloroethane		0.533	0.521	0.507	0.499	0.539	0.511	3.9
Nitrobenzene		0.345	0.325	0.332	0.319	0.349	0.330	3.9
Isophorone		0.611	0.582	0.592	0.555	0.616	0.585	4.2
2-Nitrophenol		0.169	0.172	0.182	0.178	0.194	0.180	4.8
2,4-Dimethylphenol		0.322	0.310	0.318	0.300	0.329	0.311	4.0
bis(2-Chloroethoxy) methane		0.386	0.373	0.383	0.355	0.390	0.372	4.1
2,4-Dichlorophenol		0.291	0.280	0.290	0.273	0.298	0.282	3.8
Naphthalene		1.051	1.002	1.004	0.934	1.017	0.979	5.5
4-Chloroaniline		0.414	0.408	0.409	0.379	0.421	0.398	5.3
Hexachlorobutadiene		0.195	0.188	0.198	0.184	0.200	0.190	4.3
Caprolactam			0.073	0.075	0.072	0.081	0.075	5.8
4-Chloro-3-methylphenol		0.300	0.275	0.286	0.269	0.298	0.281	5.2
2-Methylnaphthalene		0.655	0.621	0.629	0.579	0.631	0.607	6.1
Hexachlorocyclopentadiene			0.235	0.308	0.336	0.365	0.335	17.0
2,4,6-Trichlorophenol		0.385	0.380	0.394	0.365	0.408	0.385	3.7
2-Fluorobiphenyl		1.740	1.636	1.592	1.430	1.500	1.522	9.2
2,4,5-Trichlorophenol		0.407	0.403	0.421	0.391	0.416	0.405	3.0
1,1-Biphenyl		1.701	1.616	1.637	1.505	1.609	1.580	5.3
2-Chloronaphthalene		1.300	1.202	1.208	1.129	1.209	1.186	5.5
2-Nitroaniline		0.321	0.329	0.337	0.324	0.358	0.332	3.9
Dimethylphthalate		1.372	1.334	1.330	1.228	1.335	1.295	5.2
Acenaphthylene		2.110	2.006	2.011	1.861	1.995	1.952	5.6

All other compounds must meet a minimum RRF of 0.010.

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG No.: Q2413
Instrument ID: BNA_F Calibration Date(s): 06/19/2025 06/19/2025
Calibration Time(s): 17:07 20:40

LAB FILE ID:		RRF2.5 = BF142788.D			RRF005 = BF142789.D		RRF010 = BF142790.D	
		RRF020 = BF142791.D			RRF040 = BF142792.D		RRF050 = BF142793.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.297	0.279	0.290	0.277	0.294	0.284	3.6
3-Nitroaniline		0.311	0.311	0.316	0.302	0.338	0.313	4.5
Acenaphthene		1.247	1.199	1.241	1.142	1.228	1.191	4.4
2,4-Dinitrophenol			0.095	0.129	0.142	0.173	0.142	19.6
4-Nitrophenol			0.189	0.218	0.208	0.238	0.214	8.3
Dibenzofuran		1.844	1.772	1.763	1.626	1.752	1.708	6.1
2,4-Dinitrotoluene		0.373	0.378	0.397	0.367	0.407	0.378	5.9
Diethylphthalate		1.328	1.302	1.319	1.211	1.318	1.269	5.5
4-Chlorophenyl-phenylether		0.713	0.656	0.669	0.599	0.640	0.636	7.6
Fluorene		1.464	1.412	1.395	1.260	1.362	1.334	7.6
4-Nitroaniline		0.279	0.276	0.297	0.272	0.316	0.286	6.5
4,6-Dinitro-2-methylphenol			0.099	0.119	0.119	0.134	0.121	10.0
n-Nitrosodiphenylamine		0.736	0.686	0.710	0.670	0.717	0.693	4.1
2,4,6-Tribromophenol		0.208	0.207	0.214	0.200	0.219	0.206	5.0
4-Bromophenyl-phenylether		0.231	0.221	0.235	0.218	0.236	0.228	3.0
Hexachlorobenzene		0.263	0.248	0.257	0.241	0.262	0.253	3.3
Atrazine		0.176	0.172	0.180	0.174	0.195	0.179	4.5
Pentachlorophenol			0.102	0.119	0.125	0.142	0.126	11.3
Phenanthrene		1.183	1.113	1.115	1.031	1.122	1.083	6.2
Anthracene		1.226	1.123	1.149	1.064	1.140	1.111	6.3
Carbazole		1.040	0.986	0.999	0.900	1.002	0.959	6.7
Di-n-butylphthalate		0.997	1.006	1.045	0.946	1.073	0.999	4.9
Fluoranthene		1.136	1.088	1.090	0.954	1.059	1.028	8.4
Pyrene		2.110	1.940	1.968	1.742	2.030	1.875	11.4
Terphenyl-d14		1.691	1.556	1.533	1.327	1.535	1.447	13.3
Butylbenzylphthalate		0.478	0.470	0.538	0.547	0.613	0.544	10.0
3,3-Dichlorobenzidine			0.398	0.440	0.429	0.476	0.438	5.7
Benzo (a) anthracene		1.392	1.358	1.366	1.265	1.403	1.349	4.0
Chrysene		1.243	1.191	1.208	1.164	1.295	1.204	4.2
Bis (2-ethylhexyl) phthalate		0.650	0.675	0.767	0.803	0.861	0.782	11.5
Di-n-octyl phthalate			1.207	1.352	1.463	1.610	1.475	11.5
Benzo (b) fluoranthene		1.164	1.220	1.226	1.072	1.152	1.157	5.6
Benzo (k) fluoranthene		1.206	1.004	1.035	1.037	1.193	1.083	7.5
Benzo (a) pyrene		1.142	1.065	1.124	1.069	1.201	1.118	4.4
Indeno (1,2,3-cd) pyrene		1.528	1.488	1.545	1.474	1.648	1.523	4.4

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG No.: Q2413
Instrument ID: BNA_F Calibration Date(s): 06/19/2025 06/19/2025
Calibration Time(s): 17:07 20:40

LAB FILE ID:		RRF2.5 = BF142788.D			RRF005 = BF142789.D		RRF010 = BF142790.D	
		RRF020 = BF142791.D			RRF040 = BF142792.D		RRF050 = BF142793.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo (a,h) anthracene		1.249	1.214	1.293	1.180	1.337	1.238	5.1
Benzo (g,h,i) perylene		1.239	1.198	1.266	1.188	1.332	1.234	4.7
1,2,4,5-Tetrachlorobenzene		0.626	0.609	0.602	0.557	0.598	0.590	4.3
1,4-Dioxane		0.479	0.499	0.471	0.461	0.509	0.480	3.6
2,3,4,6-Tetrachlorophenol		0.327	0.327	0.339	0.312	0.345	0.326	4.4

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF061925.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jun 20 05:06:14 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BF142788.D 5 =BF142789.D 10 =BF142790.D 20 =BF142791.D 40 =BF142792.D 50 =BF142793.D 60 =BF142794.D 80 =BF142795.D

	Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----									
2)	1,4-Dioxane	0.479	0.499	0.471	0.461	0.509	0.479	0.467	0.480		3.60
3)	Pyridine	1.190	1.195	1.190	1.164	1.303	1.209	1.181	1.205		3.78
4)	n-Nitrosodimet...		0.610	0.617	0.596	0.668	0.632	0.614	0.623		3.98
5) S	2-Fluorophenol	2.475	2.460	2.460	2.299	2.525	2.342	2.256	2.403		4.26
6)	Aniline	1.951	1.944	1.891	1.818	1.982	1.869	1.746	1.886		4.41
7) S	Phenol-d6	3.024	2.962	2.846	2.681	2.959	2.760	2.610	2.834		5.53
8)	2-Chlorophenol	1.362	1.303	1.292	1.245	1.366	1.272	1.231	1.296		4.08
9)	Benzaldehyde		0.988	0.943	0.757	0.832	0.684		0.841		15.02
10) C	Phenol	1.648	1.653	1.591	1.516	1.651	1.530	1.437	1.575		5.31
11)	bis(2-Chloroet...	1.167	1.133	1.134	1.072	1.183	1.116	1.077	1.126		3.71
12)	1,3-Dichlorobe...	1.536	1.499	1.470	1.377	1.500	1.414	1.348	1.449		4.85
13) C	1,4-Dichlorobe...	1.563	1.496	1.472	1.409	1.523	1.425	1.347	1.462		5.04
14)	1,2-Dichlorobe...	1.458	1.422	1.409	1.331	1.451	1.353	1.284	1.387		4.72
15)	Benzyl Alcohol		1.023	1.025	0.976	1.106	1.032	0.985	1.025		4.51
16)	2,2'-oxybis(1-...	1.975	1.869	1.815	1.723	1.879	1.758	1.644	1.809		6.12
17)	2-Methylphenol	1.015	0.997	0.980	0.938	1.043	0.972	0.929	0.982		4.13
18)	Hexachloroethane	0.533	0.521	0.507	0.499	0.539	0.498	0.484	0.511		3.89
19) P	n-Nitroso-di-n...	0.831	0.879	0.840	0.833	0.787	0.847	0.812	0.766	0.824	4.30
20)	3+4-Methylphenols		1.292	1.262	1.176	1.312	1.197	1.108	1.224		6.36
21) I	Naphthalene-d8	-----ISTD-----									
22)	Acetophenone	0.464	0.451	0.459	0.418	0.461	0.428	0.406	0.441		5.28
23) S	Nitrobenzene-d5	0.738	0.724	0.747	0.708	0.771	0.724	0.694	0.729		3.48
24)	Nitrobenzene	0.345	0.325	0.332	0.319	0.349	0.326	0.314	0.330		3.88
25)	Isophorone	0.611	0.582	0.592	0.555	0.616	0.584	0.554	0.585		4.17
26) C	2-Nitrophenol	0.169	0.172	0.182	0.178	0.194	0.186	0.176	0.180		4.85
27)	2,4-Dimethylph...	0.322	0.310	0.318	0.300	0.329	0.307	0.294	0.311		3.99
28)	bis(2-Chloroet...	0.386	0.373	0.383	0.355	0.390	0.366	0.351	0.372		4.09
29) C	2,4-Dichloroph...	0.291	0.280	0.290	0.273	0.298	0.277	0.268	0.282		3.77
30)	1,2,4-Trichlor...	0.324	0.316	0.318	0.299	0.324	0.301	0.292	0.310		4.26
31)	Naphthalene	1.051	1.002	1.004	0.934	1.017	0.943	0.902	0.979		5.46
32)	Benzoic acid		0.118	0.147	0.157	0.181	0.180	0.169	0.159		15.08
33)	4-Chloroaniline	0.414	0.408	0.409	0.379	0.421	0.393	0.363	0.398		5.26
34) C	Hexachlorobuta...	0.195	0.188	0.198	0.184	0.200	0.184	0.179	0.190		4.30
35)	Caprolactam		0.073	0.075	0.072	0.081	0.078	0.069	0.075		5.84
36) C	4-Chloro-3-met...	0.300	0.275	0.286	0.269	0.298	0.278	0.260	0.281		5.24
37)	2-Methylnaphth...	0.655	0.621	0.629	0.579	0.631	0.583	0.550	0.607		6.12
38)	1-Methylnaphth...	0.684	0.645	0.650	0.598	0.654	0.604	0.563	0.628		6.56

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF061925.M

39) I	Acenaphthene-d10	-----ISTD-----									
40)	1,2,4,5-Tetrac...	0.626	0.609	0.602	0.557	0.598	0.577	0.563	0.590	4.27	
41) P	Hexachlorocycl...	0.235	0.308	0.336	0.365	0.376	0.389	0.335		17.02	
42) S	2,4,6-Tribromo...	0.415	0.414	0.428	0.400	0.438	0.411	0.374	0.412	4.96	
43) C	2,4,6-Trichlor...	0.385	0.380	0.394	0.365	0.408	0.389	0.371	0.385	3.71	
44)	2,4,5-Trichlor...	0.407	0.403	0.421	0.391	0.416	0.407	0.388	0.405	2.95	
45) S	2-Fluorobiphenyl	3.480	3.272	3.184	2.861	2.999	2.816	2.702	3.045	9.15	
46)	1,1'-Biphenyl	1.701	1.616	1.637	1.505	1.609	1.525	1.466	1.580	5.27	
47)	2-Chloronaphth...	1.300	1.202	1.208	1.129	1.209	1.145	1.106	1.186	5.50	
48)	2-Nitroaniline	0.321	0.329	0.337	0.324	0.358	0.337	0.320	0.332	3.93	
49)	Acenaphthylene	2.110	2.006	2.011	1.861	1.995	1.891	1.787	1.952	5.64	
50)	Dimethylphthalate	1.372	1.334	1.330	1.228	1.335	1.277	1.186	1.295	5.17	
51)	2,6-Dinitrotol...	0.297	0.279	0.290	0.277	0.294	0.284	0.268	0.284	3.57	
52) C	Acenaphthene	1.247	1.199	1.241	1.142	1.228	1.172	1.107	1.191	4.44	
53)	3-Nitroaniline	0.311	0.311	0.316	0.302	0.338	0.320	0.292	0.313	4.52	
54) P	2,4-Dinitrophenol	0.095	0.129	0.142	0.173	0.163	0.152	0.142		19.61	
55)	Dibenzofuran	1.844	1.772	1.763	1.626	1.752	1.660	1.537	1.708	6.13	
56) P	4-Nitrophenol	0.189	0.218	0.208	0.238	0.229	0.203	0.214		8.25	
57)	2,4-Dinitrotol...	0.373	0.378	0.397	0.367	0.407	0.383	0.338	0.378	5.90	
58)	Fluorene	1.464	1.412	1.395	1.260	1.362	1.260	1.183	1.334	7.58	
59)	2,3,4,6-Tetrac...	0.327	0.327	0.339	0.312	0.345	0.331	0.304	0.326	4.43	
60)	Diethylphthalate	1.328	1.302	1.319	1.211	1.318	1.266	1.141	1.269	5.50	
61)	4-Chlorophenyl...	0.713	0.656	0.669	0.599	0.640	0.609	0.568	0.636	7.64	
62)	4-Nitroaniline	0.279	0.276	0.297	0.272	0.316	0.300	0.264	0.286	6.48	
63)	Azobenzene	1.215	1.149	1.167	1.081	1.176	1.122	1.032	1.134	5.47	
64) I	Phenanthrene-d10	-----ISTD-----									
65)	4,6-Dinitro-2-...	0.099	0.119	0.119	0.134	0.130	0.125	0.121		10.04	
66) c	n-Nitrosodiphe...	0.736	0.686	0.710	0.670	0.717	0.671	0.662	0.693	4.07	
67)	4-Bromophenyl-...	0.231	0.221	0.235	0.218	0.236	0.228	0.225	0.228	3.03	
68)	Hexachlorobenzene	0.263	0.248	0.257	0.241	0.262	0.252	0.247	0.253	3.29	
69)	Atrazine	0.176	0.172	0.180	0.174	0.195	0.183	0.174	0.179	4.48	
70) C	Pentachlorophenol	0.102	0.119	0.125	0.142	0.135	0.133	0.126		11.30	
71)	Phenanthrene	1.183	1.113	1.115	1.031	1.122	1.028	0.992	1.083	6.24	
72)	Anthracene	1.226	1.123	1.149	1.064	1.140	1.065	1.011	1.111	6.35	
73)	Carbazole	1.040	0.986	0.999	0.900	1.002	0.920	0.866	0.959	6.67	
74)	Di-n-butylphth...	0.997	1.006	1.045	0.946	1.073	0.989	0.936	0.999	4.92	
75) C	Fluoranthene	1.136	1.088	1.090	0.954	1.059	0.972	0.899	1.028	8.44	
76) I	Chrysene-d12	-----ISTD-----									
77)	Benzidine	0.745	0.860	0.773	0.824	0.719	0.587	0.751		12.71	
78)	Pyrene	2.110	1.940	1.968	1.742	2.030	1.863	1.468	1.875	11.43	
79) S	Terphenyl-d14	3.383	3.111	3.066	2.655	3.070	2.772	2.208	2.895	13.30	
80)	Butylbenzylpht...	0.478	0.470	0.538	0.547	0.613	0.597	0.563	0.544	10.01	
81)	Benzo(a)anthra...	1.392	1.358	1.366	1.265	1.403	1.376	1.284	1.349	3.98	
82)	3,3'-Dichlorob...	0.398	0.440	0.429	0.476	0.445	0.438	0.438		5.73	
83)	Chrysene	1.243	1.191	1.208	1.164	1.295	1.177	1.150	1.204	4.17	
84)	Bis(2-ethylhex...	0.650	0.675	0.767	0.803	0.861	0.843	0.878	0.782	11.52	
85) c	Di-n-octyl pht...	1.207	1.352	1.463	1.610	1.569	1.651	1.475		11.54	

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF061925.M

86) I	Perylene-d12	-----ISTD-----								
87)	Indeno(1,2,3-c...	1.528	1.488	1.545	1.474	1.648	1.539	1.442	1.523	4.38
88)	Benzo(b)fluora...	1.164	1.220	1.226	1.072	1.152	1.197	1.070	1.157	5.61
89)	Benzo(k)fluora...	1.206	1.004	1.035	1.037	1.193	1.038	1.068	1.083	7.55
90) C	Benzo(a)pyrene	1.142	1.065	1.124	1.069	1.201	1.141	1.084	1.118	4.37
91)	Dibenzo(a,h)an...	1.249	1.214	1.293	1.180	1.337	1.240	1.152	1.238	5.13
92)	Benzo(g,h,i)pe...	1.239	1.198	1.266	1.188	1.332	1.255	1.161	1.234	4.66

(#) = Out of Range