

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q2413 SDG NO.: Q2413

Lab File ID: BF142787.D

DFTPP Injection Date: 06/19/2025

Instrument ID: BNA_F

DFTPP Injection Time: 16:37

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (1.8) 1
69	Mass 69 relative abundance	30.9
70	Less than 2.0% of mass 69	0.1 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.8
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	15
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.1 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF142788.D	06/19/2025	17:07
SSTDICC005	SSTDICC005	BF142789.D	06/19/2025	17:38
SSTDICC010	SSTDICC010	BF142790.D	06/19/2025	18:08
SSTDICC020	SSTDICC020	BF142791.D	06/19/2025	18:39
SSTDICCC040	SSTDICCC040	BF142792.D	06/19/2025	19:09
SSTDICC050	SSTDICC050	BF142793.D	06/19/2025	19:40
SSTDICC060	SSTDICC060	BF142794.D	06/19/2025	20:10
SSTDICC080	SSTDICC080	BF142795.D	06/19/2025	20:40