

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413
Lab File ID: VY022811.D Date Analyzed: 06/25/2025
Instrument ID: MSVOA_Y Time Analyzed: 09:13
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	448549	7.71	727879	8.61	634524	11.41
UPPER LIMIT	897098	8.207	1455760	9.11	1269050	11.914
LOWER LIMIT	224275	7.207	363940	8.11	317262	10.914
EPA SAMPLE NO.						
TP-34	392242	7.71	690772	8.61	622911	11.41
TP-77	381808	7.71	681622	8.61	597468	11.41
TP-74	385627	7.71	688505	8.61	619901	11.41
TP-73	352660	7.70	632034	8.61	567782	11.41
VY0625SBL01	322245	7.71	582244	8.62	555259	11.41
VY0625SBS01	433589	7.71	722109	8.61	625216	11.41
VY0625SBSD01	424898	7.71	717544	8.61	607296	11.41

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.