

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q2413 SAS No.: Q2413 SDG NO.: Q2413
Lab File ID: VY022838.D Date Analyzed: 06/26/2025
Instrument ID: MSVOA_Y Time Analyzed: 10:07
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	469939	7.71	769006	8.62	661912	11.41
UPPER LIMIT	939878	8.207	1538010	9.116	1323820	11.914
LOWER LIMIT	234970	7.207	384503	8.116	330956	10.914
EPA SAMPLE NO.						
TP-35	306589	7.71	584214	8.62	574106	11.41
TP-72	313811	7.71	566645	8.62	546225	11.41
VY0626SBL01	328943	7.71	604959	8.62	571725	11.41
VY0626SBS01	434636	7.71	740035	8.62	637859	11.42

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.