

Duplicate Sample Summary

Client: Aramark Uniforms	SDG No.: Q2417
Project: Monthly 2025	Sample ID: LB136277BS
Client ID: LB136277BSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TPH	mg/L	+/-18	17.0		17.2		1	1.17		06/26/2025

Duplicate Sample Summary

Client: Aramark Uniforms	SDG No.: Q2417
Project: Monthly 2025	Sample ID: Q2417-02
Client ID: COMPDUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	480		485		1	1.04		06/26/2025

Duplicate Sample Summary

Client:	Aramark Uniforms	SDG No.:	Q2417
Project:	Monthly 2025	Sample ID:	Q2418-01
Client ID:	VAC-TRUCK-4074DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	153		159		1	3.59		06/26/2025