

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2417

Project: Monthly 2025

Run No.: LB136277

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136277BS							
TPH	mg/L	20.0	17.0		85	1	78-114	06/26/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2417

Project: Monthly 2025

Run No.: LB136277

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136277BSD							
TPH	mg/L	20.0	17.2		86	1	78-114	06/26/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2417

Project: Monthly 2025

Run No.: LB136278

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136278BS							
TSS	mg/L	550	533		97	1	90-110	06/26/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2417

Project: Monthly 2025

Run No.: LB136297

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136297BS							
BOD5	mg/L	198	179		91	1	84.6-115.4	06/26/2025