


LB136644
INSTRUMENT ID : CV1

245.1

Linear



A= 0.0000e+000
B= 2.6151e-004
C= 1.0185e-003
Rho= 0.9996179
Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.0	0.000	-0.009	-0.009	-40	0.000	-40					-4
0.05	0.050	0.048	-0.002	179	0.0 %	179	0				-14
0.2	0.200	0.172	-0.028	652	0.0 %	652	0				-9
2.5	2.500	2.411	-0.089	9216	0.0 %	9216					3
5.0	5.000	5.173	0.173	19777	0.0 %	19777					1
7.5	7.500	7.609	0.109	29091	0.0 %	29091					-2
10.0	10.000	9.848	-0.152	37653	0.0 %	37653					

LB136644 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 29 Jul 2025 09:36:49

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	-40	PPB	Std	0.0000	245.1	29 Jul 2025 10:21:31
S	0.05 - 1	50.05	-	179	PPB	Std	0.0500	245.1	29 Jul 2025 10:26:43
S	0.2 - 1	50.2	-	652	PPB	Std	0.2000	245.1	29 Jul 2025 10:29:00
S	2.5 - 1	52.5	-	9216	PPB	Std	2.5000	245.1	29 Jul 2025 10:31:17
S	5.0 - 1	55	-	19777	PPB	Std	5.0000	245.1	29 Jul 2025 10:33:34
S	7.5 - 1	57.5	-	29091	PPB	Std	7.5000	245.1	29 Jul 2025 10:35:51
S	10.0 - 1	57.0	-	37653	PPB	Std	10.0000	245.1	29 Jul 2025 10:37:08
U	ICV42 - 1	ICV42	4.0094	15328	PPB	SMPL	-245.1		29 Jul 2025 10:43:25
U	ICB42 - 1	ICB42	-0.0913	-353	PPB	SMPL	-245.1		29 Jul 2025 10:45:40
U	CCV57 - 1	CCV57	5.1153	19557	PPB	SMPL	-245.1		29 Jul 2025 10:47:57
U	CCB57 - 1	CCB57	-0.0631	-245	PPB	SMPL	-245.1		29 Jul 2025 10:50:12
U	CRA - 1	CRA	0.0392	146	PPB	SMPL	-245.1		29 Jul 2025 10:55:07
U	HighStd - 1	HighStd	10.2545	39209	PPB	SMPL	-245.1		29 Jul 2025 10:57:22
U	ChkStd - 1	ChkStd	7.1763	27438	PPB	SMPL	-245.1		29 Jul 2025 10:59:38
U	PB169046BL - 1	PBW	0.0240	88	PPB	SMPL	-245.1		29 Jul 2025 11:04:40
U	PB169046BS - 1	LCSW	3.7641	14390	PPB	SMPL	-245.1		29 Jul 2025 11:06:56
U	Q2417-02 - 1	COMP	0.0447	167	PPB	SMPL	-245.1		29 Jul 2025 11:09:12
U	Q2417-02DUP - 1	COMPDUP	0.0708	267	PPB	SMPL	-245.1		29 Jul 2025 11:11:30
U	Q2417-02MS - 1	COMPMS	4.6093	17622	PPB	SMPL	-245.1		29 Jul 2025 11:24:06
U	Q2417-02MSD - 1	COMPMSD	4.2926	16411	PPB	SMPL	-245.1		29 Jul 2025 11:26:23
U	CCV58 - 1	CCV58	5.2129	19930	PPB	SMPL	-245.1		29 Jul 2025 11:32:06
U	CCB58 - 1	CCB58	-0.0772	-299	PPB	SMPL	-245.1		29 Jul 2025 11:34:20
U	Q2417-02LX5 - 1		-0.0006	-6	PPB	SMPL	-245.1		29 Jul 2025 11:36:40
U	Q2417-02A - 1		4.5751	17491	PPB	SMPL	-245.1		29 Jul 2025 11:38:54
U	CCV59 - 1	CCV59	4.9558	18947	PPB	SMPL	-245.1		29 Jul 2025 11:47:01
U	CCB59 - 1	CCB59	-0.0892	-345	PPB	SMPL	-245.1		29 Jul 2025 11:49:16