

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q2417

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV42	Mercury	4.01	4.0	100	95 - 105	CV	07/29/2025	10:43	LB136644

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q2417

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV57	Mercury	5.12	5.0	102	90 - 110	CV	07/29/2025	10:47	LB136644
CCV58	Mercury	5.21	5.0	104	90 - 110	CV	07/29/2025	11:32	LB136644
CCV59	Mercury	4.96	5.0	99	90 - 110	CV	07/29/2025	11:47	LB136644

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q2417

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Cadmium	2050	2000	103	95 - 105	P	07/29/2025	12:20	LB136648
	Copper	1020	1000	102	95 - 105	P	07/29/2025	12:20	LB136648
	Lead	4000	4000	100	95 - 105	P	07/29/2025	12:20	LB136648
	Nickel	1990	2000	100	95 - 105	P	07/29/2025	12:20	LB136648
	Zinc	2060	2000	103	95 - 105	P	07/29/2025	12:20	LB136648

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q2417

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Cadmium	5.16	6.0	86	80 - 120	P	07/29/2025	13:01	LB136648
	Copper	21.9	20.0	110	80 - 120	P	07/29/2025	13:01	LB136648
	Lead	12.1	12.0	101	80 - 120	P	07/29/2025	13:01	LB136648
	Nickel	40.3	40.0	101	80 - 120	P	07/29/2025	13:01	LB136648
	Zinc	41.8	40.0	105	80 - 120	P	07/29/2025	13:01	LB136648

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q2417

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Cadmium	2450	2500	98	90 - 110	P	07/29/2025	14:21	LB136648
	Copper	1270	1250	102	90 - 110	P	07/29/2025	14:21	LB136648
	Lead	4940	5000	99	90 - 110	P	07/29/2025	14:21	LB136648
	Nickel	2500	2500	100	90 - 110	P	07/29/2025	14:21	LB136648
	Zinc	2620	2500	105	90 - 110	P	07/29/2025	14:21	LB136648
CCV02	Cadmium	2420	2500	97	90 - 110	P	07/29/2025	15:11	LB136648
	Copper	1280	1250	103	90 - 110	P	07/29/2025	15:11	LB136648
	Lead	5000	5000	100	90 - 110	P	07/29/2025	15:11	LB136648
	Nickel	2540	2500	101	90 - 110	P	07/29/2025	15:11	LB136648
	Zinc	2620	2500	105	90 - 110	P	07/29/2025	15:11	LB136648
CCV03	Cadmium	2290	2500	91	90 - 110	P	07/29/2025	16:14	LB136648
	Copper	1270	1250	102	90 - 110	P	07/29/2025	16:14	LB136648
	Lead	4860	5000	97	90 - 110	P	07/29/2025	16:14	LB136648
	Nickel	2480	2500	99	90 - 110	P	07/29/2025	16:14	LB136648
	Zinc	2620	2500	105	90 - 110	P	07/29/2025	16:14	LB136648
CCV04	Cadmium	2270	2500	91	90 - 110	P	07/29/2025	17:20	LB136648
	Copper	1300	1250	104	90 - 110	P	07/29/2025	17:20	LB136648
	Lead	4970	5000	99	90 - 110	P	07/29/2025	17:20	LB136648
	Nickel	2550	2500	102	90 - 110	P	07/29/2025	17:20	LB136648
	Zinc	2650	2500	106	90 - 110	P	07/29/2025	17:20	LB136648
CCV05	Cadmium	2320	2500	93	90 - 110	P	07/29/2025	18:11	LB136648
	Copper	1290	1250	104	90 - 110	P	07/29/2025	18:11	LB136648
	Lead	4990	5000	100	90 - 110	P	07/29/2025	18:11	LB136648
	Nickel	2550	2500	102	90 - 110	P	07/29/2025	18:11	LB136648
	Zinc	2640	2500	106	90 - 110	P	07/29/2025	18:11	LB136648
CCV06	Cadmium	2320	2500	93	90 - 110	P	07/29/2025	18:31	LB136648
	Copper	1280	1250	103	90 - 110	P	07/29/2025	18:31	LB136648
	Lead	4950	5000	99	90 - 110	P	07/29/2025	18:31	LB136648
	Nickel	2530	2500	101	90 - 110	P	07/29/2025	18:31	LB136648
	Zinc	2650	2500	106	90 - 110	P	07/29/2025	18:31	LB136648