



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Aramark Uniforms

SDG No.: Q2417

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.039	0.05	78	60 - 140	CV	07/29/2025	10:55	LB136644
CRI01	Cadmium	5.14	6.0	86	65 - 135	P	07/29/2025	13:09	LB136648
	Copper	21.1	20.0	106	65 - 135	P	07/29/2025	13:09	LB136648
	Lead	11.5	12.0	96	65 - 135	P	07/29/2025	13:09	LB136648
	Nickel	39.0	40.0	98	65 - 135	P	07/29/2025	13:09	LB136648
	Zinc	42.4	40.0	106	65 - 135	P	07/29/2025	13:09	LB136648