

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Aramark Uniforms</u>	SDG No.: <u>Q2417</u>
Contract: <u>ARAM01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	4.35	1.0	435	-5	7	07/29/2025	13:14	LB136648
	Copper	21.7	2.0	1085	-18	22	07/29/2025	13:14	LB136648
	Lead	-4.04			-12	12	07/29/2025	13:14	LB136648
	Nickel	3.75	2.0	188	-38	42	07/29/2025	13:14	LB136648
	Zinc	1.98			-40	40	07/29/2025	13:14	LB136648
ICSAB01	Cadmium	854	972	88	826	1120	07/29/2025	13:20	LB136648
	Copper	505	511	99	434	588	07/29/2025	13:20	LB136648
	Lead	43.2	49.0	88	37	61	07/29/2025	13:20	LB136648
	Nickel	977	954	102	810	1100	07/29/2025	13:20	LB136648
	Zinc	1050	952	110	809	1095	07/29/2025	13:20	LB136648
ICSA	Cadmium	2.00	1.0	200	-5	7	07/29/2025	13:24	LB136648
	Copper	21.7	2.0	1085	-18	22	07/29/2025	13:24	LB136648
	Lead	-15.8			-12	12	07/29/2025	13:24	LB136648
	Nickel	11.1	2.0	555	-38	42	07/29/2025	13:24	LB136648
	Zinc	1.19			-40	40	07/29/2025	13:24	LB136648
ICSAB	Cadmium	845	972	87	826	1120	07/29/2025	13:28	LB136648
	Copper	532	511	104	434	588	07/29/2025	13:28	LB136648
	Lead	38.0	49.0	78	37	61	07/29/2025	13:28	LB136648
	Nickel	983	954	103	810	1100	07/29/2025	13:28	LB136648
	Zinc	1020	952	107	809	1095	07/29/2025	13:28	LB136648