

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Aramark Uniforms</u>	level: <u>low</u>	sdg no.: <u>Q2417</u>
contract: <u>ARAM01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2417-02</u>	client id: <u>COMPMS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2417-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	75 - 125	74.3	D	1.92	JD	100	72	N	P
Copper	ug/L	75 - 125	332	D	148	D	150	123		P
Lead	ug/L	75 - 125	464	D	19.4	JD	500	89		P
Mercury	ug/L	75 - 125	4.61		0.045	J	4.0	114		CV
Nickel	ug/L	75 - 125	262	D	20.8	JD	250	97		P
Zinc	ug/L	75 - 125	651	D	505	D	100	146		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Aramark Uniforms</u>	level: <u>low</u>	sdg no.: <u>Q2417</u>
contract: <u>ARAM01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2417-02</u>	client id: <u>COMPMSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2417-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	75 - 125	75.5	D	1.92	JD	100	74	N	P
Copper	ug/L	75 - 125	339	D	148	D	150	127	N	P
Lead	ug/L	75 - 125	470	D	19.4	JD	500	90		P
Mercury	ug/L	75 - 125	4.29		0.045	J	4.0	106		CV
Nickel	ug/L	75 - 125	312	D	20.8	JD	250	116		P
Zinc	ug/L	75 - 125	649	D	505	D	100	144		P