

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Aramark Uniforms</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2417</u>
Contract:	<u>ARAM01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2417-02</u>	Client ID:	<u>COMPDUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2417-02DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	1.92	JD	15.0	UD	200		P
Copper	ug/L	20	148	D	156	D	5		P
Lead	ug/L	20	19.4	JD	19.9	JD	3		P
Mercury	ug/L	20	0.045	J	0.071	J	45		CV
Nickel	ug/L	20	20.8	JD	25.5	JD	20		P
Zinc	ug/L	20	505	D	530	D	5		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Aramark Uniforms</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2417</u>
Contract:	<u>ARAM01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2417-02MS</u>	Client ID:	<u>COMPMSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2417-02MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	74.3	D	75.5	D	2		P
Copper	ug/L	20	332	D	339	D	2		P
Lead	ug/L	20	464	D	470	D	1		P
Mercury	ug/L	20	4.61		4.29		7		CV
Nickel	ug/L	20	262	D	312	D	17		P
Zinc	ug/L	20	651	D	649	D	0		P