

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Aramark Uniforms

SDG No.: Q2417

Contract: ARAM01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169046BS Mercury	ug/L	4.0	3.76		94	85 - 115	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Aramark Uniforms

SDG No.: Q2417

Contract: ARAM01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169048BS							
Cadmium	ug/L	100	86.8		87	85 - 115	P
Copper	ug/L	150	164		109	85 - 115	P
Lead	ug/L	500	480		96	85 - 115	P
Nickel	ug/L	250	258		103	85 - 115	P
Zinc	ug/L	100	113		113	85 - 115	P