

LAB CHRONICLE

OrderID:	Q2425	OrderDate:	6/25/2025 2:03:00 PM
Client:	Kleinfelder	Project:	AS Jenks School
Contact:	Mark Warchol	Location:	D51,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2425-01	COMP-1	SOIL	VOCMS Group1	8260D	06/24/25		06/25/25	06/25/25
Q2425-02	COMP-2	SOIL	VOCMS Group1	8260D	06/24/25		06/25/25	06/25/25
Q2425-03	COMP-3	SOIL	VOCMS Group1	8260D	06/24/25		06/25/25	06/25/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2425
Client: Kleinfelder

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:				0				
Total Voc :								
Total Concentration:								



SAMPLE DATA

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-1	SDG No.:	Q2425
Lab Sample ID:	Q2425-01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	76.9
Sample Wt/Vol:	4.5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022831.D	1		06/25/25 17:39	VY062525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
156-59-2	cis-1,2-Dichloroethene	1.10	U	1.10	7.20	ug/Kg
71-55-6	1,1,1-Trichloroethane	1.30	U	1.30	7.20	ug/Kg
71-43-2	Benzene	1.10	U	1.10	7.20	ug/Kg
79-01-6	Trichloroethene	1.20	U	1.20	7.20	ug/Kg
108-88-3	Toluene	1.10	U	1.10	7.20	ug/Kg
100-41-4	Ethyl Benzene	0.97	U	0.97	7.20	ug/Kg
1330-20-7	Total Xylenes	3.00	U	3.00	21.6	ug/Kg
98-82-8	Isopropylbenzene	1.10	U	1.10	7.20	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	48.5		63 - 155	97%	SPK: 50
1868-53-7	Dibromofluoromethane	49.8		70 - 134	100%	SPK: 50
2037-26-5	Toluene-d8	50.1		74 - 123	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.2		17 - 146	100%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	439000	7.707			
540-36-3	1,4-Difluorobenzene	824000	8.609			
3114-55-4	Chlorobenzene-d5	766000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	315000	13.34			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-2	SDG No.:	Q2425
Lab Sample ID:	Q2425-02	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	79.7
Sample Wt/Vol:	4.7 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022832.D	1		06/25/25 18:02	VY062525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
156-59-2	cis-1,2-Dichloroethene	1.00	U	1.00	6.70	ug/Kg
71-55-6	1,1,1-Trichloroethane	1.20	U	1.20	6.70	ug/Kg
71-43-2	Benzene	1.10	U	1.10	6.70	ug/Kg
79-01-6	Trichloroethene	1.10	U	1.10	6.70	ug/Kg
108-88-3	Toluene	1.00	U	1.00	6.70	ug/Kg
100-41-4	Ethyl Benzene	0.89	U	0.89	6.70	ug/Kg
1330-20-7	Total Xylenes	2.80	U	2.80	20.0	ug/Kg
98-82-8	Isopropylbenzene	1.00	U	1.00	6.70	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	46.2		63 - 155	92%	SPK: 50
1868-53-7	Dibromofluoromethane	49.4		70 - 134	99%	SPK: 50
2037-26-5	Toluene-d8	49.6		74 - 123	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.2		17 - 146	100%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	436000	7.707			
540-36-3	1,4-Difluorobenzene	783000	8.61			
3114-55-4	Chlorobenzene-d5	713000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	306000	13.34			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-3	SDG No.:	Q2425
Lab Sample ID:	Q2425-03	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	82.6
Sample Wt/Vol:	5.6 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022833.D	1		06/25/25 18:26	VY062525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
156-59-2	cis-1,2-Dichloroethene	0.81	U	0.81	5.40	ug/Kg
71-55-6	1,1,1-Trichloroethane	1.00	U	1.00	5.40	ug/Kg
71-43-2	Benzene	0.85	U	0.85	5.40	ug/Kg
79-01-6	Trichloroethene	0.88	U	0.88	5.40	ug/Kg
108-88-3	Toluene	0.84	U	0.84	5.40	ug/Kg
100-41-4	Ethyl Benzene	0.72	U	0.72	5.40	ug/Kg
1330-20-7	Total Xylenes	2.19	U	2.19	16.2	ug/Kg
98-82-8	Isopropylbenzene	0.84	U	0.84	5.40	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	45.7		63 - 155	91%	SPK: 50
1868-53-7	Dibromofluoromethane	49.5		70 - 134	99%	SPK: 50
2037-26-5	Toluene-d8	49.8		74 - 123	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.5		17 - 146	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	401000	7.707			
540-36-3	1,4-Difluorobenzene	713000	8.609			
3114-55-4	Chlorobenzene-d5	667000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	285000	13.34			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: Q2425

Client: Kleinfelder

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2425-01	COMP-1	1,2-Dichloroethane-d4	50	48.5	97	63	155
		Dibromofluoromethane	50	49.8	100	70	134
		Toluene-d8	50	50.0	100	74	123
		4-Bromofluorobenzene	50	50.2	100	17	146
Q2425-02	COMP-2	1,2-Dichloroethane-d4	50	46.2	92	63	155
		Dibromofluoromethane	50	49.4	99	70	134
		Toluene-d8	50	49.6	99	74	123
		4-Bromofluorobenzene	50	50.2	100	17	146
Q2425-03	COMP-3	1,2-Dichloroethane-d4	50	45.7	91	63	155
		Dibromofluoromethane	50	49.5	99	70	134
		Toluene-d8	50	49.8	100	74	123
		4-Bromofluorobenzene	50	51.5	103	17	146
VY0625SBL01	VY0625SBL01	1,2-Dichloroethane-d4	50	48.2	96	63	155
		Dibromofluoromethane	50	49.8	100	70	134
		Toluene-d8	50	49.3	99	74	123
		4-Bromofluorobenzene	50	53.3	107	17	146
VY0625SBS01	VY0625SBS01	1,2-Dichloroethane-d4	50	50.1	100	63	155
		Dibromofluoromethane	50	50.1	100	70	134
		Toluene-d8	50	51.1	102	74	123
		4-Bromofluorobenzene	50	49.7	99	17	146
VY0625SBSD01	VY0625SBSD01	1,2-Dichloroethane-d4	50	51.0	102	63	155
		Dibromofluoromethane	50	51.1	102	70	134
		Toluene-d8	50	51.5	103	74	123
		4-Bromofluorobenzene	50	49.5	99	17	146

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2425

Client: Kleinfelder

Analytical Method: SW8260D

Datafile : VY022813.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VY0625SBS01	cis-1,2-Dichloroethene	20	20.1	ug/Kg	101			82	123	
	1,1,1-Trichloroethane	20	20.2	ug/Kg	101			80	126	
	Benzene	20	20.1	ug/Kg	101			84	121	
	Trichloroethene	20	20.9	ug/Kg	104			83	122	
	Toluene	20	20.2	ug/Kg	101			83	122	
	Ethyl Benzene	20	20.1	ug/Kg	101			82	124	
	m/p-Xylenes	40	39.8	ug/Kg	100			83	124	
	o-Xylene	20	19.7	ug/Kg	99			83	123	
	Isopropylbenzene	20	20.3	ug/Kg	102			82	124	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2425

Client: Kleinfelder

Analytical Method: SW8260D

Datafile : VY022814.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VY0625SBSD01	cis-1,2-Dichloroethene	20	20.4	ug/Kg	102	1		82	123	20
	1,1,1-Trichloroethane	20	20.5	ug/Kg	103	2		80	126	20
	Benzene	20	20.4	ug/Kg	102	1		84	121	20
	Trichloroethene	20	20.4	ug/Kg	102	2		83	122	20
	Toluene	20	20.0	ug/Kg	100	1		83	122	20
	Ethyl Benzene	20	20.6	ug/Kg	103	2		82	124	20
	m/p-Xylenes	40	41.0	ug/Kg	103	3		83	124	20
	o-Xylene	20	20.2	ug/Kg	101	2		83	123	20
	Isopropylbenzene	20	21.3	ug/Kg	106	4		82	124	20

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VY0625SBL01

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM Case No.: Q2425

SAS No.: Q2425 SDG NO.: Q2425

Lab File ID: VY022812.D

Lab Sample ID: VY0625SBL01

Date Analyzed: 06/25/2025

Time Analyzed: 09:46

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) Y

Instrument ID: MSVOA_Y

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VY0625SBS01	VY0625SBS01	VY022813.D	06/25/2025
VY0625SBSD01	VY0625SBSD01	VY022814.D	06/25/2025
COMP-1	Q2425-01	VY022831.D	06/25/2025
COMP-2	Q2425-02	VY022832.D	06/25/2025
COMP-3	Q2425-03	VY022833.D	06/25/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: POWE02
Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425
Lab File ID: VY022775.D BFB Injection Date: 06/23/2025
Instrument ID: MSVOA_Y BFB Injection Time: 10:17
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.8
75	30.0 - 60.0% of mass 95	56.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.9 (1.1) 1
174	50.0 - 100.0% of mass 95	81.9
175	5.0 - 9.0% of mass 174	6 (7.4) 1
176	95.0 - 101.0% of mass 174	78.2 (95.5) 1
177	5.0 - 9.0% of mass 176	5.1 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VY022776.D	06/23/2025	13:38
VSTDIC010	VSTDIC010	VY022777.D	06/23/2025	14:00
VSTDIC020	VSTDIC020	VY022778.D	06/23/2025	14:23
VSTDIC050	VSTDIC050	VY022779.D	06/23/2025	14:46
VSTDIC100	VSTDIC100	VY022780.D	06/23/2025	15:08
VSTDIC150	VSTDIC150	VY022781.D	06/23/2025	15:31

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: POWE02
Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425
Lab File ID: VY022810.D BFB Injection Date: 06/25/2025
Instrument ID: MSVOA_Y BFB Injection Time: 08:41
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23
75	30.0 - 60.0% of mass 95	56.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.9
173	Less than 2.0% of mass 174	0.9 (1.2) 1
174	50.0 - 100.0% of mass 95	80.5
175	5.0 - 9.0% of mass 174	6 (7.5) 1
176	95.0 - 101.0% of mass 174	79.5 (98.8) 1
177	5.0 - 9.0% of mass 176	5.2 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY022811.D	06/25/2025	09:13
VY0625SBL01	VY0625SBL01	VY022812.D	06/25/2025	09:46
VY0625SBS01	VY0625SBS01	VY022813.D	06/25/2025	10:18
VY0625SBSD01	VY0625SBSD01	VY022814.D	06/25/2025	10:40
COMP-1	Q2425-01	VY022831.D	06/25/2025	17:39
COMP-2	Q2425-02	VY022832.D	06/25/2025	18:02
COMP-3	Q2425-03	VY022833.D	06/25/2025	18:26

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: POWE02
Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425
Lab File ID: VY022811.D Date Analyzed: 06/25/2025
Instrument ID: MSVOA_Y Time Analyzed: 09:13
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	448549	7.71	727879	8.61	634524	11.41
UPPER LIMIT	897098	8.207	1455760	9.11	1269050	11.914
LOWER LIMIT	224275	7.207	363940	8.11	317262	10.914
EPA SAMPLE NO.						
COMP-1	439139	7.71	824260	8.61	766377	11.41
COMP-2	436434	7.71	782882	8.61	712732	11.41
COMP-3	401005	7.71	712820	8.61	666622	11.41
VY0625SBL01	322245	7.71	582244	8.62	555259	11.41
VY0625SBS01	433589	7.71	722109	8.61	625216	11.41
VY0625SBSD01	424898	7.71	717544	8.61	607296	11.41

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: POWE02
 Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425
 Lab File ID: VY022811.D Date Analyzed: 06/25/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:13
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	316858	13.34				
UPPER LIMIT	633716	13.84				
LOWER LIMIT	158429	12.84				
EPA SAMPLE NO.						
COMP-1	315122	13.34				
COMP-2	306171	13.34				
COMP-3	284845	13.34				
VY0625SBL01	248657	13.35				
VY0625SBS01	303331	13.35				
VY0625SBSD01	290027	13.34				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Kleinfelder		Date Collected:	
Project:	AS Jenks School		Date Received:	
Client Sample ID:	VY0625SBL01		SDG No.:	Q2425
Lab Sample ID:	VY0625SBL01		Matrix:	SOIL
Analytical Method:	8260D		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022812.D	1		06/25/25 09:46	VY062525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.75	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.93	U	0.93	5.00	ug/Kg
71-43-2	Benzene	0.79	U	0.79	5.00	ug/Kg
79-01-6	Trichloroethene	0.81	U	0.81	5.00	ug/Kg
108-88-3	Toluene	0.78	U	0.78	5.00	ug/Kg
100-41-4	Ethyl Benzene	0.67	U	0.67	5.00	ug/Kg
1330-20-7	Total Xylenes	2.02	U	2.02	15.0	ug/Kg
98-82-8	Isopropylbenzene	0.78	U	0.78	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	48.2		63 - 155	96%	SPK: 50
1868-53-7	Dibromofluoromethane	49.8		70 - 134	100%	SPK: 50
2037-26-5	Toluene-d8	49.3		74 - 123	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.3		17 - 146	107%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	322000	7.707			
540-36-3	1,4-Difluorobenzene	582000	8.616			
3114-55-4	Chlorobenzene-d5	555000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	249000	13.346			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	AS Jenks School	Date Received:	
Client Sample ID:	VY0625SBS01	SDG No.:	Q2425
Lab Sample ID:	VY0625SBS01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022813.D	1		06/25/25 10:18	VY062525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
156-59-2	cis-1,2-Dichloroethene	20.1		0.75	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	20.2		0.93	5.00	ug/Kg
71-43-2	Benzene	20.1		0.79	5.00	ug/Kg
79-01-6	Trichloroethene	20.9		0.81	5.00	ug/Kg
108-88-3	Toluene	20.2		0.78	5.00	ug/Kg
100-41-4	Ethyl Benzene	20.1		0.67	5.00	ug/Kg
1330-20-7	Total Xylenes	59.5		2.02	15.0	ug/Kg
98-82-8	Isopropylbenzene	20.3		0.78	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	50.1		63 - 155	100%	SPK: 50
1868-53-7	Dibromofluoromethane	50.1		70 - 134	100%	SPK: 50
2037-26-5	Toluene-d8	51.1		74 - 123	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.7		17 - 146	99%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	434000	7.707			
540-36-3	1,4-Difluorobenzene	722000	8.61			
3114-55-4	Chlorobenzene-d5	625000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	303000	13.346			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	AS Jenks School	Date Received:	
Client Sample ID:	VY0625SBSD01	SDG No.:	Q2425
Lab Sample ID:	VY0625SBSD01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022814.D	1		06/25/25 10:40	VY062525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
156-59-2	cis-1,2-Dichloroethene	20.4		0.75	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	20.5		0.93	5.00	ug/Kg
71-43-2	Benzene	20.4		0.79	5.00	ug/Kg
79-01-6	Trichloroethene	20.4		0.81	5.00	ug/Kg
108-88-3	Toluene	20.0		0.78	5.00	ug/Kg
100-41-4	Ethyl Benzene	20.6		0.67	5.00	ug/Kg
1330-20-7	Total Xylenes	61.2		2.02	15.0	ug/Kg
98-82-8	Isopropylbenzene	21.3		0.78	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.0		63 - 155	102%	SPK: 50
1868-53-7	Dibromofluoromethane	51.1		70 - 134	102%	SPK: 50
2037-26-5	Toluene-d8	51.5		74 - 123	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.5		17 - 146	99%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	425000	7.707			
540-36-3	1,4-Difluorobenzene	718000	8.61			
3114-55-4	Chlorobenzene-d5	607000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	290000	13.34			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>POWE02</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2425</u>
Instrument ID:	<u>MSVOA_Y</u>	SAS No.:	<u>Q2425</u>
Heated Purge: (Y/N)	<u>Y</u>	SDG No.:	<u>Q2425</u>
GC Column:	<u>RXI-624</u>	Calibration Date(s):	<u>06/23/2025</u>
ID:	<u>0.25 (mm)</u>	Calibration Time(s):	<u>13:38</u>
			<u>15:31</u>

LAB FILE ID:		RRF005 = VY022776.D		RRF010 = VY022777.D		RRF020 = VY022778.D		
		RRF050 = VY022779.D		RRF100 = VY022780.D		RRF150 = VY022781.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
cis-1,2-Dichloroethene	0.606	0.689	0.687	0.685	0.687	0.678	0.672	4.8
1,1,1-Trichloroethane	0.847	0.945	0.973	0.950	0.939	0.923	0.929	4.7
Benzene	1.248	1.433	1.451	1.464	1.467	1.440	1.417	5.9
Trichloroethene	0.305	0.364	0.382	0.372	0.360	0.350	0.356	7.6
Toluene	0.747	0.873	0.908	0.926	0.955	0.954	0.894	8.8
Ethyl Benzene	1.644	1.881	1.971	2.029	2.040	2.018	1.930	7.9
m/p-Xylenes	0.624	0.722	0.759	0.782	0.800	0.791	0.746	8.8
o-Xylene	0.578	0.674	0.708	0.734	0.759	0.765	0.703	10
Isopropylbenzene	3.354	3.764	3.823	3.778	3.709	3.759	3.698	4.7
1,2-Dichloroethane-d4	0.568	0.550	0.557	0.559	0.571	0.545	0.558	1.8
Dibromofluoromethane	0.306	0.297	0.295	0.304	0.314	0.308	0.304	2.3
Toluene-d8	1.182	1.148	1.186	1.215	1.262	1.247	1.207	3.6
4-Bromofluorobenzene	0.368	0.362	0.370	0.385	0.423	0.421	0.388	7

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG No.: Q2425

Instrument ID: MSVOA_Y Calibration Date/Time: 06/25/2025 09:13

Lab File ID: VY022811.D Init. Calib. Date(s): 06/23/2025 06/23/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:38 15:31

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
cis-1,2-Dichloroethene	0.672	0.648		-3.57	20
1,1,1-Trichloroethane	0.929	0.916		-1.4	20
Benzene	1.417	1.415		-0.14	20
Trichloroethene	0.356	0.364		2.25	20
Toluene	0.894	0.909		1.68	20
Ethyl Benzene	1.930	1.981		2.64	20
m/p-Xylenes	0.746	0.775		3.89	20
o-Xylene	0.703	0.725		3.13	20
Isopropylbenzene	3.698	3.805		2.89	20
1,2-Dichloroethane-d4	0.558	0.536		-3.94	20
Dibromofluoromethane	0.304	0.309		1.64	20
Toluene-d8	1.207	1.256		4.06	20
4-Bromofluorobenzene	0.388	0.398		2.58	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

LAB CHRONICLE

OrderID:	Q2425	OrderDate:	6/25/2025 2:03:00 PM
Client:	Kleinfelder	Project:	AS Jenks School
Contact:	Mark Warchol	Location:	D51,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2425-01	COMP-1	SOIL	SVOCMS Group1	8270E	06/24/25	06/26/25	06/26/25	06/25/25
Q2425-02	COMP-2	SOIL	SVOCMS Group1	8270E	06/24/25	06/26/25	06/27/25	06/25/25
Q2425-03	COMP-3	SOIL	SVOCMS Group1	8270E	06/24/25	06/26/25	06/27/25	06/25/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2425
Client: Kleinfelder

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :				0.000				
Total Svoc :					0.00			
Total Concentration:					0.00			



SAMPLE DATA

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-1	SDG No.:	Q2425
Lab Sample ID:	Q2425-01	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	76.9
Sample Wt/Vol:	30.05	Units:	g
Soil Aliquot Vol:			uL
Extraction Type :		Decanted :	N
Injection Volume :		GPC Factor :	1.0
Prep Method :	SW3541	Level :	LOW
		Test:	SVOCMS Group1
		Final Vol:	1000
		GPC Cleanup :	N
		PH :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142874.D	1	06/26/25 09:20	06/26/25 18:05	PB168625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	29.5	U	29.5	220	ug/Kg
86-73-7	Fluorene	32.8	U	32.8	220	ug/Kg
85-01-8	Phenanthrene	27.1	U	27.1	220	ug/Kg
120-12-7	Anthracene	43.2	U	43.2	220	ug/Kg
129-00-0	Pyrene	46.7	U	46.7	220	ug/Kg
56-55-3	Benzo(a)anthracene	29.9	U	29.9	220	ug/Kg
218-01-9	Chrysene	25.8	U	25.8	220	ug/Kg
205-99-2	Benzo(b)fluoranthene	24.7	U	24.7	220	ug/Kg
50-32-8	Benzo(a)pyrene	38.3	U	38.3	220	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	37.8	U	37.8	220	ug/Kg
191-24-2	Benzo(g,h,i)perylene	33.4	U	33.4	220	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	50.4		18 - 107	50%	SPK: 100
321-60-8	2-Fluorobiphenyl	51.1		20 - 109	51%	SPK: 100
1718-51-0	Terphenyl-d14	50.4		10 - 105	50%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	80000	6.875			
1146-65-2	Naphthalene-d8	306000	8.157			
15067-26-2	Acenaphthene-d10	164000	9.916			
1517-22-2	Phenanthrene-d10	267000	11.404			
1719-03-5	Chrysene-d12	136000	14.051			
1520-96-3	Perylene-d12	164000	15.539			

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-1	SDG No.:	Q2425
Lab Sample ID:	Q2425-01	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	76.9
Sample Wt/Vol:	30.05	Units:	g
Soil Aliquot Vol:			uL
Extraction Type :		Decanted :	N
Injection Volume :		GPC Factor :	1.0
Prep Method :	SW3541	Level :	LOW
		Final Vol:	1000 uL
		Test:	SVOCMS Group1
		GPC Cleanup :	N
		PH :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142874.D	1	06/26/25 09:20	06/26/25 18:05	PB168625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-2	SDG No.:	Q2425
Lab Sample ID:	Q2425-02	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	79.7
Sample Wt/Vol:	30.07 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142887.D	1	06/26/25 09:20	06/27/25 12:21	PB168625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	28.4	U	28.4	210	ug/Kg
86-73-7	Fluorene	31.7	U	31.7	210	ug/Kg
85-01-8	Phenanthrene	26.2	U	26.2	210	ug/Kg
120-12-7	Anthracene	41.7	U	41.7	210	ug/Kg
129-00-0	Pyrene	45.1	U	45.1	210	ug/Kg
56-55-3	Benzo(a)anthracene	28.8	U	28.8	210	ug/Kg
218-01-9	Chrysene	24.9	U	24.9	210	ug/Kg
205-99-2	Benzo(b)fluoranthene	23.8	U	23.8	210	ug/Kg
50-32-8	Benzo(a)pyrene	36.9	U	36.9	210	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	36.4	U	36.4	210	ug/Kg
191-24-2	Benzo(g,h,i)perylene	32.2	U	32.2	210	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	83.7		18 - 107	84%	SPK: 100
321-60-8	2-Fluorobiphenyl	79.9		20 - 109	80%	SPK: 100
1718-51-0	Terphenyl-d14	63.3		10 - 105	63%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	69200	6.875			
1146-65-2	Naphthalene-d8	263000	8.157			
15067-26-2	Acenaphthene-d10	136000	9.916			
1517-22-2	Phenanthrene-d10	203000	11.404			
1719-03-5	Chrysene-d12	150000	14.051			
1520-96-3	Perylene-d12	149000	15.545			

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-2	SDG No.:	Q2425
Lab Sample ID:	Q2425-02	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	79.7
Sample Wt/Vol:	30.07	Units:	g
Soil Aliquot Vol:			uL
Extraction Type :		Decanted :	N
Injection Volume :		GPC Factor :	1.0
Prep Method :	SW3541	Level :	LOW
		Final Vol:	1000 uL
		Test:	SVOCMS Group1
		GPC Cleanup :	N
		PH :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142887.D	1	06/26/25 09:20	06/27/25 12:21	PB168625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-3	SDG No.:	Q2425
Lab Sample ID:	Q2425-03	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	82.6
Sample Wt/Vol:	30.02 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142888.D	1	06/26/25 09:20	06/27/25 12:50	PB168625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	27.5	U	27.5	210	ug/Kg
86-73-7	Fluorene	30.6	U	30.6	210	ug/Kg
85-01-8	Phenanthrene	25.3	U	25.3	210	ug/Kg
120-12-7	Anthracene	40.3	U	40.3	210	ug/Kg
129-00-0	Pyrene	43.6	U	43.6	210	ug/Kg
56-55-3	Benzo(a)anthracene	27.8	U	27.8	210	ug/Kg
218-01-9	Chrysene	24.1	U	24.1	210	ug/Kg
205-99-2	Benzo(b)fluoranthene	23.0	U	23.0	210	ug/Kg
50-32-8	Benzo(a)pyrene	35.7	U	35.7	210	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	35.2	U	35.2	210	ug/Kg
191-24-2	Benzo(g,h,i)perylene	31.1	U	31.1	210	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	49.2		18 - 107	49%	SPK: 100
321-60-8	2-Fluorobiphenyl	52.4		20 - 109	52%	SPK: 100
1718-51-0	Terphenyl-d14	40.1		10 - 105	40%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	69900	6.875			
1146-65-2	Naphthalene-d8	264000	8.157			
15067-26-2	Acenaphthene-d10	130000	9.916			
1517-22-2	Phenanthrene-d10	197000	11.404			
1719-03-5	Chrysene-d12	148000	14.051			
1520-96-3	Perylene-d12	148000	15.545			

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-3	SDG No.:	Q2425
Lab Sample ID:	Q2425-03	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	82.6
Sample Wt/Vol:	30.02	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	SVOCMS Group1
	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	1.0	GPC Cleanup :
		N	PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142888.D	1	06/26/25 09:20	06/27/25 12:50	PB168625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q2425

Client: Kleinfelder

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168625BL	PB168625BL	Nitrobenzene-d5	100	81.6	82		18	107
		2-Fluorobiphenyl	100	77.4	77		20	109
		Terphenyl-d14	100	78.5	78		10	105
PB168625BS	PB168625BS	Nitrobenzene-d5	100	77.0	77		18	107
		2-Fluorobiphenyl	100	74.7	75		20	109
		Terphenyl-d14	100	78.2	78		10	105
Q2416-01MS	MH-G/HMS	Nitrobenzene-d5	100	45.0	45		18	107
		2-Fluorobiphenyl	100	46.3	46		20	109
		Terphenyl-d14	100	45.4	45		10	105
Q2416-01MSD	MH-G/HMSD	Nitrobenzene-d5	100	46.2	46		18	107
		2-Fluorobiphenyl	100	46.1	46		20	109
		Terphenyl-d14	100	45.3	45		10	105
Q2425-01	COMP-1	Nitrobenzene-d5	100	50.4	50		18	107
		2-Fluorobiphenyl	100	51.1	51		20	109
		Terphenyl-d14	100	50.4	50		10	105
Q2425-02	COMP-2	Nitrobenzene-d5	100	83.7	84		18	107
		2-Fluorobiphenyl	100	79.9	80		20	109
		Terphenyl-d14	100	63.3	63		10	105
Q2425-03	COMP-3	Nitrobenzene-d5	100	49.2	49		18	107
		2-Fluorobiphenyl	100	52.4	52		20	109
		Terphenyl-d14	100	40.1	40		10	105

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2425

Client: Kleinfelder

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q2416-01MS	Client Sample ID:	MH-G/HMS					DataFile:	BF142872.D		
Naphthalene	1100	0	940	ug/Kg	85				51	121	
Fluorene	1100	0	950	ug/Kg	86				53	118	
Phenanthrene	1100	0	950	ug/Kg	86				52	128	
Anthracene	1100	0	950	ug/Kg	86				62	124	
Pyrene	1100	0	900	ug/Kg	82				37	122	
Benzo(a)anthracene	1100	0	960	ug/Kg	87				53	119	
Chrysene	1100	0	940	ug/Kg	85				57	121	
Benzo(b)fluoranthene	1100	0	1000	ug/Kg	91				52	117	
Benzo(a)pyrene	1100	0	980	ug/Kg	89				70	142	
Indeno(1,2,3-cd)pyrene	1100	0	660	ug/Kg	60				40	129	
Benzo(g,h,i)perylene	1100	0	590	ug/Kg	54				24	125	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2425

Client: Kleinfelder

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q2416-01MSD	Client Sample ID:	MH-G/HMSD					DataFile:	BF142873.D		
Naphthalene	1100	0	960	ug/Kg	87		2		51	121	20
Fluorene	1100	0	960	ug/Kg	87		1		53	118	20
Phenanthrene	1100	0	970	ug/Kg	88		2		52	128	20
Anthracene	1100	0	970	ug/Kg	88		2		62	124	20
Pyrene	1100	0	890	ug/Kg	81		1		37	122	20
Benzo(a)anthracene	1100	0	970	ug/Kg	88		1		53	119	20
Chrysene	1100	0	970	ug/Kg	88		3		57	121	20
Benzo(b)fluoranthene	1100	0	1100	ug/Kg	100		9		52	117	20
Benzo(a)pyrene	1100	0	1000	ug/Kg	91		2		70	142	20
Indeno(1,2,3-cd)pyrene	1100	0	730	ug/Kg	66		10		40	129	20
Benzo(g,h,i)perylene	1100	0	660	ug/Kg	60		11		24	125	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2425

Client: Kleinfelder

Analytical Method: 8270E DataFile: BF142885.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		
								Qual	Low	High	RPD
PB168625BS	Naphthalene	1700	1400	ug/Kg	82				62	100	
	Fluorene	1700	1400	ug/Kg	82				61	101	
	Phenanthrene	1700	1400	ug/Kg	82				59	103	
	Anthracene	1700	1400	ug/Kg	82				61	105	
	Pyrene	1700	1400	ug/Kg	82				59	103	
	Benzo(a)anthracene	1700	1500	ug/Kg	88				60	102	
	Chrysene	1700	1500	ug/Kg	88				59	101	
	Benzo(b)fluoranthene	1700	1500	ug/Kg	88				62	109	
	Benzo(a)pyrene	1700	1500	ug/Kg	88				63	103	
	Indeno(1,2,3-cd)pyrene	1700	1400	ug/Kg	82				63	101	
	Benzo(g,h,i)perylene	1700	1400	ug/Kg	82				70	108	

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB168625BL

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM Case No.: Q2425

SAS No.: Q2425 SDG NO.: Q2425

Lab File ID: BF142884.D

Lab Sample ID: PB168625BL

Instrument ID: BNA_F

Date Extracted: 06/26/2025

Matrix: (soil/water) SOIL

Date Analyzed: 06/27/2025

Level: (low/med) LOW

Time Analyzed: 10:49

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB168625BS	PB168625BS	BF142885.D	06/27/2025
COMP-2	Q2425-02	BF142887.D	06/27/2025
COMP-3	Q2425-03	BF142888.D	06/27/2025
MH-G/HMS	Q2416-01MS	BF142872.D	06/26/2025
MH-G/HMSD	Q2416-01MSD	BF142873.D	06/26/2025
COMP-1	Q2425-01	BF142874.D	06/26/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM

SAS No.: Q2425 SDG NO.: Q2425

Lab File ID: BF142787.D

DFTPP Injection Date: 06/19/2025

Instrument ID: BNA_F

DFTPP Injection Time: 16:37

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (1.8) 1
69	Mass 69 relative abundance	30.9
70	Less than 2.0% of mass 69	0.1 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.8
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	15
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.1 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF142788.D	06/19/2025	17:07
SSTDICC005	SSTDICC005	BF142789.D	06/19/2025	17:38
SSTDICC010	SSTDICC010	BF142790.D	06/19/2025	18:08
SSTDICC020	SSTDICC020	BF142791.D	06/19/2025	18:39
SSTDICCC040	SSTDICCC040	BF142792.D	06/19/2025	19:09
SSTDICC050	SSTDICC050	BF142793.D	06/19/2025	19:40
SSTDICC060	SSTDICC060	BF142794.D	06/19/2025	20:10
SSTDICC080	SSTDICC080	BF142795.D	06/19/2025	20:40

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM

SAS No.: Q2425 SDG NO.: Q2425

Lab File ID: BF142856.D

DFTPP Injection Date: 06/26/2025

Instrument ID: BNA_F

DFTPP Injection Time: 09:10

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (1.8) 1
69	Mass 69 relative abundance	33.7
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	15.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.4 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142857.D	06/26/2025	09:39
MH-G/HMS	Q2416-01MS	BF142872.D	06/26/2025	17:06
MH-G/HMSD	Q2416-01MSD	BF142873.D	06/26/2025	17:36
COMP-1	Q2425-01	BF142874.D	06/26/2025	18:05

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM

SAS No.: Q2425 SDG NO.: Q2425

Lab File ID: BF142881.D

DFTPP Injection Date: 06/27/2025

Instrument ID: BNA_F

DFTPP Injection Time: 09:21

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (1.8) 1
69	Mass 69 relative abundance	32.3
70	Less than 2.0% of mass 69	0.1 (0.4) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.4
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	15.1
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.2 (19.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142882.D	06/27/2025	09:51
PB168625BL	PB168625BL	BF142884.D	06/27/2025	10:49
PB168625BS	PB168625BS	BF142885.D	06/27/2025	11:19
COMP-2	Q2425-02	BF142887.D	06/27/2025	12:21
COMP-3	Q2425-03	BF142888.D	06/27/2025	12:50

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/26/2025

Lab File ID: BF142857.D Time Analyzed: 09:39

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	77664	6.875	293795	8.16	161192	9.92
UPPER LIMIT	155328	7.375	587590	8.663	322384	10.422
LOWER LIMIT	38832	6.375	146898	7.663	80596	9.422
EPA SAMPLE NO.						
01 MH-G/HMS	84939	6.88	322606	8.16	168684	9.92
02 MH-G/HMSD	79872	6.88	303973	8.16	162686	9.92
03 COMP-1	80015	6.88	306273	8.16	163617	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/26/2025

Lab File ID: BF142857.D Time Analyzed: 09:39

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	275904	11.41	159836	14.057	144296	15.545
UPPER LIMIT	551808	11.91	319672	14.557	288592	16.045
LOWER LIMIT	137952	10.91	79918	13.557	72148	15.045
EPA SAMPLE NO.						
01 MH-G/HMS	277774	11.41	149832	14.06	169288	15.55
02 MH-G/HMSD	266635	11.41	149914	14.06	166770	15.55
03 COMP-1	266840	11.40	136110	14.05	163547	15.54

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/27/2025

Lab File ID: BF142882.D Time Analyzed: 09:51

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	70500	6.875	264550	8.16	140000	9.92
UPPER LIMIT	141000	7.375	529100	8.663	280000	10.422
LOWER LIMIT	35250	6.375	132275	7.663	70000	9.422
EPA SAMPLE NO.						
01 PB168625BL	77055	6.88	295430	8.16	162519	9.92
02 PB168625BS	79288	6.88	306414	8.16	167156	9.92
03 COMP-2	69223	6.88	263260	8.16	135660	9.92
04 COMP-3	69872	6.88	263807	8.16	130470	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/27/2025

Lab File ID: BF142882.D Time Analyzed: 09:51

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	234291	11.41	139168	14.057	151209	15.545
UPPER LIMIT	468582	11.91	278336	14.557	302418	16.045
LOWER LIMIT	117146	10.91	69584	13.557	75604.5	15.045
EPA SAMPLE NO.						
01 PB168625BL	284575	11.40	168482	14.05	158706	15.55
02 PB168625BS	286462	11.41	164874	14.06	166752	15.55
03 COMP-2	203415	11.40	149806	14.05	148502	15.55
04 COMP-3	197251	11.40	147517	14.05	148110	15.55

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	AS Jenks School	Date Received:	
Client Sample ID:	PB168625BL	SDG No.:	Q2425
Lab Sample ID:	PB168625BL	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	100
Sample Wt/Vol:	30.01 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142884.D	1	06/26/25 09:20	06/27/25 10:49	PB168625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	22.7	U	22.7	170	ug/Kg
86-73-7	Fluorene	25.3	U	25.3	170	ug/Kg
85-01-8	Phenanthrene	20.9	U	20.9	170	ug/Kg
120-12-7	Anthracene	33.3	U	33.3	170	ug/Kg
129-00-0	Pyrene	36.0	U	36.0	170	ug/Kg
56-55-3	Benzo(a)anthracene	23.0	U	23.0	170	ug/Kg
218-01-9	Chrysene	19.9	U	19.9	170	ug/Kg
205-99-2	Benzo(b)fluoranthene	19.0	U	19.0	170	ug/Kg
50-32-8	Benzo(a)pyrene	29.5	U	29.5	170	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	29.1	U	29.1	170	ug/Kg
191-24-2	Benzo(g,h,i)perylene	25.7	U	25.7	170	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	81.6		18 - 107	82%	SPK: 100
321-60-8	2-Fluorobiphenyl	77.4		20 - 109	77%	SPK: 100
1718-51-0	Terphenyl-d14	78.5		10 - 105	78%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	77100	6.875			
1146-65-2	Naphthalene-d8	295000	8.157			
15067-26-2	Acenaphthene-d10	163000	9.916			
1517-22-2	Phenanthrene-d10	285000	11.404			
1719-03-5	Chrysene-d12	168000	14.051			
1520-96-3	Perylene-d12	159000	15.545			

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	AS Jenks School	Date Received:	
Client Sample ID:	PB168625BL	SDG No.:	Q2425
Lab Sample ID:	PB168625BL	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	100
Sample Wt/Vol:	30.01 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142884.D	1	06/26/25 09:20	06/27/25 10:49	PB168625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	AS Jenks School	Date Received:	
Client Sample ID:	PB168625BS	SDG No.:	Q2425
Lab Sample ID:	PB168625BS	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	100
Sample Wt/Vol:	30.02 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142885.D	1	06/26/25 09:20	06/27/25 11:19	PB168625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	1400		22.7	170	ug/Kg
86-73-7	Fluorene	1400		25.3	170	ug/Kg
85-01-8	Phenanthrene	1400		20.9	170	ug/Kg
120-12-7	Anthracene	1400		33.3	170	ug/Kg
129-00-0	Pyrene	1400		36.0	170	ug/Kg
56-55-3	Benzo(a)anthracene	1500		23.0	170	ug/Kg
218-01-9	Chrysene	1500		19.9	170	ug/Kg
205-99-2	Benzo(b)fluoranthene	1500		19.0	170	ug/Kg
50-32-8	Benzo(a)pyrene	1500		29.5	170	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	1400		29.1	170	ug/Kg
191-24-2	Benzo(g,h,i)perylene	1400		25.7	170	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	77.0		18 - 107	77%	SPK: 100
321-60-8	2-Fluorobiphenyl	74.7		20 - 109	75%	SPK: 100
1718-51-0	Terphenyl-d14	78.2		10 - 105	78%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	79300	6.875			
1146-65-2	Naphthalene-d8	306000	8.163			
15067-26-2	Acenaphthene-d10	167000	9.922			
1517-22-2	Phenanthrene-d10	286000	11.41			
1719-03-5	Chrysene-d12	165000	14.057			
1520-96-3	Perylene-d12	167000	15.545			

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	AS Jenks School	Date Received:	
Client Sample ID:	PB168625BS	SDG No.:	Q2425
Lab Sample ID:	PB168625BS	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	100
Sample Wt/Vol:	30.02 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142885.D	1	06/26/25 09:20	06/27/25 11:19	PB168625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/25/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	MH-G/HMS	SDG No.:	Q2425
Lab Sample ID:	Q2416-01MS	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	89.7
Sample Wt/Vol:	50.09 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142872.D	1	06/26/25 09:20	06/26/25 17:06	PB168625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	940		15.2	110	ug/Kg
86-73-7	Fluorene	950		16.9	110	ug/Kg
85-01-8	Phenanthrene	950		14.0	110	ug/Kg
120-12-7	Anthracene	950		22.2	110	ug/Kg
129-00-0	Pyrene	900		24.0	110	ug/Kg
56-55-3	Benzo(a)anthracene	960		15.4	110	ug/Kg
218-01-9	Chrysene	940		13.3	110	ug/Kg
205-99-2	Benzo(b)fluoranthene	1000		12.7	110	ug/Kg
50-32-8	Benzo(a)pyrene	980		19.7	110	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	660		19.4	110	ug/Kg
191-24-2	Benzo(g,h,i)perylene	590		17.2	110	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	45.0		18 - 107	45%	SPK: 100
321-60-8	2-Fluorobiphenyl	46.3		20 - 109	46%	SPK: 100
1718-51-0	Terphenyl-d14	45.4		10 - 105	45%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	84900	6.875			
1146-65-2	Naphthalene-d8	323000	8.163			
15067-26-2	Acenaphthene-d10	169000	9.922			
1517-22-2	Phenanthrene-d10	278000	11.41			
1719-03-5	Chrysene-d12	150000	14.057			
1520-96-3	Perylene-d12	169000	15.545			

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/25/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	MH-G/HMS	SDG No.:	Q2425
Lab Sample ID:	Q2416-01MS	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	89.7
Sample Wt/Vol:	50.09 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142872.D	1	06/26/25 09:20	06/26/25 17:06	PB168625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/25/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	MH-G/HMSD	SDG No.:	Q2425
Lab Sample ID:	Q2416-01MSD	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	89.7
Sample Wt/Vol:	50.04 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142873.D	1	06/26/25 09:20	06/26/25 17:36	PB168625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	960		15.2	110	ug/Kg
86-73-7	Fluorene	960		16.9	110	ug/Kg
85-01-8	Phenanthrene	970		14.0	110	ug/Kg
120-12-7	Anthracene	970		22.3	110	ug/Kg
129-00-0	Pyrene	890		24.1	110	ug/Kg
56-55-3	Benzo(a)anthracene	970		15.4	110	ug/Kg
218-01-9	Chrysene	970		13.3	110	ug/Kg
205-99-2	Benzo(b)fluoranthene	1100		12.7	110	ug/Kg
50-32-8	Benzo(a)pyrene	1000		19.7	110	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	730		19.4	110	ug/Kg
191-24-2	Benzo(g,h,i)perylene	660		17.2	110	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	46.2		18 - 107	46%	SPK: 100
321-60-8	2-Fluorobiphenyl	46.1		20 - 109	46%	SPK: 100
1718-51-0	Terphenyl-d14	45.3		10 - 105	45%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	79900	6.875			
1146-65-2	Naphthalene-d8	304000	8.163			
15067-26-2	Acenaphthene-d10	163000	9.922			
1517-22-2	Phenanthrene-d10	267000	11.41			
1719-03-5	Chrysene-d12	150000	14.057			
1520-96-3	Perylene-d12	167000	15.545			

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/25/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	MH-G/HMSD	SDG No.:	Q2425
Lab Sample ID:	Q2416-01MSD	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	89.7
Sample Wt/Vol:	50.04 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142873.D	1	06/26/25 09:20	06/26/25 17:36	PB168625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF061925.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jun 20 05:06:14 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BF142788.D 5 =BF142789.D 10 =BF142790.D 20 =BF142791.D 40 =BF142792.D 50 =BF142793.D 60 =BF142794.D 80 =BF142795.D

	Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----									
2)	1,4-Dioxane	0.479	0.499	0.471	0.461	0.509	0.479	0.467	0.480		3.60
3)	Pyridine	1.190	1.195	1.190	1.164	1.303	1.209	1.181	1.205		3.78
4)	n-Nitrosodimet...	0.610	0.617	0.596	0.668	0.632	0.614	0.623			3.98
5) S	2-Fluorophenol	2.475	2.460	2.460	2.299	2.525	2.342	2.256	2.403		4.26
6)	Aniline	1.951	1.944	1.891	1.818	1.982	1.869	1.746	1.886		4.41
7) S	Phenol-d6	3.024	2.962	2.846	2.681	2.959	2.760	2.610	2.834		5.53
8)	2-Chlorophenol	1.362	1.303	1.292	1.245	1.366	1.272	1.231	1.296		4.08
9)	Benzaldehyde	0.988	0.943	0.757	0.832	0.684		0.841			15.02
10) C	Phenol	1.648	1.653	1.591	1.516	1.651	1.530	1.437	1.575		5.31
11)	bis(2-Chloroet...	1.167	1.133	1.134	1.072	1.183	1.116	1.077	1.126		3.71
12)	1,3-Dichlorobe...	1.536	1.499	1.470	1.377	1.500	1.414	1.348	1.449		4.85
13) C	1,4-Dichlorobe...	1.563	1.496	1.472	1.409	1.523	1.425	1.347	1.462		5.04
14)	1,2-Dichlorobe...	1.458	1.422	1.409	1.331	1.451	1.353	1.284	1.387		4.72
15)	Benzyl Alcohol	1.023	1.025	0.976	1.106	1.032	0.985	1.025			4.51
16)	2,2'-oxybis(1-...	1.975	1.869	1.815	1.723	1.879	1.758	1.644	1.809		6.12
17)	2-Methylphenol	1.015	0.997	0.980	0.938	1.043	0.972	0.929	0.982		4.13
18)	Hexachloroethane	0.533	0.521	0.507	0.499	0.539	0.498	0.484	0.511		3.89
19) P	n-Nitroso-di-n...	0.831	0.879	0.840	0.833	0.787	0.847	0.812	0.766	0.824	4.30
20)	3+4-Methylphenols	1.292	1.262	1.176	1.312	1.197	1.108	1.224			6.36
21) I	Naphthalene-d8	-----ISTD-----									
22)	Acetophenone	0.464	0.451	0.459	0.418	0.461	0.428	0.406	0.441		5.28
23) S	Nitrobenzene-d5	0.738	0.724	0.747	0.708	0.771	0.724	0.694	0.729		3.48
24)	Nitrobenzene	0.345	0.325	0.332	0.319	0.349	0.326	0.314	0.330		3.88
25)	Isophorone	0.611	0.582	0.592	0.555	0.616	0.584	0.554	0.585		4.17
26) C	2-Nitrophenol	0.169	0.172	0.182	0.178	0.194	0.186	0.176	0.180		4.85
27)	2,4-Dimethylph...	0.322	0.310	0.318	0.300	0.329	0.307	0.294	0.311		3.99
28)	bis(2-Chloroet...	0.386	0.373	0.383	0.355	0.390	0.366	0.351	0.372		4.09
29) C	2,4-Dichloroph...	0.291	0.280	0.290	0.273	0.298	0.277	0.268	0.282		3.77
30)	1,2,4-Trichlor...	0.324	0.316	0.318	0.299	0.324	0.301	0.292	0.310		4.26
31)	Naphthalene	1.051	1.002	1.004	0.934	1.017	0.943	0.902	0.979		5.46
32)	Benzoic acid	0.118	0.147	0.157	0.181	0.180	0.169	0.159			15.08
33)	4-Chloroaniline	0.414	0.408	0.409	0.379	0.421	0.393	0.363	0.398		5.26
34) C	Hexachlorobuta...	0.195	0.188	0.198	0.184	0.200	0.184	0.179	0.190		4.30
35)	Caprolactam	0.073	0.075	0.072	0.081	0.078	0.069	0.075			5.84
36) C	4-Chloro-3-met...	0.300	0.275	0.286	0.269	0.298	0.278	0.260	0.281		5.24
37)	2-Methylnaphth...	0.655	0.621	0.629	0.579	0.631	0.583	0.550	0.607		6.12
38)	1-Methylnaphth...	0.684	0.645	0.650	0.598	0.654	0.604	0.563	0.628		6.56

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF061925.M

39) I	Acenaphthene-d10	-----ISTD-----								
40)	1,2,4,5-Tetrac...	0.626	0.609	0.602	0.557	0.598	0.577	0.563	0.590	4.27
41) P	Hexachlorocycl...		0.235	0.308	0.336	0.365	0.376	0.389	0.335	17.02
42) S	2,4,6-Tribromo...	0.415	0.414	0.428	0.400	0.438	0.411	0.374	0.412	4.96
43) C	2,4,6-Trichlor...	0.385	0.380	0.394	0.365	0.408	0.389	0.371	0.385	3.71
44)	2,4,5-Trichlor...	0.407	0.403	0.421	0.391	0.416	0.407	0.388	0.405	2.95
45) S	2-Fluorobiphenyl	3.480	3.272	3.184	2.861	2.999	2.816	2.702	3.045	9.15
46)	1,1'-Biphenyl	1.701	1.616	1.637	1.505	1.609	1.525	1.466	1.580	5.27
47)	2-Chloronaphth...	1.300	1.202	1.208	1.129	1.209	1.145	1.106	1.186	5.50
48)	2-Nitroaniline	0.321	0.329	0.337	0.324	0.358	0.337	0.320	0.332	3.93
49)	Acenaphthylene	2.110	2.006	2.011	1.861	1.995	1.891	1.787	1.952	5.64
50)	Dimethylphthalate	1.372	1.334	1.330	1.228	1.335	1.277	1.186	1.295	5.17
51)	2,6-Dinitrotol...	0.297	0.279	0.290	0.277	0.294	0.284	0.268	0.284	3.57
52) C	Acenaphthene	1.247	1.199	1.241	1.142	1.228	1.172	1.107	1.191	4.44
53)	3-Nitroaniline	0.311	0.311	0.316	0.302	0.338	0.320	0.292	0.313	4.52
54) P	2,4-Dinitrophenol		0.095	0.129	0.142	0.173	0.163	0.152	0.142	19.61
55)	Dibenzofuran	1.844	1.772	1.763	1.626	1.752	1.660	1.537	1.708	6.13
56) P	4-Nitrophenol		0.189	0.218	0.208	0.238	0.229	0.203	0.214	8.25
57)	2,4-Dinitrotol...	0.373	0.378	0.397	0.367	0.407	0.383	0.338	0.378	5.90
58)	Fluorene	1.464	1.412	1.395	1.260	1.362	1.260	1.183	1.334	7.58
59)	2,3,4,6-Tetrac...	0.327	0.327	0.339	0.312	0.345	0.331	0.304	0.326	4.43
60)	Diethylphthalate	1.328	1.302	1.319	1.211	1.318	1.266	1.141	1.269	5.50
61)	4-Chlorophenyl...	0.713	0.656	0.669	0.599	0.640	0.609	0.568	0.636	7.64
62)	4-Nitroaniline	0.279	0.276	0.297	0.272	0.316	0.300	0.264	0.286	6.48
63)	Azobenzene	1.215	1.149	1.167	1.081	1.176	1.122	1.032	1.134	5.47
64) I	Phenanthrene-d10	-----ISTD-----								
65)	4,6-Dinitro-2-...		0.099	0.119	0.119	0.134	0.130	0.125	0.121	10.04
66) c	n-Nitrosodiphe...	0.736	0.686	0.710	0.670	0.717	0.671	0.662	0.693	4.07
67)	4-Bromophenyl-...	0.231	0.221	0.235	0.218	0.236	0.228	0.225	0.228	3.03
68)	Hexachlorobenzene	0.263	0.248	0.257	0.241	0.262	0.252	0.247	0.253	3.29
69)	Atrazine	0.176	0.172	0.180	0.174	0.195	0.183	0.174	0.179	4.48
70) C	Pentachlorophenol		0.102	0.119	0.125	0.142	0.135	0.133	0.126	11.30
71)	Phenanthrene	1.183	1.113	1.115	1.031	1.122	1.028	0.992	1.083	6.24
72)	Anthracene	1.226	1.123	1.149	1.064	1.140	1.065	1.011	1.111	6.35
73)	Carbazole	1.040	0.986	0.999	0.900	1.002	0.920	0.866	0.959	6.67
74)	Di-n-butylphth...	0.997	1.006	1.045	0.946	1.073	0.989	0.936	0.999	4.92
75) C	Fluoranthene	1.136	1.088	1.090	0.954	1.059	0.972	0.899	1.028	8.44
76) I	Chrysene-d12	-----ISTD-----								
77)	Benzidine		0.745	0.860	0.773	0.824	0.719	0.587	0.751	12.71
78)	Pyrene	2.110	1.940	1.968	1.742	2.030	1.863	1.468	1.875	11.43
79) S	Terphenyl-d14	3.383	3.111	3.066	2.655	3.070	2.772	2.208	2.895	13.30
80)	Butylbenzylpht...	0.478	0.470	0.538	0.547	0.613	0.597	0.563	0.544	10.01
81)	Benzo(a)anthra...	1.392	1.358	1.366	1.265	1.403	1.376	1.284	1.349	3.98
82)	3,3'-Dichlorob...		0.398	0.440	0.429	0.476	0.445	0.438	0.438	5.73
83)	Chrysene	1.243	1.191	1.208	1.164	1.295	1.177	1.150	1.204	4.17
84)	Bis(2-ethylhex...	0.650	0.675	0.767	0.803	0.861	0.843	0.878	0.782	11.52
85) c	Di-n-octyl pht...		1.207	1.352	1.463	1.610	1.569	1.651	1.475	11.54

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF061925.M

86) I	Perylene-d12	-----ISTD-----								
87)	Indeno(1,2,3-c...	1.528	1.488	1.545	1.474	1.648	1.539	1.442	1.523	4.38
88)	Benzo(b)fluora...	1.164	1.220	1.226	1.072	1.152	1.197	1.070	1.157	5.61
89)	Benzo(k)fluora...	1.206	1.004	1.035	1.037	1.193	1.038	1.068	1.083	7.55
90) C	Benzo(a)pyrene	1.142	1.065	1.124	1.069	1.201	1.141	1.084	1.118	4.37
91)	Dibenzo(a,h)an...	1.249	1.214	1.293	1.180	1.337	1.240	1.152	1.238	5.13
92)	Benzo(g,h,i)pe...	1.239	1.198	1.266	1.188	1.332	1.255	1.161	1.234	4.66

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG No.: Q2425

Instrument ID: BNA_F Calibration Date/Time: 06/26/2025 09:39

Lab File ID: BF142857.D Init. Calib. Date(s): 06/19/2025 06/19/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 17:07 20:40

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.201	1.244		3.6	
Phenol-d6	1.417	1.459		3.0	
Nitrobenzene-d5	0.365	0.443		21.4	
Naphthalene	0.979	1.015		3.7	
2-Fluorobiphenyl	1.522	1.697		11.5	
Fluorene	1.334	1.364		2.2	
2,4,6-Tribromophenol	0.206	0.223		8.3	
Phenanthrene	1.083	1.100		1.6	
Anthracene	1.111	1.136		2.3	
Pyrene	1.875	1.932		3.0	
Terphenyl-d14	1.447	1.677		15.9	
Benzo (a) anthracene	1.349	1.377		2.1	
Chrysene	1.204	1.250		3.8	
Benzo (b) fluoranthene	1.157	1.283		10.9	
Benzo (a) pyrene	1.118	1.173		4.9	20.0
Indeno (1,2,3-cd) pyrene	1.523	1.584		4.0	
Benzo (g,h,i) perylene	1.234	1.288		4.4	

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG No.: Q2425

Instrument ID: BNA_F Calibration Date/Time: 06/27/2025 09:51

Lab File ID: BF142882.D Init. Calib. Date(s): 06/19/2025 06/19/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 17:07 20:40

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.201	1.268		5.6	
Phenol-d6	1.417	1.473		4.0	
Nitrobenzene-d5	0.365	0.451		23.6	
Naphthalene	0.979	1.021		4.3	
2-Fluorobiphenyl	1.522	1.783		17.1	
Fluorene	1.334	1.383		3.7	
2,4,6-Tribromophenol	0.206	0.217		5.3	
Phenanthrene	1.083	1.105		2.0	
Anthracene	1.111	1.136		2.3	
Pyrene	1.875	1.800		-4.0	
Terphenyl-d14	1.447	1.565		8.2	
Benzo (a) anthracene	1.349	1.390		3.0	
Chrysene	1.204	1.236		2.7	
Benzo (b) fluoranthene	1.157	1.220		5.4	
Benzo (a) pyrene	1.118	1.169		4.6	20.0
Indeno (1,2,3-cd) pyrene	1.523	1.469		-3.5	
Benzo (g,h,i) perylene	1.234	1.183		-4.1	

All other compounds must meet a minimum RRF of 0.010.

LAB CHRONICLE

OrderID: Q2425	OrderDate: 6/25/2025 2:03:00 PM
Client: Kleinfelder	Project: AS Jenks School
Contact: Mark Warchol	Location: D51,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2425-01	COMP-1	SOIL			06/24/25			06/25/25
			PCB Group1	8082A		06/26/25	06/26/25	
			PESTICIDE Group1	8081B		06/26/25	06/27/25	
Q2425-02	COMP-2	SOIL			06/24/25			06/25/25
			PCB Group1	8082A		06/26/25	06/26/25	
			PESTICIDE Group1	8081B		06/26/25	06/27/25	
Q2425-03	COMP-3	SOIL			06/24/25			06/25/25
			PCB Group1	8082A		06/26/25	06/26/25	
			PESTICIDE Group1	8081B		06/26/25	06/27/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2425

Order ID: Q2425

Client: Kleinfelder

Project ID: AS Jenks School

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000

A
B
C
D
E
F
G
H



SAMPLE DATA

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-1	SDG No.:	Q2425
Lab Sample ID:	Q2425-01	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	76.9
Sample Wt/Vol:	30.01	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PESTICIDE Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089195.D	1	06/26/25 08:20	06/27/25 19:41	PB168623

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
309-00-2	Aldrin	0.16	U	0.16	2.20	ug/kg
60-57-1	Dieldrin	0.18	U	0.18	2.20	ug/kg
72-55-9	4,4-DDE	0.18	U	0.18	2.20	ug/kg
72-54-8	4,4-DDD	0.20	U	0.20	2.20	ug/kg
50-29-3	4,4-DDT	0.18	U	0.18	2.20	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	18.1		20 - 144	91%	SPK: 20
877-09-8	Tetrachloro-m-xylene	14.7		19 - 148	74%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-2	SDG No.:	Q2425
Lab Sample ID:	Q2425-02	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	79.7
Sample Wt/Vol:	30.04	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PESTICIDE Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089196.D	1	06/26/25 08:20	06/27/25 19:54	PB168623

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
309-00-2	Aldrin	0.15	U	0.15	2.10	ug/kg
60-57-1	Dieldrin	0.18	U	0.18	2.10	ug/kg
72-55-9	4,4-DDE	0.18	U	0.18	2.10	ug/kg
72-54-8	4,4-DDD	0.19	U	0.19	2.10	ug/kg
50-29-3	4,4-DDT	0.18	U	0.18	2.10	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	14.8		20 - 144	74%	SPK: 20
877-09-8	Tetrachloro-m-xylene	16.9		19 - 148	85%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-3	SDG No.:	Q2425
Lab Sample ID:	Q2425-03	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	82.6
Sample Wt/Vol:	30.07	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PESTICIDE Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089197.D	1	06/26/25 08:20	06/27/25 20:08	PB168623

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
309-00-2	Aldrin	0.14	U	0.14	2.10	ug/kg
60-57-1	Dieldrin	0.17	U	0.17	2.10	ug/kg
72-55-9	4,4-DDE	0.17	U	0.17	2.10	ug/kg
72-54-8	4,4-DDD	0.18	U	0.18	2.10	ug/kg
50-29-3	4,4-DDT	0.17	U	0.17	2.10	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	17.4		20 - 144	87%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.0		19 - 148	100%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



QC SUMMARY

Surrogate Summary

SDG No.: Q2425

Client: Kleinfelder

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD088990.D	PIBLK-PD088990.D	Decachlorobiphenyl	1	20	18.2	91		57	171
		Tetrachloro-m-xylene	1	20	16.1	80		61	148
		Decachlorobiphenyl	2	20	18.6	93		57	171
		Tetrachloro-m-xylene	2	20	17.7	88		61	148
I.BLK-PD089157.D	PIBLK-PD089157.D	Decachlorobiphenyl	1	20	21.3	106		57	171
		Tetrachloro-m-xylene	1	20	20.9	104		61	148
		Decachlorobiphenyl	2	20	20.8	104		57	171
		Tetrachloro-m-xylene	2	20	22.2	111		61	148
PB168623BL	PB168623BL	Decachlorobiphenyl	1	20	17.6	88		20	144
		Tetrachloro-m-xylene	1	20	17.9	89		19	148
		Decachlorobiphenyl	2	20	16.3	81		20	144
		Tetrachloro-m-xylene	2	20	19.3	97		19	148
PB168623BS	PB168623BS	Decachlorobiphenyl	1	20	22.1	111		20	144
		Tetrachloro-m-xylene	1	20	21.3	106		19	148
		Decachlorobiphenyl	2	20	20.9	105		20	144
		Tetrachloro-m-xylene	2	20	22.9	114		19	148
I.BLK-PD089171.D	PIBLK-PD089171.D	Decachlorobiphenyl	1	20	19.8	99		57	171
		Tetrachloro-m-xylene	1	20	20.5	103		61	148
		Decachlorobiphenyl	2	20	17.4	87		57	171
		Tetrachloro-m-xylene	2	20	21.6	108		61	148
I.BLK-PD089190.D	PIBLK-PD089190.D	Decachlorobiphenyl	1	20	22.1	110		57	171
		Tetrachloro-m-xylene	1	20	22.8	114		61	148
		Decachlorobiphenyl	2	20	21.5	107		57	171
		Tetrachloro-m-xylene	2	20	24.4	122		61	148
Q2425-01	COMP-1	Decachlorobiphenyl	1	20	14.3	72		20	144
		Tetrachloro-m-xylene	1	20	13.7	68		19	148
		Decachlorobiphenyl	2	20	18.1	91		20	144
		Tetrachloro-m-xylene	2	20	14.7	74		19	148
Q2425-02	COMP-2	Decachlorobiphenyl	1	20	14.8	74		20	144
		Tetrachloro-m-xylene	1	20	15.6	78		19	148
		Decachlorobiphenyl	2	20	14.5	73		20	144
		Tetrachloro-m-xylene	2	20	16.9	85		19	148
Q2425-03	COMP-3	Decachlorobiphenyl	1	20	17.4	87		20	144
		Tetrachloro-m-xylene	1	20	18.5	93		19	148
		Decachlorobiphenyl	2	20	16.6	83		20	144
		Tetrachloro-m-xylene	2	20	20.0	100		19	148
Q2425-03MS	COMP-3MS	Decachlorobiphenyl	1	20	16.0	80		20	144
		Tetrachloro-m-xylene	1	20	17.0	85		19	148
		Decachlorobiphenyl	2	20	16.5	82		20	144
		Tetrachloro-m-xylene	2	20	18.5	92		19	148
Q2425-03MSD	COMP-3MSD	Decachlorobiphenyl	1	20	16.8	84		20	144

Surrogate Summary

SDG No.: Q2425

Client: Kleinfelder

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q2425-03MSD	COMP-3MSD	Tetrachloro-m-xylene	1	20	17.4	87		19	148
		Decachlorobiphenyl	2	20	16.7	83		20	144
		Tetrachloro-m-xylene	2	20	18.4	92		19	148
I.BLK-PD089201.D	PIBLK-PD089201.D	Decachlorobiphenyl	1	20	22.4	112		57	171
		Tetrachloro-m-xylene	1	20	22.9	114		61	148
		Decachlorobiphenyl	2	20	21.9	109		57	171
		Tetrachloro-m-xylene	2	20	24.5	122		61	148

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2425</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Kleinfelder</u>	DataFile :	<u>PD089198.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2425-03MS (Column 1)	COMP-3MS Aldrin	20.12	0	17.6	ug/kg	87				49	139	
	Dieldrin	20.12	0	17.4	ug/kg	86				47	161	
	4,4'-DDE	20.12	0	17.2	ug/kg	85				55	136	
	4,4'-DDD	20.12	0	18.6	ug/kg	92				47	163	
	4,4'-DDT	20.12	0	11.5	ug/kg	57				51	146	
Client Sample ID: Q2425-03MS (Column 2)	COMP-3MS Aldrin	20.12	0	17.9	ug/kg	89				49	139	
	Dieldrin	20.12	0	17.6	ug/kg	87				47	161	
	4,4'-DDE	20.12	0	17.7	ug/kg	88				55	136	
	4,4'-DDD	20.12	0	17.8	ug/kg	88				47	163	
	4,4'-DDT	20.12	0	13.3	ug/kg	66				51	146	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2425</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Kleinfelder</u>	DataFile :	<u>PD089199.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2425-03MSD (Column 1)	COMP-3MSD											
	Aldrin	20.16	0	18.2	ug/kg	90		3		49	139	20
	Dieldrin	20.16	0	17.8	ug/kg	88		2		47	161	20
	4,4'-DDE	20.16	0	18.2	ug/kg	90		6		55	136	20
	4,4'-DDD	20.16	0	17.7	ug/kg	88		4		47	163	20
	4,4'-DDT	20.16	0	10.2	ug/kg	51		11		51	146	20
Client Sample ID: Q2425-03MSD (Column 2)	COMP-3MSD											
	Aldrin	20.16	0	18.3	ug/kg	91		2		49	139	20
	Dieldrin	20.16	0	17.9	ug/kg	89		2		47	161	20
	4,4'-DDE	20.16	0	18.4	ug/kg	91		3		55	136	20
	4,4'-DDD	20.16	0	17.1	ug/kg	85		3		47	163	20
	4,4'-DDT	20.16	0	12.9	ug/kg	64		3		51	146	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2425</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Kleinfelder</u>	Datafile :	<u>PD089167.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB168623BS (Column 1)	Aldrin	16.65	16.2	ug/kg	97				82	124		
	Dieldrin	16.65	15.9	ug/kg	95				85	121		
	4,4'-DDE	16.65	15.9	ug/kg	95				81	123		
	4,4'-DDD	16.65	16.1	ug/kg	97				80	131		
	4,4'-DDT	16.65	13.9	ug/kg	83				70	129		
PB168623BS (Column 2)	Aldrin	16.65	15.9	ug/kg	95				82	124		
	Dieldrin	16.65	15.5	ug/kg	93				85	121		
	4,4'-DDE	16.65	15.5	ug/kg	93				81	123		
	4,4'-DDD	16.65	16.0	ug/kg	96				80	131		
	4,4'-DDT	16.65	13.9	ug/kg	83				70	129		

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB168623BL

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM Case No.: Q2425

SAS No.: Q2425 SDG NO.: Q2425

Lab Sample ID: PB168623BL

Lab File ID: PD089166.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 06/26/2025

Date Analyzed (1): 06/26/2025

Date Analyzed (2): 06/26/2025

Time Analyzed (1): 13:44

Time Analyzed (2): 13:44

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB168623BS	PB168623BS	PD089167.D	06/26/2025	06/26/2025
COMP-1	Q2425-01	PD089195.D	06/27/2025	06/27/2025
COMP-2	Q2425-02	PD089196.D	06/27/2025	06/27/2025
COMP-3	Q2425-03	PD089197.D	06/27/2025	06/27/2025
COMP-3MS	Q2425-03MS	PD089198.D	06/27/2025	06/27/2025
COMP-3MSD	Q2425-03MSD	PD089199.D	06/27/2025	06/27/2025

COMMENTS: _____



QC SAMPLE DATA

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	AS Jenks School	Date Received:	
Client Sample ID:	PB168623BL	SDG No.:	Q2425
Lab Sample ID:	PB168623BL	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	100
Sample Wt/Vol:	30.02	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PESTICIDE Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Decanted:	
		Final Vol:	10000
			uL
		Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089166.D	1	06/26/25 08:20	06/26/25 13:44	PB168623

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
309-00-2	Aldrin	0.12	U	0.12	1.70	ug/kg
60-57-1	Dieldrin	0.14	U	0.14	1.70	ug/kg
72-55-9	4,4-DDE	0.14	U	0.14	1.70	ug/kg
72-54-8	4,4-DDD	0.15	U	0.15	1.70	ug/kg
50-29-3	4,4-DDT	0.14	U	0.14	1.70	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	17.6		20 - 144	88%	SPK: 20
877-09-8	Tetrachloro-m-xylene	19.3		19 - 148	97%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/17/25
Project:	AS Jenks School	Date Received:	06/17/25
Client Sample ID:	PIBLK-PD088990.D	SDG No.:	Q2425
Lab Sample ID:	I.BLK-PD088990.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PESTICIDE Group1
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088990.D	1		06/17/25	PD061825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
309-00-2	Aldrin	0.0036	U	0.0036	0.050	ug/L
60-57-1	Dieldrin	0.0036	U	0.0036	0.050	ug/L
72-55-9	4,4-DDE	0.0037	U	0.0037	0.050	ug/L
72-54-8	4,4-DDD	0.0071	U	0.0071	0.050	ug/L
50-29-3	4,4-DDT	0.0035	U	0.0035	0.050	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	18.6		57 - 171	93%	SPK: 20
877-09-8	Tetrachloro-m-xylene	17.7		61 - 148	88%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/26/25
Project:	AS Jenks School	Date Received:	06/26/25
Client Sample ID:	PIBLK-PD089157.D	SDG No.:	Q2425
Lab Sample ID:	I.BLK-PD089157.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PESTICIDE Group1
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089157.D	1		06/26/25	pd062625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
309-00-2	Aldrin	0.0036	U	0.0036	0.050	ug/L
60-57-1	Dieldrin	0.0036	U	0.0036	0.050	ug/L
72-55-9	4,4-DDE	0.0037	U	0.0037	0.050	ug/L
72-54-8	4,4-DDD	0.0071	U	0.0071	0.050	ug/L
50-29-3	4,4-DDT	0.0035	U	0.0035	0.050	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	21.3		57 - 171	106%	SPK: 20
877-09-8	Tetrachloro-m-xylene	22.2		61 - 148	111%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/26/25
Project:	AS Jenks School	Date Received:	06/26/25
Client Sample ID:	PIBLK-PD089171.D	SDG No.:	Q2425
Lab Sample ID:	I.BLK-PD089171.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PESTICIDE Group1
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089171.D	1		06/26/25	pd062625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
309-00-2	Aldrin	0.0036	U	0.0036	0.050	ug/L
60-57-1	Dieldrin	0.0036	U	0.0036	0.050	ug/L
72-55-9	4,4-DDE	0.0037	U	0.0037	0.050	ug/L
72-54-8	4,4-DDD	0.0071	U	0.0071	0.050	ug/L
50-29-3	4,4-DDT	0.0035	U	0.0035	0.050	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	19.8		57 - 171	99%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.6		61 - 148	108%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/27/25
Project:	AS Jenks School	Date Received:	06/27/25
Client Sample ID:	PIBLK-PD089190.D	SDG No.:	Q2425
Lab Sample ID:	I.BLK-PD089190.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PESTICIDE Group1
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089190.D	1		06/27/25	pd062825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
309-00-2	Aldrin	0.0036	U	0.0036	0.050	ug/L
60-57-1	Dieldrin	0.0036	U	0.0036	0.050	ug/L
72-55-9	4,4-DDE	0.0037	U	0.0037	0.050	ug/L
72-54-8	4,4-DDD	0.0071	U	0.0071	0.050	ug/L
50-29-3	4,4-DDT	0.0035	U	0.0035	0.050	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	22.1		57 - 171	110%	SPK: 20
877-09-8	Tetrachloro-m-xylene	24.4		61 - 148	122%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/27/25
Project:	AS Jenks School	Date Received:	06/27/25
Client Sample ID:	PIBLK-PD089201.D	SDG No.:	Q2425
Lab Sample ID:	I.BLK-PD089201.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PESTICIDE Group1
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089201.D	1		06/27/25	pd062825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
309-00-2	Aldrin	0.0036	U	0.0036	0.050	ug/L
60-57-1	Dieldrin	0.0036	U	0.0036	0.050	ug/L
72-55-9	4,4-DDE	0.0037	U	0.0037	0.050	ug/L
72-54-8	4,4-DDD	0.0071	U	0.0071	0.050	ug/L
50-29-3	4,4-DDT	0.0035	U	0.0035	0.050	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	22.4		57 - 171	112%	SPK: 20
877-09-8	Tetrachloro-m-xylene	24.5		61 - 148	122%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	AS Jenks School	Date Received:	
Client Sample ID:	PB168623BS	SDG No.:	Q2425
Lab Sample ID:	PB168623BS	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	100
Sample Wt/Vol:	30.03	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PESTICIDE Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089167.D	1	06/26/25 08:20	06/26/25 14:59	PB168623

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
309-00-2	Aldrin	16.2		0.12	1.70	ug/kg
60-57-1	Dieldrin	15.9		0.14	1.70	ug/kg
72-55-9	4,4-DDE	15.9		0.14	1.70	ug/kg
72-54-8	4,4-DDD	16.1		0.15	1.70	ug/kg
50-29-3	4,4-DDT	13.9		0.14	1.70	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	22.1		20 - 144	111%	SPK: 20
877-09-8	Tetrachloro-m-xylene	22.9		19 - 148	114%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-3MS	SDG No.:	Q2425
Lab Sample ID:	Q2425-03MS	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	82.6
Sample Wt/Vol:	30.08	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PESTICIDE Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089198.D	1	06/26/25 08:20	06/27/25 20:22	PB168623

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
309-00-2	Aldrin	17.9		0.14	2.10	ug/kg
60-57-1	Dieldrin	17.6		0.17	2.10	ug/kg
72-55-9	4,4-DDE	17.7		0.17	2.10	ug/kg
72-54-8	4,4-DDD	18.6		0.18	2.10	ug/kg
50-29-3	4,4-DDT	13.3		0.17	2.10	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	16.5		20 - 144	82%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.5		19 - 148	92%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-3MSD	SDG No.:	Q2425
Lab Sample ID:	Q2425-03MSD	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	82.6
Sample Wt/Vol:	30.03	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PESTICIDE Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089199.D	1	06/26/25 08:20	06/27/25 20:35	PB168623

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
309-00-2	Aldrin	18.3		0.15	2.10	ug/kg
60-57-1	Dieldrin	17.9		0.17	2.10	ug/kg
72-55-9	4,4-DDE	18.4		0.17	2.10	ug/kg
72-54-8	4,4-DDD	17.7		0.18	2.10	ug/kg
50-29-3	4,4-DDT	12.9		0.17	2.10	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	16.8		20 - 144	84%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.4		19 - 148	92%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



CALIBRATION SUMMARY

A
B
C
D
E
F
G
H

GC Column: ZB-MR1 ID: 0.32 (mm)

[illegible]

A
B
C
D
E
F
G
H

Calibration Times: 15:52 16:47

LAB FILE ID:	RT 100 =	<u>PD088993.D</u>	RT 075 =	<u>PD088994.D</u>
	RT 050 =	PD088995.D	RT 025 =	PD088996.D
			RT 005 =	PD088997.D

[illegible]

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM **Case No.:** Q2425 **SAS No.:** Q2425 **SDG NO.:** Q2425

Instrument ID: ECD_D **Calibration Date(s):** 06/17/2025 06/17/2025
Calibration Times: 15:52 16:47

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 100 = <u>PD088993.D</u>		CF 075 = <u>PD088994.D</u>			
CF 050 = <u>PD088995.D</u>		CF 025 = <u>PD088996.D</u>		CF 005 = <u>PD088997.D</u>			
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	3557450000	3439230000	3414120000	3228670000	3520780000	3432050000	4
4,4'-DDE	4564130000	4401240000	4312380000	4092310000	4435110000	4361030000	4
4,4'-DDT	3928430000	3802000000	3763430000	3580670000	3866490000	3788210000	3
Aldrin	5504150000	5312640000	5254800000	4989300000	5504000000	5312980000	4
Decachlorobiphenyl	3598890000	3628150000	3773200000	3897960000	4739940000	3927630000	12
Dieldrin	4933650000	4797080000	4750270000	4527510000	4978810000	4797470000	4
Tetrachloro-m-xylene	2812300000	2762040000	2767280000	2751820000	3172230000	2853130000	6

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM **Case No.:** Q2425 **SAS No.:** Q2425 **SDG NO.:** Q2425

Instrument ID: ECD_D **Calibration Date(s):** 06/17/2025 06/17/2025
Calibration Times: 15:52 16:47

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD088993.D</u>	CF 075 =	<u>PD088994.D</u>		
CF 050 =	<u>PD088995.D</u>	CF 025 =	<u>PD088996.D</u>	CF 005 =	<u>PD088997.D</u>		
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	17754100000	17899700000	18191100000	18805800000	22996200000	19129400000	11
4,4'-DDE	21343500000	21440200000	21823600000	22564100000	27409000000	22916100000	11
4,4'-DDT	19391000000	19327800000	19594100000	20005200000	23078500000	20279300000	8
Aldrin	22719200000	22747700000	23159400000	23852800000	28607200000	24217300000	10
Decachlorobiphenyl	17946500000	17939700000	18428600000	19427500000	25107500000	19770000000	15
Dieldrin	21438700000	21622000000	22120600000	22941800000	27919200000	23208500000	12
Tetrachloro-m-xylene	15879800000	15866400000	15978000000	16730900000	20310700000	16953200000	11

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 11:47 Initial Calibration Time(s): 15:52 16:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.07	9.07	8.97	9.17	0.00
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00
Aldrin	5.27	5.27	5.17	5.37	0.00
Dieldrin	6.35	6.35	6.25	6.45	0.00
4,4'-DDE	6.20	6.20	6.10	6.30	0.00
4,4'-DDD	6.71	6.71	6.61	6.81	0.01
4,4'-DDT	7.02	7.02	6.92	7.12	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 11:47 Initial Calibration Time(s): 15:52 16:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	8.07	8.07	7.97	8.17	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
Aldrin	4.37	4.37	4.27	4.47	0.00
Dieldrin	5.51	5.51	5.41	5.61	0.00
4,4'-DDE	5.38	5.38	5.28	5.48	0.00
4,4'-DDD	5.93	5.93	5.83	6.03	0.00
4,4'-DDT	6.18	6.18	6.08	6.28	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL01 Date Analyzed: 06/26/2025

Lab Sample No.: PSTDCCC050 Data File : PD089159.D Time Analyzed: 11:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.705	6.605	6.805	53.110	50.000	6.2
4,4'-DDE	6.196	6.096	6.296	52.180	50.000	4.4
4,4'-DDT	7.021	6.921	7.121	48.890	50.000	-2.2
Aldrin	5.271	5.171	5.371	54.570	50.000	9.1
Decachlorobiphenyl	9.072	8.972	9.172	49.280	50.000	-1.4
Dieldrin	6.347	6.247	6.447	53.280	50.000	6.6
Tetrachloro-m-xylene	3.551	3.450	3.650	53.870	50.000	7.7

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL01 Date Analyzed: 06/26/2025

Lab Sample No.: PSTDCCC050 Data File : PD089159.D Time Analyzed: 11:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.929	5.830	6.030	51.260	50.000	2.5
4,4'-DDE	5.375	5.275	5.475	51.130	50.000	2.3
4,4'-DDT	6.183	6.084	6.284	48.440	50.000	-3.1
Aldrin	4.369	4.269	4.469	52.050	50.000	4.1
Decachlorobiphenyl	8.071	7.972	8.172	47.520	50.000	-5.0
Dieldrin	5.512	5.413	5.613	51.120	50.000	2.2
Tetrachloro-m-xylene	2.881	2.780	2.980	52.560	50.000	5.1

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 17:20 Initial Calibration Time(s): 15:52 16:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.08	9.07	8.97	9.17	-0.01
Tetrachloro-m-xylene	3.56	3.55	3.45	3.65	-0.01
Aldrin	5.28	5.27	5.17	5.37	-0.01
Dieldrin	6.35	6.35	6.25	6.45	0.00
4,4'-DDE	6.20	6.20	6.10	6.30	0.00
4,4'-DDD	6.71	6.71	6.61	6.81	0.00
4,4'-DDT	7.03	7.02	6.92	7.12	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 17:20 Initial Calibration Time(s): 15:52 16:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.07	8.07	7.97	8.17	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
Aldrin	4.37	4.37	4.27	4.47	0.00
Dieldrin	5.51	5.51	5.41	5.61	0.00
4,4'-DDE	5.38	5.38	5.28	5.48	0.00
4,4'-DDD	5.93	5.93	5.83	6.03	0.00
4,4'-DDT	6.18	6.18	6.08	6.28	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL02 Date Analyzed: 06/26/2025

Lab Sample No.: PSTDCCC050 Data File : PD089172.D Time Analyzed: 17:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.709	6.605	6.805	52.210	50.000	4.4
4,4'-DDE	6.200	6.096	6.296	53.450	50.000	6.9
4,4'-DDT	7.025	6.921	7.121	46.980	50.000	-6.0
Aldrin	5.275	5.171	5.371	55.770	50.000	11.5
Decachlorobiphenyl	9.077	8.972	9.172	49.220	50.000	-1.6
Dieldrin	6.351	6.247	6.447	53.140	50.000	6.3
Tetrachloro-m-xylene	3.555	3.450	3.650	57.010	50.000	14.0

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL02 Date Analyzed: 06/26/2025

Lab Sample No.: PSTDCCC050 Data File : PD089172.D Time Analyzed: 17:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.930	5.830	6.030	49.850	50.000	-0.3
4,4'-DDE	5.376	5.275	5.475	51.580	50.000	3.2
4,4'-DDT	6.184	6.084	6.284	46.860	50.000	-6.3
Aldrin	4.369	4.269	4.469	53.840	50.000	7.7
Decachlorobiphenyl	8.073	7.972	8.172	44.560	50.000	-10.9
Dieldrin	5.513	5.413	5.613	51.790	50.000	3.6
Tetrachloro-m-xylene	2.881	2.780	2.980	56.090	50.000	12.2

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/27/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 18:46 Initial Calibration Time(s): 15:52 16:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.07	9.07	8.97	9.17	0.00
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00
Aldrin	5.27	5.27	5.17	5.37	0.00
Dieldrin	6.35	6.35	6.25	6.45	0.00
4,4'-DDE	6.20	6.20	6.10	6.30	0.00
4,4'-DDD	6.70	6.71	6.61	6.81	0.01
4,4'-DDT	7.02	7.02	6.92	7.12	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/27/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 18:46 Initial Calibration Time(s): 15:52 16:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.07	8.07	7.97	8.17	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
Aldrin	4.37	4.37	4.27	4.47	0.00
Dieldrin	5.51	5.51	5.41	5.61	0.00
4,4'-DDE	5.37	5.38	5.28	5.48	0.01
4,4'-DDD	5.93	5.93	5.83	6.03	0.00
4,4'-DDT	6.18	6.18	6.08	6.28	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/27/2025

Lab Sample No.: PSTDCCC050 Data File : PD089192.D Time Analyzed: 18:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.704	6.605	6.805	57.030	50.000	14.1
4,4'-DDE	6.195	6.096	6.296	55.130	50.000	10.3
4,4'-DDT	7.020	6.921	7.121	43.230	50.000	-13.5
Aldrin	5.270	5.171	5.371	57.560	50.000	15.1
Decachlorobiphenyl	9.071	8.972	9.172	50.650	50.000	1.3
Dieldrin	6.346	6.247	6.447	56.130	50.000	12.3
Tetrachloro-m-xylene	3.550	3.450	3.650	58.070	50.000	16.1

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/27/2025

Lab Sample No.: PSTDCCC050 Data File : PD089192.D Time Analyzed: 18:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.928	5.830	6.030	53.960	50.000	7.9
4,4'-DDE	5.374	5.275	5.475	54.320	50.000	8.6
4,4'-DDT	6.182	6.084	6.284	43.220	50.000	-13.6
Aldrin	4.367	4.269	4.469	55.730	50.000	11.5
Decachlorobiphenyl	8.070	7.972	8.172	50.160	50.000	0.3
Dieldrin	5.511	5.413	5.613	54.310	50.000	8.6
Tetrachloro-m-xylene	2.881	2.780	2.980	58.440	50.000	16.9

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/27/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 21:16 Initial Calibration Time(s): 15:52 16:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.07	9.07	8.97	9.17	0.00
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00
Aldrin	5.27	5.27	5.17	5.37	0.00
Dieldrin	6.35	6.35	6.25	6.45	0.01
4,4'-DDE	6.19	6.20	6.10	6.30	0.01
4,4'-DDD	6.70	6.71	6.61	6.81	0.01
4,4'-DDT	7.02	7.02	6.92	7.12	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/27/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 21:16 Initial Calibration Time(s): 15:52 16:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.07	8.07	7.97	8.17	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
Aldrin	4.37	4.37	4.27	4.47	0.00
Dieldrin	5.51	5.51	5.41	5.61	0.00
4,4'-DDE	5.37	5.38	5.28	5.48	0.01
4,4'-DDD	5.93	5.93	5.83	6.03	0.00
4,4'-DDT	6.18	6.18	6.08	6.28	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL04 Date Analyzed: 06/27/2025

Lab Sample No.: PSTDCCC050 Data File : PD089202.D Time Analyzed: 21:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.704	6.605	6.805	57.920	50.000	15.8
4,4'-DDE	6.194	6.096	6.296	54.700	50.000	9.4
4,4'-DDT	7.019	6.921	7.121	42.000	50.000	-16.0
Aldrin	5.270	5.171	5.371	56.890	50.000	13.8
Decachlorobiphenyl	9.070	8.972	9.172	50.550	50.000	1.1
Dieldrin	6.345	6.247	6.447	55.790	50.000	11.6
Tetrachloro-m-xylene	3.550	3.450	3.650	57.800	50.000	15.6

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL04 Date Analyzed: 06/27/2025

Lab Sample No.: PSTDCCC050 Data File : PD089202.D Time Analyzed: 21:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.928	5.830	6.030	54.200	50.000	8.4
4,4'-DDE	5.373	5.275	5.475	53.740	50.000	7.5
4,4'-DDT	6.181	6.084	6.284	41.310	50.000	-17.4
Aldrin	4.367	4.269	4.469	55.520	50.000	11.0
Decachlorobiphenyl	8.070	7.972	8.172	50.420	50.000	0.8
Dieldrin	5.511	5.413	5.613	53.700	50.000	7.4
Tetrachloro-m-xylene	2.881	2.780	2.980	58.420	50.000	16.8

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: **POWE02**

Lab Code: **CHEM** Case No.: **Q2425** SAS No.: **Q2425** SDG NO.: **Q2425**

GC Column: **ZB-MR1** ID: **0.32** (mm) Initi. Calib. Date(s): **06/17/2025** **06/17/2025**

Client Sample No. (PEM): **PEM - PD088991.D** Date Analyzed: **06/17/2025**

Lab Sample No.(PEM): **PEM** Time Analyzed: **15:25**

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.071	8.970	9.170	21.440	20.000	7.2
Tetrachloro-m-xylene	3.550	3.500	3.600	19.860	20.000	-0.7
alpha-BHC	3.999	3.950	4.050	8.970	10.000	-10.3
beta-BHC	4.514	4.460	4.560	9.720	10.000	-2.8
gamma-BHC (Lindane)	4.330	4.280	4.380	9.320	10.000	-6.8
Endrin	6.574	6.500	6.640	50.260	50.000	0.5
4,4'-DDT	7.020	6.950	7.090	100.820	100.000	0.8
Methoxychlor	7.493	7.420	7.560	229.880	250.000	-8.0

GC Column: **ZB-MR2** ID: **0.32** (mm) Initi. Calib. Date(s): **06/17/2025** **06/17/2025**

Client Sample No. (PEM): **PEM - PD088991.D** Date Analyzed: **06/17/2025**

Lab Sample No.(PEM): **PEM** Time Analyzed: **15:25**

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.072	7.970	8.170	21.290	20.000	6.5
Tetrachloro-m-xylene	2.880	2.830	2.930	20.440	20.000	2.2
alpha-BHC	3.393	3.340	3.440	10.570	10.000	5.7
beta-BHC	4.025	3.970	4.080	10.750	10.000	7.5
gamma-BHC (Lindane)	3.729	3.680	3.780	10.580	10.000	5.8
Endrin	5.789	5.720	5.860	50.490	50.000	1.0
4,4'-DDT	6.183	6.110	6.250	96.080	100.000	-3.9
Methoxychlor	6.754	6.680	6.820	198.860	250.000	-20.5

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: **POWE02**

Lab Code: **CHEM** Case No.: **Q2425** SAS No.: **Q2425** SDG NO.: **Q2425**

GC Column: **ZB-MR1** ID: **0.32** (mm) Initi. Calib. Date(s): **06/17/2025** **06/17/2025**

Client Sample No. (PEM): **PEM - PD089158.D** Date Analyzed: **06/26/2025**

Lab Sample No.(PEM): **PEM** Time Analyzed: **11:33**

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	9.072	8.970	9.170	20.370	20.000	1.9
Tetrachloro-m-xylene	3.550	3.500	3.600	20.980	20.000	4.9
alpha-BHC	3.999	3.950	4.050	9.740	10.000	-2.6
beta-BHC	4.516	4.470	4.570	10.730	10.000	7.3
gamma-BHC (Lindane)	4.331	4.280	4.380	10.120	10.000	1.2
Endrin	6.574	6.500	6.640	48.910	50.000	-2.2
4,4'-DDT	7.021	6.950	7.090	92.590	100.000	-7.4
Methoxychlor	7.493	7.420	7.560	206.780	250.000	-17.3

GC Column: **ZB-MR2** ID: **0.32** (mm) Initi. Calib. Date(s): **06/17/2025** **06/17/2025**

Client Sample No. (PEM): **PEM - PD089158.D** Date Analyzed: **06/26/2025**

Lab Sample No.(PEM): **PEM** Time Analyzed: **11:33**

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	8.071	7.970	8.170	20.100	20.000	0.5
Tetrachloro-m-xylene	2.880	2.830	2.930	21.410	20.000	7.1
alpha-BHC	3.393	3.340	3.440	11.020	10.000	10.2
beta-BHC	4.024	3.970	4.070	10.860	10.000	8.6
gamma-BHC (Lindane)	3.729	3.680	3.780	11.120	10.000	11.2
Endrin	5.788	5.720	5.860	48.080	50.000	-3.8
4,4'-DDT	6.182	6.110	6.250	89.550	100.000	-10.5
Methoxychlor	6.752	6.680	6.820	175.300	250.000	-29.9

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No. (PEM): PEM - PD089191.D Date Analyzed: 06/27/2025

Lab Sample No.(PEM): PEM Time Analyzed: 18:32

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	9.071	8.970	9.170	20.920	20.000	4.6
Tetrachloro-m-xylene	3.550	3.500	3.600	22.080	20.000	10.4
alpha-BHC	3.998	3.950	4.050	10.090	10.000	0.9
beta-BHC	4.515	4.460	4.570	11.260	10.000	12.6
gamma-BHC (Lindane)	4.328	4.280	4.380	10.260	10.000	2.6
Endrin	6.572	6.500	6.640	51.080	50.000	2.2
4,4'-DDT	7.020	6.950	7.090	80.720	100.000	-19.3
Methoxychlor	7.492	7.420	7.560	182.490	250.000	-27.0

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No. (PEM): PEM - PD089191.D Date Analyzed: 06/27/2025

Lab Sample No.(PEM): PEM Time Analyzed: 18:32

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	8.071	7.970	8.170	20.430	20.000	2.2
Tetrachloro-m-xylene	2.880	2.830	2.930	23.060	20.000	15.3
alpha-BHC	3.392	3.340	3.440	11.830	10.000	18.3
beta-BHC	4.025	3.970	4.080	10.760	10.000	7.6
gamma-BHC (Lindane)	3.729	3.680	3.780	11.860	10.000	18.6
Endrin	5.788	5.720	5.860	53.330	50.000	6.7
4,4'-DDT	6.182	6.110	6.250	80.360	100.000	-19.6
Methoxychlor	6.753	6.680	6.820	160.570	250.000	-35.8

Analytical Sequence

Client: Kleinfelder	SDG No.: Q2425
Project: AS Jenks School	Instrument ID: ECD_D
GC Column: ZB-MR1	ID: 0.32 (mm) Inst. Calib. Date(s): 06/17/2025 06/17/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	06/17/2025	15:11	PD088990.D	9.07	3.55
PEM	PEM	06/17/2025	15:25	PD088991.D	9.07	3.55
RESCHK	RESCHK	06/17/2025	15:39	PD088992.D	9.07	3.55
PSTDICC100	PSTDICC100	06/17/2025	15:52	PD088993.D	9.07	3.55
PSTDICC075	PSTDICC075	06/17/2025	16:06	PD088994.D	9.07	3.55
PSTDICC050	PSTDICC050	06/17/2025	16:20	PD088995.D	9.07	3.55
PSTDICC025	PSTDICC025	06/17/2025	16:33	PD088996.D	9.07	3.55
PSTDICC005	PSTDICC005	06/17/2025	16:47	PD088997.D	9.07	3.55
PCHLORICC500	PCHLORICC500	06/17/2025	17:28	PD089000.D	9.07	3.55
PTOXICC500	PTOXICC500	06/17/2025	18:36	PD089005.D	9.07	3.55
IBLK	IBLK	06/26/2025	11:20	PD089157.D	9.07	3.55
PEM	PEM	06/26/2025	11:33	PD089158.D	9.07	3.55
PSTDCCC050	PSTDCCC050	06/26/2025	11:47	PD089159.D	9.07	3.55
PB168623BL	PB168623BL	06/26/2025	13:44	PD089166.D	9.07	3.55
PB168623BS	PB168623BS	06/26/2025	14:59	PD089167.D	9.08	3.56
IBLK	IBLK	06/26/2025	16:35	PD089171.D	9.08	3.56
PSTDCCC050	PSTDCCC050	06/26/2025	17:20	PD089172.D	9.08	3.56
IBLK	IBLK	06/27/2025	18:19	PD089190.D	9.07	3.55
PEM	PEM	06/27/2025	18:32	PD089191.D	9.07	3.55
PSTDCCC050	PSTDCCC050	06/27/2025	18:46	PD089192.D	9.07	3.55
COMP-1	Q2425-01	06/27/2025	19:41	PD089195.D	9.07	3.55
COMP-2	Q2425-02	06/27/2025	19:54	PD089196.D	9.07	3.55
COMP-3	Q2425-03	06/27/2025	20:08	PD089197.D	9.07	3.55
COMP-3MS	Q2425-03MS	06/27/2025	20:22	PD089198.D	9.07	3.55
COMP-3MSD	Q2425-03MSD	06/27/2025	20:35	PD089199.D	9.07	3.55
IBLK	IBLK	06/27/2025	21:03	PD089201.D	9.07	3.55
PSTDCCC050	PSTDCCC050	06/27/2025	21:16	PD089202.D	9.07	3.55

Analytical Sequence

Client: Kleinfelder	SDG No.: Q2425
Project: AS Jenks School	Instrument ID: ECD_D
GC Column: ZB-MR2	ID: 0.32 (mm) Inst. Calib. Date(s): 06/17/2025 06/17/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	06/17/2025	15:11	PD088990.D	8.07	2.88
PEM	PEM	06/17/2025	15:25	PD088991.D	8.07	2.88
RESCHK	RESCHK	06/17/2025	15:39	PD088992.D	8.07	2.88
PSTDICC100	PSTDICC100	06/17/2025	15:52	PD088993.D	8.07	2.88
PSTDICC075	PSTDICC075	06/17/2025	16:06	PD088994.D	8.07	2.88
PSTDICC050	PSTDICC050	06/17/2025	16:20	PD088995.D	8.07	2.88
PSTDICC025	PSTDICC025	06/17/2025	16:33	PD088996.D	8.07	2.88
PSTDICC005	PSTDICC005	06/17/2025	16:47	PD088997.D	8.07	2.88
PCHLORICC500	PCHLORICC500	06/17/2025	17:28	PD089000.D	8.07	2.88
PTOXICC500	PTOXICC500	06/17/2025	18:36	PD089005.D	8.07	2.88
IBLK	IBLK	06/26/2025	11:20	PD089157.D	8.07	2.88
PEM	PEM	06/26/2025	11:33	PD089158.D	8.07	2.88
PSTDCCC050	PSTDCCC050	06/26/2025	11:47	PD089159.D	8.07	2.88
PB168623BL	PB168623BL	06/26/2025	13:44	PD089166.D	8.07	2.88
PB168623BS	PB168623BS	06/26/2025	14:59	PD089167.D	8.07	2.88
IBLK	IBLK	06/26/2025	16:35	PD089171.D	8.07	2.88
PSTDCCC050	PSTDCCC050	06/26/2025	17:20	PD089172.D	8.07	2.88
IBLK	IBLK	06/27/2025	18:19	PD089190.D	8.07	2.88
PEM	PEM	06/27/2025	18:32	PD089191.D	8.07	2.88
PSTDCCC050	PSTDCCC050	06/27/2025	18:46	PD089192.D	8.07	2.88
COMP-1	Q2425-01	06/27/2025	19:41	PD089195.D	8.07	2.88
COMP-2	Q2425-02	06/27/2025	19:54	PD089196.D	8.07	2.88
COMP-3	Q2425-03	06/27/2025	20:08	PD089197.D	8.07	2.88
COMP-3MS	Q2425-03MS	06/27/2025	20:22	PD089198.D	8.07	2.88
COMP-3MSD	Q2425-03MSD	06/27/2025	20:35	PD089199.D	8.07	2.88
IBLK	IBLK	06/27/2025	21:03	PD089201.D	8.07	2.88
PSTDCCC050	PSTDCCC050	06/27/2025	21:16	PD089202.D	8.07	2.88

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

COMP-3MS

Contract: POWE02

Lab Code: CHEM Case No.: Q2425

SAS No.: Q2425 SDG NO.: Q2425

Lab Sample ID: Q2425-03MS

Date(s) Analyzed: 06/27/2025 06/27/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDD	1	6.70	6.65	6.75	18.6	4.4
	2	5.93	5.88	5.98	17.8	
4,4'-DDT	1	7.02	6.97	7.07	11.5	14.5
	2	6.18	6.13	6.23	13.3	
Aldrin	1	5.27	5.22	5.32	17.6	1.7
	2	4.37	4.32	4.42	17.9	
4,4'-DDE	1	6.20	6.15	6.25	17.2	2.9
	2	5.37	5.32	5.42	17.7	
Dieldrin	1	6.35	6.30	6.40	17.4	1.1
	2	5.51	5.46	5.56	17.6	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

COMP-3MSD

Contract: POWE02

Lab Code: CHEM Case No.: Q2425

SAS No.: Q2425 SDG NO.: Q2425

Lab Sample ID: Q2425-03MSD

Date(s) Analyzed: 06/27/2025 06/27/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDD	1	6.70	6.65	6.75	17.7	3.4
	2	5.93	5.88	5.98	17.1	
4,4'-DDT	1	7.02	6.97	7.07	10.2	23.4
	2	6.18	6.13	6.23	12.9	
Aldrin	1	5.27	5.22	5.32	18.2	0.5
	2	4.37	4.32	4.42	18.3	
4,4'-DDE	1	6.20	6.15	6.25	18.2	1.1
	2	5.37	5.32	5.42	18.4	
Dieldrin	1	6.35	6.30	6.40	17.8	0.6
	2	5.51	5.46	5.56	17.9	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB168623BS

Contract: POWE02

Lab Code: CHEM Case No.: Q2425

SAS No.: Q2425 SDG NO.: Q2425

Lab Sample ID: PB168623BS

Date(s) Analyzed: 06/26/2025 06/26/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDD	1	6.71	6.66	6.76	16.1	0.6
	2	5.93	5.88	5.98	16.0	
4,4'-DDE	1	6.20	6.15	6.25	15.9	2.5
	2	5.38	5.33	5.43	15.5	
4,4'-DDT	1	7.03	6.98	7.08	13.9	0
	2	6.18	6.13	6.23	13.9	
Aldrin	1	5.28	5.23	5.33	16.2	1.9
	2	4.37	4.32	4.42	15.9	
Dieldrin	1	6.35	6.30	6.40	15.9	2.5
	2	5.51	5.46	5.56	15.5	

LAB CHRONICLE

OrderID:	Q2425	OrderDate:	6/25/2025 2:03:00 PM
Client:	Kleinfelder	Project:	AS Jenks School
Contact:	Mark Warchol	Location:	D51,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2425-01	COMP-1	SOIL	PCB Group1	8082A	06/24/25	06/26/25	06/26/25	06/25/25
Q2425-02	COMP-2	SOIL	PCB Group1	8082A	06/24/25	06/26/25	06/26/25	06/25/25
Q2425-03	COMP-3	SOIL	PCB Group1	8082A	06/24/25	06/26/25	06/26/25	06/25/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.:	Q2425	Order ID:	Q2425
Client:	Kleinfelder	Project ID:	AS Jenks School

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :				Total Concentration:	0.000			



SAMPLE DATA

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-1	SDG No.:	Q2425
Lab Sample ID:	Q2425-01	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	76.9
Sample Wt/Vol:	30.01	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO111866.D	1	06/26/25 08:20	06/26/25 13:12	PB168622

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	5.10	U	5.10	22.1	ug/kg
11097-69-1	Aroclor-1254	4.20	U	4.20	22.1	ug/kg
11096-82-5	Aroclor-1260	4.20	U	4.20	22.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.4		32 - 144	82%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.5		32 - 175	82%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-2	SDG No.:	Q2425
Lab Sample ID:	Q2425-02	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	79.7
Sample Wt/Vol:	30.04	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO111867.D	1	06/26/25 08:20	06/26/25 13:30	PB168622

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.90	U	4.90	21.3	ug/kg
11097-69-1	Aroclor-1254	4.00	U	4.00	21.3	ug/kg
11096-82-5	Aroclor-1260	4.00	U	4.00	21.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.6		32 - 144	93%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.8		32 - 175	84%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-3	SDG No.:	Q2425
Lab Sample ID:	Q2425-03	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	82.6
Sample Wt/Vol:	30.07	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO111868.D	1	06/26/25 08:20	06/26/25 13:48	PB168622

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.80	U	4.80	20.5	ug/kg
11097-69-1	Aroclor-1254	3.90	U	3.90	20.5	ug/kg
11096-82-5	Aroclor-1260	3.90	U	3.90	20.5	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.1		32 - 144	106%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.7		32 - 175	98%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit



QC SUMMARY

Surrogate Summary

SDG No.: Q2425

Client: Kleinfelder

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO111586.D	PIBLK-PO111586.D	Tetrachloro-m-xylene	1	20	20.3	102		60	140
		Decachlorobiphenyl	1	20	20.7	104		60	140
		Tetrachloro-m-xylene	2	20	19.0	95		60	140
		Decachlorobiphenyl	2	20	20.5	103		60	140
I.BLK-PO111862.D	PIBLK-PO111862.D	Tetrachloro-m-xylene	1	20	20.5	102		60	140
		Decachlorobiphenyl	1	20	19.8	99		60	140
		Tetrachloro-m-xylene	2	20	18.4	92		60	140
		Decachlorobiphenyl	2	20	19.4	97		60	140
Q2425-01	COMP-1	Tetrachloro-m-xylene	1	20	16.4	82		32	144
		Decachlorobiphenyl	1	20	15.8	79		32	175
		Tetrachloro-m-xylene	2	20	14.6	73		32	144
		Decachlorobiphenyl	2	20	16.5	82		32	175
Q2425-02	COMP-2	Tetrachloro-m-xylene	1	20	18.6	93		32	144
		Decachlorobiphenyl	1	20	16.5	83		32	175
		Tetrachloro-m-xylene	2	20	16.9	85		32	144
		Decachlorobiphenyl	2	20	16.8	84		32	175
Q2425-03	COMP-3	Tetrachloro-m-xylene	1	20	21.1	106		32	144
		Decachlorobiphenyl	1	20	19.2	96		32	175
		Tetrachloro-m-xylene	2	20	19.4	97		32	144
		Decachlorobiphenyl	2	20	19.7	98		32	175
I.BLK-PO111875.D	PIBLK-PO111875.D	Tetrachloro-m-xylene	1	20	20.0	100		60	140
		Decachlorobiphenyl	1	20	19.2	96		60	140
		Tetrachloro-m-xylene	2	20	18.1	90		60	140
		Decachlorobiphenyl	2	20	19.5	97		60	140
I.BLK-PP072990.D	PIBLK-PP072990.D	Tetrachloro-m-xylene	1	20	17.3	86		60	140
		Decachlorobiphenyl	1	20	18.3	91		60	140
		Tetrachloro-m-xylene	2	20	18.0	90		60	140
		Decachlorobiphenyl	2	20	17.2	86		60	140
I.BLK-PP073283.D	PIBLK-PP073283.D	Tetrachloro-m-xylene	1	20	18.9	94		60	140
		Decachlorobiphenyl	1	20	18.6	93		60	140
		Tetrachloro-m-xylene	2	20	18.8	94		60	140
		Decachlorobiphenyl	2	20	20.9	105		60	140
PB168622BL	PB168622BL	Tetrachloro-m-xylene	1	20	19.3	96		32	144
		Decachlorobiphenyl	1	20	20.0	100		32	175
		Tetrachloro-m-xylene	2	20	18.8	94		32	144
		Decachlorobiphenyl	2	20	21.1	106		32	175
PB168622BS	PB168622BS	Tetrachloro-m-xylene	1	20	20.2	101		32	144
		Decachlorobiphenyl	1	20	20.0	100		32	175
		Tetrachloro-m-xylene	2	20	18.6	93		32	144
		Decachlorobiphenyl	2	20	23.0	115		32	175
Q2409-02MS	COP-SOIL-PILEMS	Tetrachloro-m-xylene	1	20	20.0	100		32	144

Surrogate Summary

SDG No.: Q2425

Client: Kleinfelder

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q2409-02MS	COP-SOIL-PILEMS	Decachlorobiphenyl	1	20	22.5	112		32	175
		Tetrachloro-m-xylene	2	20	18.9	94		32	144
		Decachlorobiphenyl	2	20	22.7	113		32	175
Q2409-02MSD	COP-SOIL-PILEMSD	Tetrachloro-m-xylene	1	20	19.3	97		32	144
		Decachlorobiphenyl	1	20	23.0	115		32	175
		Tetrachloro-m-xylene	2	20	19.2	96		32	144
		Decachlorobiphenyl	2	20	22.7	114		32	175
		Tetrachloro-m-xylene	1	20	18.3	91		60	140
I.BLK-PP073298.D	PIBLK-PP073298.D	Decachlorobiphenyl	1	20	18.9	95		60	140
		Tetrachloro-m-xylene	2	20	18.7	93		60	140
		Decachlorobiphenyl	2	20	20.5	103		60	140

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2425</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Kleinfelder</u>	DataFile :	<u>PP073292.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2409-02MS (Column 1)	COP-SOIL-PILEMS AR1016	172.7	0	163	ug/kg	94				55	146	
	AR1260	172.7	31	195	ug/kg	95				54	119	
Client Sample ID: Q2409-02MS (Column 2)	COP-SOIL-PILEMS AR1016	172.7	0	154	ug/kg	89				55	146	
	AR1260	172.7	28	169	ug/kg	98				54	119	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2425</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Kleinfelder</u>	DataFile :	<u>PP073293.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2409-02MSD (Column 1)	COP-SOIL-PILEMSD AR1016	172.5	0	163	ug/kg	94		0		55	146	15
	AR1260	172.5	31	202	ug/kg	99		4		54	119	15
Client Sample ID: Q2409-02MSD (Column 2)	COP-SOIL-PILEMSD AR1016	172.5	0	158	ug/kg	92		3		55	146	15
	AR1260	172.5	28	174	ug/kg	101		3		54	119	15

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2425</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Kleinfelder</u>	Datafile :	<u>PP073290.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB168622BS (Column 1)	AR1016	166.6	160	ug/kg	96				71	120		
	AR1260	166.6	159	ug/kg	95				65	130		
PB168622BS (Column 2)	AR1016	166.6	149	ug/kg	89				71	120		
	AR1260	166.6	157	ug/kg	94				65	130		

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB168622BL

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM Case No.: Q2425

SAS No.: Q2425 SDG NO.: Q2425

Lab Sample ID: PB168622BL

Lab File ID: PP073289.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 06/26/2025

Date Analyzed (1): 06/26/2025

Date Analyzed (2): 06/26/2025

Time Analyzed (1): 12:25

Time Analyzed (2): 12:25

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
COMP-1	Q2425-01	PO111866.D	06/26/2025	06/26/2025
COMP-2	Q2425-02	PO111867.D	06/26/2025	06/26/2025
COMP-3	Q2425-03	PO111868.D	06/26/2025	06/26/2025
PB168622BS	PB168622BS	PP073290.D	06/26/2025	06/26/2025
COP-SOIL-PILEMS	Q2409-02MS	PP073292.D	06/26/2025	06/26/2025
COP-SOIL-PILEMSD	Q2409-02MSD	PP073293.D	06/26/2025	06/26/2025

COMMENTS: _____



CALIBRATION SUMMARY

RETENTION TIMES OF INITIAL CALIBRATION

Contract:	POWE02						
Lab Code:	CHEM	Case No.:	Q2425	SAS No.:	Q2425	SDG NO.:	Q2425
Instrument ID:	ECD_O	Calibration Date(s):		06/11/2025	06/11/2025		
		Calibration Times:		10:40	18:31		

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PO111587.D RT 750 = PO111588.D
RT 500 = PO111589.D RT 250 = PO111590.D RT 050 = PO111591.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Contract:	POWE02						
Lab Code:	CHEM	Case No.:	Q2425	SAS No.:	Q2425	SDG NO.:	Q2425
Instrument ID:	ECD_O	Calibration Date(s):		06/11/2025	06/11/2025		
		Calibration Times:		10:40	18:31		

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PO111587.D RT 750 = PO111588.D
RT 500 = PO111589.D RT 250 = PO111590.D RT 050 = PO111591.D

[illegible]

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM **Case No.:** Q2425 **SAS No.:** Q2425 **SDG NO.:** Q2425

Instrument ID: ECD_O **Calibration Date(s):** 06/11/2025 06/11/2025
Calibration Times: 10:40 18:31

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 =	PO111587.D	CF 750 =	PO111588.D			
CF 500 =		PO111589.D	CF 250 =	PO111590.D	CF 050 =	PO111591.D		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	220511854	230179169	244104994	255923844	242047420	238553456	6
Aroclor-1016-2	(2)	319525390	333752768	345607420	360231916	304600680	332743635	7
Aroclor-1016-3	(3)	218699222	228973452	241664208	255477324	205889780	230140797	8
Aroclor-1016-4	(4)	174886969	182974663	191580278	204290796	173581000	185462741	7
Aroclor-1016-5	(5)	177464979	184697736	193390828	207039800	170624200	186643509	8
Aroclor-1260-1	(1)	321794228	337454645	358028956	426721964	332609080	355321775	12
Aroclor-1260-2	(2)	433743230	452724105	474832224	506755684	494307920	472472633	6
Aroclor-1260-3	(3)	393522712	410790572	429624930	441759844	447671720	424673956	5
Aroclor-1260-4	(4)	282329615	299358140	315388464	327883556	347275780	314447111	8
Aroclor-1260-5	(5)	781713901	810719289	832144010	867226572	826835640	823727882	4
Decachlorobiphenyl		4901811500	5109294200	5313242720	5581051480	5340821400	5249244260	5
Tetrachloro-m-xylene		5558158730	5743696387	5927517640	6035511000	5532196200	5759415991	4
Aroclor-1254-1	(1)	375291496	390445984	409053510	418326128	532236220	425070668	15
Aroclor-1254-2	(2)	334554268	347730616	365467502	374564136	475325660	379528436	15
Aroclor-1254-3	(3)	528273778	541799637	566990470	566909168	652766900	571347991	8
Aroclor-1254-4	(4)	330811190	344690741	356705876	353681912	358524400	348882824	3
Aroclor-1254-5	(5)	481176015	491854427	511981248	504869196	570769000	512129977	7
Decachlorobiphenyl		5009210160	5167761240	5403649020	5380598200	6191612600	5430566244	8
Tetrachloro-m-xylene		5724223470	6030704747	6021777640	5926293200	6653717200	6071343251	6

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM **Case No.:** Q2425 **SAS No.:** Q2425 **SDG NO.:** Q2425

Instrument ID: ECD_O **Calibration Date(s):** 06/11/2025 06/11/2025
Calibration Times: 10:40 18:31

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 =	PO111587.D	CF 750 =	PO111588.D			
CF 500 =		PO111589.D	CF 250 =	PO111590.D	CF 050 =	PO111591.D		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	180839726	189231809	196263600	204676936	212823680	196767150	6
Aroclor-1016-2	(2)	269525798	277038755	287808724	296313292	293716580	284880630	4
Aroclor-1016-3	(3)	141401189	146814288	152988496	158088464	151904200	150239327	4
Aroclor-1016-4	(4)	112384684	117918833	124672358	128074364	125886140	121787276	5
Aroclor-1016-5	(5)	145637432	151733320	158595618	164650128	175375940	159198488	7
Aroclor-1260-1	(1)	232933493	242290633	253160914	265400132	260795120	250916058	5
Aroclor-1260-2	(2)	272930024	283658681	293882326	306210008	313403160	294016840	6
Aroclor-1260-3	(3)	248293091	256892847	265761412	279567532	285382000	267179376	6
Aroclor-1260-4	(4)	171150736	180595663	190038668	201212768	207483680	190096303	8
Aroclor-1260-5	(5)	395557404	413148045	424269490	442012232	450112560	425019946	5
Decachlorobiphenyl		1664373530	1742647173	1808169220	1884747160	1788371000	1777661617	5
Tetrachloro-m-xylene		5585912110	5701804613	5792994800	5795065440	5210609600	5617277313	4
Aroclor-1254-1	(1)	296936934	303927659	319123596	320561628	374491800	323008323	9
Aroclor-1254-2	(2)	255453905	262352463	275811104	277251804	325698480	279313551	10
Aroclor-1254-3	(3)	392851690	400205872	416928596	411547332	473734340	419053566	8
Aroclor-1254-4	(4)	214635691	222689013	230857806	231773900	239050800	227801442	4
Aroclor-1254-5	(5)	301414788	308036695	323788090	319823096	374889540	325590442	9
Decachlorobiphenyl		1681329490	1723219680	1812787460	1794853560	2027152800	1807868598	7
Tetrachloro-m-xylene		5554099840	5630982267	5714907100	5492046840	5804310800	5639269369	2

A
B
C
D
E
F
G

Calibration Times: 10:04 19:37

LAB FILE ID: RT 1000 = PP072991.D RT 750 = PP072992.D
RT 500 = PP072993.D RT 250 = PP072994.D RT 050 = PP072995.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM

Case No.: Q2425

SAS No.: Q2425

SDG NO.: Q2425

Instrument ID: ECD_P

Calibration Date(s): 06/17/2025 06/17/2025

Calibration Times: 10:04 19:37

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PP072991.D RT 750 = PP072992.D
RT 500 = PP072993.D RT 250 = PP072994.D RT 050 = PP072995.D

[illegible]

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM **Case No.:** Q2425 **SAS No.:** Q2425 **SDG NO.:** Q2425

Instrument ID: ECD_P **Calibration Date(s):** 06/17/2025 06/17/2025
Calibration Times: 10:04 19:37

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 =	PP072991.D	CF 750 =	PP072992.D			
CF 500 =		PP072993.D	CF 250 =	PP072994.D	CF 050 =	PP072995.D		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	63671603	67644257	70148492	76963128	84930220	72671540	12
Aroclor-1016-2	(2)	97440661	102354848	107008936	114458460	98840440	104020669	7
Aroclor-1016-3	(3)	59983524	63267185	65041372	73554480	64895380	65348388	8
Aroclor-1016-4	(4)	50589003	52912481	54400336	56287652	47214800	52280854	7
Aroclor-1016-5	(5)	46240143	47840919	49349310	51433860	55161280	50005102	7
Aroclor-1260-1	(1)	82538787	86792979	90284630	94211564	87238820	88213356	5
Aroclor-1260-2	(2)	125600539	132072107	137363486	143645412	140737120	135883733	5
Aroclor-1260-3	(3)	106201830	111185451	115341698	120182220	117849880	114152216	5
Aroclor-1260-4	(4)	98670506	102821608	101520658	108282220	103932620	103045522	3
Aroclor-1260-5	(5)	227637841	236485680	245198018	255372264	238418920	240622545	4
Decachlorobiphenyl		1620648950	1677444467	1749703060	1784234960	1712434400	1708893167	4
Tetrachloro-m-xylene		1977356170	2059482187	2109826840	2243117400	2164249200	2110806359	5
Aroclor-1254-1	(1)	73763285	78885443	81757104	88891384	97572240	84173891	11
Aroclor-1254-2	(2)	111451506	118674340	123357828	133620984	139899720	125400876	9
Aroclor-1254-3	(3)	119235877	126285657	130365552	141133040	139034740	131210973	7
Aroclor-1254-4	(4)	108754762	114515340	117946542	128101128	129602040	119783962	7
Aroclor-1254-5	(5)	106940669	112254864	115991910	123909512	112185340	114256459	6
Decachlorobiphenyl		1648218000	1739306053	1775656020	1935470960	1659491600	1751628527	7
Tetrachloro-m-xylene		1872735760	1960277133	2019595300	2159825040	2074265000	2017339647	5

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM **Case No.:** Q2425 **SAS No.:** Q2425 **SDG NO.:** Q2425

Instrument ID: ECD_P **Calibration Date(s):** 06/17/2025 06/17/2025
Calibration Times: 10:04 19:37

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 =	PP072991.D	CF 750 =	PP072992.D			
CF 500 =		PP072993.D	CF 250 =	PP072994.D	CF 050 =	PP072995.D		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	58734267	64000205	67674152	72695560	80484240	68717685	12
Aroclor-1016-2	(2)	90382349	94344453	98038690	104433472	115993580	100638509	10
Aroclor-1016-3	(3)	48136215	50748025	53595712	56975444	64125620	54716203	11
Aroclor-1016-4	(4)	38107893	40907291	43245840	46954372	53532680	44549615	13
Aroclor-1016-5	(5)	49203503	51615780	54910048	58505108	62177080	55282304	9
Aroclor-1260-1	(1)	84892100	88579871	95448406	100236824	104495860	94730612	9
Aroclor-1260-2	(2)	105588352	108160505	115763440	121715240	130953840	116436275	9
Aroclor-1260-3	(3)	96095206	97763095	104346708	109521728	113078520	104161051	7
Aroclor-1260-4	(4)	77788306	79915727	86740682	92092068	102155800	87738517	11
Aroclor-1260-5	(5)	196041650	198840268	209548214	218270728	244301260	213400424	9
Decachlorobiphenyl		1225771100	1248497173	1342901640	1435827000	1420581600	1334715703	7
Tetrachloro-m-xylene		1662896120	1767446653	1788437460	1893713440	1896490200	1801796775	5
Aroclor-1254-1	(1)	94131500	102369100	108358230	120483728	122907600	109650032	11
Aroclor-1254-2	(2)	80645865	88030495	92596756	103733056	107725280	94546290	12
Aroclor-1254-3	(3)	128776319	138178049	147317104	161124892	162339800	147547233	10
Aroclor-1254-4	(4)	83562949	90268741	96675708	106896836	108101020	97101051	11
Aroclor-1254-5	(5)	116424396	122854644	129598642	142067368	138868340	129962678	8
Decachlorobiphenyl		1234350420	1314370427	1354328000	1438665160	1427909200	1353924641	6
Tetrachloro-m-xylene		1670033590	1775346533	1836392320	1937020080	1788575000	1801473505	5

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 10:06 Initial Calibration Time(s): 10:40 18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.77	4.77	4.67	4.87	0.00
Aroclor-1016-2	(2)	4.79	4.78	4.68	4.88	-0.01
Aroclor-1016-3	(3)	4.84	4.84	4.74	4.94	0.00
Aroclor-1016-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5	(5)	5.22	5.22	5.12	5.32	0.00
Aroclor-1260-1	(1)	6.26	6.26	6.16	6.36	0.00
Aroclor-1260-2	(2)	6.45	6.45	6.35	6.55	0.00
Aroclor-1260-3	(3)	6.81	6.81	6.71	6.91	0.00
Aroclor-1260-4	(4)	7.07	7.07	6.97	7.17	0.00
Aroclor-1260-5	(5)	7.32	7.32	7.22	7.42	0.00
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.71	8.71	8.61	8.81	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 10:06 Initial Calibration Time(s): 10:40 18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.75	4.75	4.65	4.85	0.00
Aroclor-1016-2	(2)	4.77	4.77	4.67	4.87	0.00
Aroclor-1016-3	(3)	4.94	4.94	4.84	5.04	0.00
Aroclor-1016-4	(4)	4.98	4.99	4.89	5.09	0.01
Aroclor-1016-5	(5)	5.20	5.20	5.10	5.30	0.00
Aroclor-1260-1	(1)	6.23	6.23	6.13	6.33	0.00
Aroclor-1260-2	(2)	6.41	6.42	6.32	6.52	0.01
Aroclor-1260-3	(3)	6.57	6.57	6.47	6.67	0.00
Aroclor-1260-4	(4)	7.04	7.04	6.94	7.14	0.00
Aroclor-1260-5	(5)	7.28	7.28	7.18	7.38	0.00
Tetrachloro-m-xylene		3.67	3.67	3.57	3.77	0.00
Decachlorobiphenyl		8.66	8.66	8.56	8.76	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL01 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PO111858.D Time Analyzed: 10:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.767	4.666	4.866	493.110	500.000	-1.4
Aroclor-1016-2	4.786	4.684	4.884	509.120	500.000	1.8
Aroclor-1016-3	4.842	4.741	4.941	505.140	500.000	1.0
Aroclor-1016-4	4.962	4.861	5.061	503.830	500.000	0.8
Aroclor-1016-5	5.220	5.118	5.318	530.310	500.000	6.1
Aroclor-1260-1	6.258	6.157	6.357	513.670	500.000	2.7
Aroclor-1260-2	6.448	6.346	6.546	553.480	500.000	10.7
Aroclor-1260-3	6.813	6.713	6.913	575.780	500.000	15.2
Aroclor-1260-4	7.074	6.972	7.172	564.630	500.000	12.9
Aroclor-1260-5	7.316	7.216	7.416	545.190	500.000	9.0
Decachlorobiphenyl	8.712	8.612	8.812	48.170	50.000	-3.7
Tetrachloro-m-xylene	3.679	3.577	3.777	53.720	50.000	7.4

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL01 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PO111858.D Time Analyzed: 10:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.747	4.651	4.851	495.750	500.000	-0.9
Aroclor-1016-2	4.766	4.669	4.869	504.470	500.000	0.9
Aroclor-1016-3	4.941	4.844	5.044	501.230	500.000	0.2
Aroclor-1016-4	4.984	4.887	5.087	498.050	500.000	-0.4
Aroclor-1016-5	5.196	5.099	5.299	502.020	500.000	0.4
Aroclor-1260-1	6.226	6.129	6.329	501.900	500.000	0.4
Aroclor-1260-2	6.414	6.316	6.516	526.020	500.000	5.2
Aroclor-1260-3	6.566	6.469	6.669	497.850	500.000	-0.4
Aroclor-1260-4	7.036	6.939	7.139	497.700	500.000	-0.5
Aroclor-1260-5	7.277	7.181	7.381	508.850	500.000	1.8
Decachlorobiphenyl	8.658	8.561	8.761	47.550	50.000	-4.9
Tetrachloro-m-xylene	3.671	3.573	3.773	49.570	50.000	-0.9

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 15:37 Initial Calibration Time(s): 10:40 18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.76	4.77	4.67	4.87	0.01
Aroclor-1016-2	(2)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-3	(3)	4.84	4.84	4.74	4.94	0.00
Aroclor-1016-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5	(5)	5.22	5.22	5.12	5.32	0.00
Aroclor-1260-1	(1)	6.25	6.26	6.16	6.36	0.01
Aroclor-1260-2	(2)	6.44	6.45	6.35	6.55	0.01
Aroclor-1260-3	(3)	6.81	6.81	6.71	6.91	0.00
Aroclor-1260-4	(4)	7.07	7.07	6.97	7.17	0.00
Aroclor-1260-5	(5)	7.31	7.32	7.22	7.42	0.01
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.71	8.71	8.61	8.81	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 15:37 Initial Calibration Time(s): 10:40 18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.75	4.75	4.65	4.85	0.00
Aroclor-1016-2	(2)	4.77	4.77	4.67	4.87	0.00
Aroclor-1016-3	(3)	4.94	4.94	4.84	5.04	0.00
Aroclor-1016-4	(4)	4.98	4.99	4.89	5.09	0.01
Aroclor-1016-5	(5)	5.20	5.20	5.10	5.30	0.00
Aroclor-1260-1	(1)	6.23	6.23	6.13	6.33	0.00
Aroclor-1260-2	(2)	6.41	6.42	6.32	6.52	0.01
Aroclor-1260-3	(3)	6.57	6.57	6.47	6.67	0.00
Aroclor-1260-4	(4)	7.04	7.04	6.94	7.14	0.00
Aroclor-1260-5	(5)	7.28	7.28	7.18	7.38	0.00
Tetrachloro-m-xylene		3.67	3.67	3.57	3.77	0.00
Decachlorobiphenyl		8.66	8.66	8.56	8.76	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL02 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PO111871.D Time Analyzed: 15:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.764	4.666	4.866	475.480	500.000	-4.9
Aroclor-1016-2	4.783	4.684	4.884	494.560	500.000	-1.1
Aroclor-1016-3	4.840	4.741	4.941	485.280	500.000	-2.9
Aroclor-1016-4	4.960	4.861	5.061	484.640	500.000	-3.1
Aroclor-1016-5	5.217	5.118	5.318	505.120	500.000	1.0
Aroclor-1260-1	6.254	6.157	6.357	486.680	500.000	-2.7
Aroclor-1260-2	6.444	6.346	6.546	527.440	500.000	5.5
Aroclor-1260-3	6.811	6.713	6.913	531.610	500.000	6.3
Aroclor-1260-4	7.070	6.972	7.172	535.590	500.000	7.1
Aroclor-1260-5	7.313	7.216	7.416	516.860	500.000	3.4
Decachlorobiphenyl	8.708	8.612	8.812	45.020	50.000	-10.0
Tetrachloro-m-xylene	3.676	3.577	3.777	51.620	50.000	3.2

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL02 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PO111871.D Time Analyzed: 15:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.749	4.651	4.851	492.000	500.000	-1.6
Aroclor-1016-2	4.767	4.669	4.869	498.410	500.000	-0.3
Aroclor-1016-3	4.942	4.844	5.044	493.840	500.000	-1.2
Aroclor-1016-4	4.984	4.887	5.087	490.000	500.000	-2.0
Aroclor-1016-5	5.196	5.099	5.299	486.890	500.000	-2.6
Aroclor-1260-1	6.226	6.129	6.329	482.400	500.000	-3.5
Aroclor-1260-2	6.414	6.316	6.516	502.890	500.000	0.6
Aroclor-1260-3	6.566	6.469	6.669	475.680	500.000	-4.9
Aroclor-1260-4	7.036	6.939	7.139	469.160	500.000	-6.2
Aroclor-1260-5	7.277	7.181	7.381	474.960	500.000	-5.0
Decachlorobiphenyl	8.656	8.561	8.761	46.490	50.000	-7.0
Tetrachloro-m-xylene	3.672	3.573	3.773	48.950	50.000	-2.1

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 09:04 Initial Calibration Time(s): 10:04 19:37

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.64	5.65	5.55	5.75	0.01
Aroclor-1016-2	(2)	5.66	5.67	5.57	5.77	0.01
Aroclor-1016-3	(3)	5.73	5.73	5.63	5.83	0.01
Aroclor-1016-4	(4)	5.82	5.83	5.73	5.93	0.01
Aroclor-1016-5	(5)	6.11	6.12	6.02	6.22	0.01
Aroclor-1260-1	(1)	7.23	7.24	7.14	7.34	0.01
Aroclor-1260-2	(2)	7.49	7.49	7.39	7.59	0.00
Aroclor-1260-3	(3)	7.84	7.85	7.75	7.95	0.01
Aroclor-1260-4	(4)	8.07	8.07	7.97	8.17	0.00
Aroclor-1260-5	(5)	8.38	8.39	8.29	8.49	0.01
Tetrachloro-m-xylene		4.49	4.49	4.39	4.59	0.00
Decachlorobiphenyl		10.18	10.19	10.09	10.29	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 09:04 Initial Calibration Time(s): 10:04 19:37

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.86	4.87	4.77	4.97	0.01
Aroclor-1016-2	(2)	4.88	4.89	4.79	4.99	0.01
Aroclor-1016-3	(3)	5.06	5.06	4.96	5.16	0.00
Aroclor-1016-4	(4)	5.10	5.10	5.00	5.20	0.00
Aroclor-1016-5	(5)	5.31	5.32	5.22	5.42	0.01
Aroclor-1260-1	(1)	6.34	6.35	6.25	6.45	0.01
Aroclor-1260-2	(2)	6.53	6.54	6.44	6.64	0.01
Aroclor-1260-3	(3)	6.68	6.69	6.59	6.79	0.01
Aroclor-1260-4	(4)	7.15	7.16	7.06	7.26	0.01
Aroclor-1260-5	(5)	7.40	7.40	7.30	7.50	0.01
Tetrachloro-m-xylene		3.78	3.79	3.69	3.89	0.01
Decachlorobiphenyl		8.79	8.80	8.70	8.90	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PP073279.D Time Analyzed: 09:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.641	5.546	5.746	485.990	500.000	-2.8
Aroclor-1016-2	5.662	5.567	5.767	533.210	500.000	6.6
Aroclor-1016-3	5.725	5.629	5.829	520.180	500.000	4.0
Aroclor-1016-4	5.822	5.727	5.927	549.860	500.000	10.0
Aroclor-1016-5	6.114	6.019	6.219	524.760	500.000	5.0
Aroclor-1260-1	7.231	7.136	7.336	560.880	500.000	12.2
Aroclor-1260-2	7.485	7.389	7.589	543.950	500.000	8.8
Aroclor-1260-3	7.842	7.747	7.947	544.400	500.000	8.9
Aroclor-1260-4	8.066	7.971	8.171	548.260	500.000	9.7
Aroclor-1260-5	8.384	8.288	8.488	516.290	500.000	3.3
Decachlorobiphenyl	10.181	10.087	10.287	50.830	50.000	1.7
Tetrachloro-m-xylene	4.490	4.394	4.594	51.690	50.000	3.4

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PP073279.D Time Analyzed: 09:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.862	4.768	4.968	473.800	500.000	-5.2
Aroclor-1016-2	4.880	4.786	4.986	482.540	500.000	-3.5
Aroclor-1016-3	5.056	4.962	5.162	483.380	500.000	-3.3
Aroclor-1016-4	5.098	5.004	5.204	477.720	500.000	-4.5
Aroclor-1016-5	5.312	5.218	5.418	546.190	500.000	9.2
Aroclor-1260-1	6.342	6.249	6.449	495.490	500.000	-0.9
Aroclor-1260-2	6.531	6.438	6.638	513.470	500.000	2.7
Aroclor-1260-3	6.683	6.589	6.789	485.640	500.000	-2.9
Aroclor-1260-4	7.152	7.060	7.260	489.320	500.000	-2.1
Aroclor-1260-5	7.395	7.302	7.502	489.940	500.000	-2.0
Decachlorobiphenyl	8.788	8.697	8.897	50.670	50.000	1.3
Tetrachloro-m-xylene	3.783	3.688	3.888	47.840	50.000	-4.3

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 14:37 Initial Calibration Time(s): 10:04 19:37

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.64	5.65	5.55	5.75	0.01
Aroclor-1016-2	(2)	5.67	5.67	5.57	5.77	0.01
Aroclor-1016-3	(3)	5.73	5.73	5.63	5.83	0.00
Aroclor-1016-4	(4)	5.82	5.83	5.73	5.93	0.01
Aroclor-1016-5	(5)	6.12	6.12	6.02	6.22	0.00
Aroclor-1260-1	(1)	7.23	7.24	7.14	7.34	0.01
Aroclor-1260-2	(2)	7.49	7.49	7.39	7.59	0.00
Aroclor-1260-3	(3)	7.84	7.85	7.75	7.95	0.01
Aroclor-1260-4	(4)	8.07	8.07	7.97	8.17	0.00
Aroclor-1260-5	(5)	8.39	8.39	8.29	8.49	0.00
Tetrachloro-m-xylene		4.49	4.49	4.39	4.59	0.00
Decachlorobiphenyl		10.18	10.19	10.09	10.29	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 14:37 Initial Calibration Time(s): 10:04 19:37

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.87	4.87	4.77	4.97	0.01
Aroclor-1016-2	(2)	4.88	4.89	4.79	4.99	0.01
Aroclor-1016-3	(3)	5.06	5.06	4.96	5.16	0.00
Aroclor-1016-4	(4)	5.10	5.10	5.00	5.20	0.00
Aroclor-1016-5	(5)	5.31	5.32	5.22	5.42	0.01
Aroclor-1260-1	(1)	6.34	6.35	6.25	6.45	0.01
Aroclor-1260-2	(2)	6.53	6.54	6.44	6.64	0.01
Aroclor-1260-3	(3)	6.69	6.69	6.59	6.79	0.01
Aroclor-1260-4	(4)	7.16	7.16	7.06	7.26	0.00
Aroclor-1260-5	(5)	7.40	7.40	7.30	7.50	0.00
Tetrachloro-m-xylene		3.79	3.79	3.69	3.89	0.00
Decachlorobiphenyl		8.79	8.80	8.70	8.90	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL04 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PP073294.D Time Analyzed: 14:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.643	5.546	5.746	477.730	500.000	-4.5
Aroclor-1016-2	5.665	5.567	5.767	514.610	500.000	2.9
Aroclor-1016-3	5.727	5.629	5.829	507.950	500.000	1.6
Aroclor-1016-4	5.824	5.727	5.927	534.040	500.000	6.8
Aroclor-1016-5	6.116	6.019	6.219	509.130	500.000	1.8
Aroclor-1260-1	7.233	7.136	7.336	554.370	500.000	10.9
Aroclor-1260-2	7.486	7.389	7.589	530.980	500.000	6.2
Aroclor-1260-3	7.844	7.747	7.947	525.830	500.000	5.2
Aroclor-1260-4	8.068	7.971	8.171	527.090	500.000	5.4
Aroclor-1260-5	8.386	8.288	8.488	505.830	500.000	1.2
Decachlorobiphenyl	10.182	10.087	10.287	51.630	50.000	3.3
Tetrachloro-m-xylene	4.492	4.394	4.594	49.270	50.000	-1.5

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL04 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PP073294.D Time Analyzed: 14:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.865	4.768	4.968	475.950	500.000	-4.8
Aroclor-1016-2	4.882	4.786	4.986	486.130	500.000	-2.8
Aroclor-1016-3	5.059	4.962	5.162	484.990	500.000	-3.0
Aroclor-1016-4	5.100	5.004	5.204	478.600	500.000	-4.3
Aroclor-1016-5	5.314	5.218	5.418	545.720	500.000	9.1
Aroclor-1260-1	6.344	6.249	6.449	518.030	500.000	3.6
Aroclor-1260-2	6.534	6.438	6.638	531.220	500.000	6.2
Aroclor-1260-3	6.685	6.589	6.789	499.440	500.000	-0.1
Aroclor-1260-4	7.155	7.060	7.260	498.560	500.000	-0.3
Aroclor-1260-5	7.396	7.302	7.502	503.650	500.000	0.7
Decachlorobiphenyl	8.790	8.697	8.897	52.420	50.000	4.8
Tetrachloro-m-xylene	3.786	3.688	3.888	49.210	50.000	-1.6

Analytical Sequence

Client: Kleinfelder

SDG No.: Q2425

Project: AS Jenks School

Instrument ID: ECD_O

GC Column: ZB-MR1

ID: 0.32 (mm)

Inst. Calib. Date(s): 06/11/2025

06/11/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	06/11/2025	10:21	PO111586.D	8.71	3.68
AR1660ICC1000	AR1660ICC1000	06/11/2025	10:40	PO111587.D	8.71	3.68
AR1660ICC750	AR1660ICC750	06/11/2025	10:58	PO111588.D	8.71	3.68
AR1660ICC500	AR1660ICC500	06/11/2025	11:17	PO111589.D	8.71	3.68
AR1660ICC250	AR1660ICC250	06/11/2025	11:35	PO111590.D	8.71	3.68
AR1660ICC050	AR1660ICC050	06/11/2025	11:53	PO111591.D	8.71	3.68
AR1221ICC500	AR1221ICC500	06/11/2025	12:12	PO111592.D	8.71	3.68
AR1232ICC500	AR1232ICC500	06/11/2025	12:30	PO111593.D	8.71	3.68
AR1242ICC500	AR1242ICC500	06/11/2025	13:25	PO111596.D	8.71	3.68
AR1248ICC500	AR1248ICC500	06/11/2025	15:14	PO111601.D	8.71	3.68
AR1254ICC1000	AR1254ICC1000	06/11/2025	16:06	PO111604.D	8.71	3.68
AR1254ICC750	AR1254ICC750	06/11/2025	16:25	PO111605.D	8.71	3.68
AR1254ICC500	AR1254ICC500	06/11/2025	16:43	PO111606.D	8.71	3.68
AR1254ICC250	AR1254ICC250	06/11/2025	17:00	PO111607.D	8.71	3.68
AR1254ICC050	AR1254ICC050	06/11/2025	17:18	PO111608.D	8.71	3.68
AR1262ICC500	AR1262ICC500	06/11/2025	17:36	PO111609.D	8.71	3.68
AR1268ICC500	AR1268ICC500	06/11/2025	18:31	PO111612.D	8.71	3.68
AR1660CCC500	AR1660CCC500	06/26/2025	10:06	PO111858.D	8.71	3.68
IBLK	IBLK	06/26/2025	11:20	PO111862.D	8.71	3.68
COMP-1	Q2425-01	06/26/2025	13:12	PO111866.D	8.71	3.67
COMP-2	Q2425-02	06/26/2025	13:30	PO111867.D	8.71	3.68
COMP-3	Q2425-03	06/26/2025	13:48	PO111868.D	8.71	3.67
AR1660CCC500	AR1660CCC500	06/26/2025	15:37	PO111871.D	8.71	3.68
IBLK	IBLK	06/26/2025	16:51	PO111875.D	8.71	3.68
IBLK	IBLK	06/17/2025	09:47	PP072990.D	10.19	4.49
AR1660ICC1000	AR1660ICC1000	06/17/2025	10:04	PP072991.D	10.19	4.49
AR1660ICC750	AR1660ICC750	06/17/2025	10:20	PP072992.D	10.19	4.49
AR1660ICC500	AR1660ICC500	06/17/2025	10:37	PP072993.D	10.19	4.49
AR1660ICC250	AR1660ICC250	06/17/2025	10:53	PP072994.D	10.19	4.49
AR1660ICC050	AR1660ICC050	06/17/2025	11:43	PP072995.D	10.19	4.50
AR1221ICC500	AR1221ICC500	06/17/2025	12:00	PP072996.D	10.18	4.49
AR1232ICC500	AR1232ICC500	06/17/2025	12:16	PP072997.D	10.19	4.50
AR1242ICC500	AR1242ICC500	06/17/2025	15:00	PP073000.D	10.19	4.49
AR1248ICC500	AR1248ICC500	06/17/2025	16:37	PP073005.D	10.19	4.49
AR1254ICC1000	AR1254ICC1000	06/17/2025	17:26	PP073008.D	10.19	4.49
AR1254ICC750	AR1254ICC750	06/17/2025	17:43	PP073009.D	10.19	4.49
AR1254ICC500	AR1254ICC500	06/17/2025	17:59	PP073010.D	10.19	4.49
AR1254ICC250	AR1254ICC250	06/17/2025	18:15	PP073011.D	10.19	4.49
AR1254ICC050	AR1254ICC050	06/17/2025	18:32	PP073012.D	10.19	4.49
AR1262ICC500	AR1262ICC500	06/17/2025	18:48	PP073013.D	10.19	4.49
AR1268ICC500	AR1268ICC500	06/17/2025	19:37	PP073016.D	10.18	4.49
AR1660CCC500	AR1660CCC500	06/26/2025	09:04	PP073279.D	10.18	4.49

Analytical Sequence

I.BLK	I.BLK	06/26/2025	10:09	PP073283.D	10.18	4.49
PB168622BL	PB168622BL	06/26/2025	12:25	PP073289.D	10.18	4.49
PB168622BS	PB168622BS	06/26/2025	12:42	PP073290.D	10.18	4.49
COP-SOIL-PILEMS	Q2409-02MS	06/26/2025	13:15	PP073292.D	10.18	4.49
COP-SOIL-PILEMSD	Q2409-02MSD	06/26/2025	13:31	PP073293.D	10.18	4.49
AR1660CCC500	AR1660CCC500	06/26/2025	14:37	PP073294.D	10.18	4.49
I.BLK	I.BLK	06/26/2025	15:42	PP073298.D	10.18	4.49

Analytical Sequence

Client: Kleinfelder

SDG No.: Q2425

Project: AS Jenks School

Instrument ID: ECD_O

GC Column: ZB-MR2

ID: 0.32 (mm)

Inst. Calib. Date(s): 06/11/2025

06/11/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	06/11/2025	10:21	PO111586.D	8.66	3.67
AR1660ICC1000	AR1660ICC1000	06/11/2025	10:40	PO111587.D	8.66	3.67
AR1660ICC750	AR1660ICC750	06/11/2025	10:58	PO111588.D	8.66	3.67
AR1660ICC500	AR1660ICC500	06/11/2025	11:17	PO111589.D	8.66	3.67
AR1660ICC250	AR1660ICC250	06/11/2025	11:35	PO111590.D	8.66	3.67
AR1660ICC050	AR1660ICC050	06/11/2025	11:53	PO111591.D	8.66	3.67
AR1221ICC500	AR1221ICC500	06/11/2025	12:12	PO111592.D	8.66	3.67
AR1232ICC500	AR1232ICC500	06/11/2025	12:30	PO111593.D	8.66	3.67
AR1242ICC500	AR1242ICC500	06/11/2025	13:25	PO111596.D	8.66	3.67
AR1248ICC500	AR1248ICC500	06/11/2025	15:14	PO111601.D	8.66	3.67
AR1254ICC1000	AR1254ICC1000	06/11/2025	16:06	PO111604.D	8.66	3.67
AR1254ICC750	AR1254ICC750	06/11/2025	16:25	PO111605.D	8.66	3.67
AR1254ICC500	AR1254ICC500	06/11/2025	16:43	PO111606.D	8.66	3.67
AR1254ICC250	AR1254ICC250	06/11/2025	17:00	PO111607.D	8.66	3.67
AR1254ICC050	AR1254ICC050	06/11/2025	17:18	PO111608.D	8.66	3.67
AR1262ICC500	AR1262ICC500	06/11/2025	17:36	PO111609.D	8.66	3.67
AR1268ICC500	AR1268ICC500	06/11/2025	18:31	PO111612.D	8.66	3.67
AR1660CCC500	AR1660CCC500	06/26/2025	10:06	PO111858.D	8.66	3.67
IBLK	IBLK	06/26/2025	11:20	PO111862.D	8.66	3.67
COMP-1	Q2425-01	06/26/2025	13:12	PO111866.D	8.66	3.67
COMP-2	Q2425-02	06/26/2025	13:30	PO111867.D	8.66	3.67
COMP-3	Q2425-03	06/26/2025	13:48	PO111868.D	8.66	3.67
AR1660CCC500	AR1660CCC500	06/26/2025	15:37	PO111871.D	8.66	3.67
IBLK	IBLK	06/26/2025	16:51	PO111875.D	8.66	3.67
IBLK	IBLK	06/17/2025	09:47	PP072990.D	8.80	3.79
AR1660ICC1000	AR1660ICC1000	06/17/2025	10:04	PP072991.D	8.80	3.79
AR1660ICC750	AR1660ICC750	06/17/2025	10:20	PP072992.D	8.80	3.79
AR1660ICC500	AR1660ICC500	06/17/2025	10:37	PP072993.D	8.80	3.79
AR1660ICC250	AR1660ICC250	06/17/2025	10:53	PP072994.D	8.80	3.79
AR1660ICC050	AR1660ICC050	06/17/2025	11:43	PP072995.D	8.80	3.79
AR1221ICC500	AR1221ICC500	06/17/2025	12:00	PP072996.D	8.80	3.79
AR1232ICC500	AR1232ICC500	06/17/2025	12:16	PP072997.D	8.80	3.79
AR1242ICC500	AR1242ICC500	06/17/2025	15:00	PP073000.D	8.80	3.79
AR1248ICC500	AR1248ICC500	06/17/2025	16:37	PP073005.D	8.80	3.79
AR1254ICC1000	AR1254ICC1000	06/17/2025	17:26	PP073008.D	8.80	3.79
AR1254ICC750	AR1254ICC750	06/17/2025	17:43	PP073009.D	8.80	3.79
AR1254ICC500	AR1254ICC500	06/17/2025	17:59	PP073010.D	8.80	3.79
AR1254ICC250	AR1254ICC250	06/17/2025	18:15	PP073011.D	8.80	3.79
AR1254ICC050	AR1254ICC050	06/17/2025	18:32	PP073012.D	8.80	3.79
AR1262ICC500	AR1262ICC500	06/17/2025	18:48	PP073013.D	8.80	3.79
AR1268ICC500	AR1268ICC500	06/17/2025	19:37	PP073016.D	8.80	3.79
AR1660CCC500	AR1660CCC500	06/26/2025	09:04	PP073279.D	8.79	3.78

Analytical Sequence

I.BLK	I.BLK	06/26/2025	10:09	PP073283.D	8.79	3.78
PB168622BL	PB168622BL	06/26/2025	12:25	PP073289.D	8.79	3.79
PB168622BS	PB168622BS	06/26/2025	12:42	PP073290.D	8.79	3.79
COP-SOIL-PILEMS	Q2409-02MS	06/26/2025	13:15	PP073292.D	8.79	3.79
COP-SOIL-PILEMSD	Q2409-02MSD	06/26/2025	13:31	PP073293.D	8.79	3.79
AR1660CCC500	AR1660CCC500	06/26/2025	14:37	PP073294.D	8.79	3.79
I.BLK	I.BLK	06/26/2025	15:42	PP073298.D	8.79	3.78



QC SAMPLE DATA

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	AS Jenks School	Date Received:	
Client Sample ID:	PB168622BL	SDG No.:	Q2425
Lab Sample ID:	PB168622BL	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	100
Sample Wt/Vol:	30.01	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP073289.D	1	06/26/25 08:20	06/26/25 12:25	PB168622

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	3.90	U	3.90	17.0	ug/kg
11097-69-1	Aroclor-1254	3.20	U	3.20	17.0	ug/kg
11096-82-5	Aroclor-1260	3.20	U	3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.3		32 - 144	96%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.1		32 - 175	106%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/11/25
Project:	AS Jenks School	Date Received:	06/11/25
Client Sample ID:	PIBLK-PO111586.D	SDG No.:	Q2425
Lab Sample ID:	I.BLK-PO111586.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO111586.D	1		06/11/25	po061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.0		60 - 140	95%	SPK: 20
2051-24-3	Decachlorobiphenyl	20.5		60 - 140	103%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/26/25
Project:	AS Jenks School	Date Received:	06/26/25
Client Sample ID:	PIBLK-PO111862.D	SDG No.:	Q2425
Lab Sample ID:	I.BLK-PO111862.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO111862.D	1		06/26/25	po062625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.4		60 - 140	92%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.4		60 - 140	97%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/26/25
Project:	AS Jenks School	Date Received:	06/26/25
Client Sample ID:	PIBLK-PO111875.D	SDG No.:	Q2425
Lab Sample ID:	I.BLK-PO111875.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO111875.D	1		06/26/25	po062625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.1		60 - 140	90%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.2		60 - 140	96%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/17/25
Project:	AS Jenks School	Date Received:	06/17/25
Client Sample ID:	PIBLK-PP072990.D	SDG No.:	Q2425
Lab Sample ID:	I.BLK-PP072990.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072990.D	1		06/17/25	pp061725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.3		60 - 140	86%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.2		60 - 140	86%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/26/25
Project:	AS Jenks School	Date Received:	06/26/25
Client Sample ID:	PIBLK-PP073283.D	SDG No.:	Q2425
Lab Sample ID:	I.BLK-PP073283.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP073283.D	1		06/26/25	pp062625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.8		60 - 140	94%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.6		60 - 140	93%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/26/25
Project:	AS Jenks School	Date Received:	06/26/25
Client Sample ID:	PIBLK-PP073298.D	SDG No.:	Q2425
Lab Sample ID:	I.BLK-PP073298.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP073298.D	1		06/26/25	pp062625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.3		60 - 140	91%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.9		60 - 140	95%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	AS Jenks School	Date Received:	
Client Sample ID:	PB168622BS	SDG No.:	Q2425
Lab Sample ID:	PB168622BS	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	100
Sample Wt/Vol:	30.02	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP073290.D	1	06/26/25 08:20	06/26/25 12:42	PB168622

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	160		3.90	17.0	ug/kg
11097-69-1	Aroclor-1254	3.20	U	3.20	17.0	ug/kg
11096-82-5	Aroclor-1260	159		3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.2		32 - 144	101%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.0		32 - 175	115%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/24/25
Client Sample ID:	COP-SOIL-PILEMS	SDG No.:	Q2425
Lab Sample ID:	Q2409-02MS	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	96.4
Sample Wt/Vol:	30.03	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP073292.D	1	06/26/25 08:20	06/26/25 13:15	PB168622

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	163		4.10	17.6	ug/kg
11097-69-1	Aroclor-1254	135		3.30	17.6	ug/kg
11096-82-5	Aroclor-1260	195		3.30	17.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.0		32 - 144	100%	SPK: 20
2051-24-3	Decachlorobiphenyl	22.7		32 - 175	113%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/24/25
Client Sample ID:	COP-SOIL-PILEMSD	SDG No.:	Q2425
Lab Sample ID:	Q2409-02MSD	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	96.4
Sample Wt/Vol:	30.06	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP073293.D	1	06/26/25 08:20	06/26/25 13:31	PB168622

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	163		4.10	17.6	ug/kg
11097-69-1	Aroclor-1254	126		3.30	17.6	ug/kg
11096-82-5	Aroclor-1260	202		3.30	17.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.3		32 - 144	97%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.0		32 - 175	115%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

LAB CHRONICLE

OrderID:	Q2425	OrderDate:	6/25/2025 2:03:00 PM
Client:	Kleinfelder	Project:	AS Jenks School
Contact:	Mark Warchol	Location:	D51,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2425-01	COMP-1	SOIL			06/24/25			06/25/25
			Mercury	7471B		06/26/25	06/26/25	
			Metals ICP-Group1	6010D		06/26/25	06/27/25	
Q2425-02	COMP-2	SOIL			06/24/25			06/25/25
			Mercury	7471B		06/26/25	06/26/25	
			Metals ICP-Group1	6010D		06/26/25	06/27/25	
Q2425-03	COMP-3	SOIL			06/24/25			06/25/25
			Mercury	7471B		06/26/25	06/26/25	
			Metals ICP-Group1	6010D		06/26/25	06/27/25	

Hit Summary Sheet SW-846

SDG No.: Q2425
Client: Kleinfelder

Order ID: Q2425
Project ID: AS Jenks School

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : COMP-1								
Q2425-01	COMP-1	SOIL	Aluminum	7700		0.99	5.88	mg/Kg
Q2425-01	COMP-1	SOIL	Antimony	1.37	J	0.26	2.94	mg/Kg
Q2425-01	COMP-1	SOIL	Arsenic	6.92		0.22	1.18	mg/Kg
Q2425-01	COMP-1	SOIL	Barium	76.0		0.86	5.88	mg/Kg
Q2425-01	COMP-1	SOIL	Beryllium	0.75		0.029	0.35	mg/Kg
Q2425-01	COMP-1	SOIL	Chromium	12.1		0.055	0.59	mg/Kg
Q2425-01	COMP-1	SOIL	Cobalt	8.27		0.12	1.77	mg/Kg
Q2425-01	COMP-1	SOIL	Copper	71.1		0.26	1.18	mg/Kg
Q2425-01	COMP-1	SOIL	Iron	14900		4.70	5.88	mg/Kg
Q2425-01	COMP-1	SOIL	Lead	109		0.15	0.71	mg/Kg
Q2425-01	COMP-1	SOIL	Manganese	342		0.17	1.18	mg/Kg
Q2425-01	COMP-1	SOIL	Mercury	2.07	D	0.087	0.16	mg/Kg
Q2425-01	COMP-1	SOIL	Nickel	11.4		0.15	2.35	mg/Kg
Q2425-01	COMP-1	SOIL	Potassium	611		32.6	118	mg/Kg
Q2425-01	COMP-1	SOIL	Selenium	4.44		0.31	1.18	mg/Kg
Q2425-01	COMP-1	SOIL	Silver	0.45	J	0.14	0.59	mg/Kg
Q2425-01	COMP-1	SOIL	Sodium	55.9	J	20.9	118	mg/Kg
Q2425-01	COMP-1	SOIL	Vanadium	19.9		0.29	2.35	mg/Kg
Q2425-01	COMP-1	SOIL	Zinc	68.6		0.13	2.35	mg/Kg
Client ID : COMP-2								
Q2425-02	COMP-2	SOIL	Aluminum	5300		0.95	5.63	mg/Kg
Q2425-02	COMP-2	SOIL	Antimony	0.72	J	0.25	2.81	mg/Kg
Q2425-02	COMP-2	SOIL	Arsenic	6.56		0.21	1.13	mg/Kg
Q2425-02	COMP-2	SOIL	Barium	84.9		0.82	5.63	mg/Kg
Q2425-02	COMP-2	SOIL	Beryllium	0.70		0.028	0.34	mg/Kg
Q2425-02	COMP-2	SOIL	Cadmium	0.047	J	0.027	0.34	mg/Kg
Q2425-02	COMP-2	SOIL	Chromium	8.04		0.053	0.56	mg/Kg
Q2425-02	COMP-2	SOIL	Cobalt	7.63		0.11	1.69	mg/Kg
Q2425-02	COMP-2	SOIL	Copper	15.1		0.25	1.13	mg/Kg
Q2425-02	COMP-2	SOIL	Iron	8290		4.49	5.63	mg/Kg
Q2425-02	COMP-2	SOIL	Lead	115		0.15	0.68	mg/Kg
Q2425-02	COMP-2	SOIL	Manganese	429		0.16	1.13	mg/Kg
Q2425-02	COMP-2	SOIL	Mercury	2.14	D	0.087	0.16	mg/Kg
Q2425-02	COMP-2	SOIL	Nickel	9.81		0.15	2.25	mg/Kg
Q2425-02	COMP-2	SOIL	Potassium	649		31.2	113	mg/Kg
Q2425-02	COMP-2	SOIL	Selenium	2.43		0.29	1.13	mg/Kg
Q2425-02	COMP-2	SOIL	Silver	0.40	J	0.14	0.56	mg/Kg

Hit Summary Sheet
SW-846

SDG No.: Q2425

Order ID: Q2425

Client: Kleinfelder

Project ID: AS Jenks School

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q2425-02	COMP-2	SOIL	Sodium	60.9	J	20.0	113	mg/Kg
Q2425-02	COMP-2	SOIL	Thallium	0.29	J	0.26	2.25	mg/Kg
Q2425-02	COMP-2	SOIL	Vanadium	11.3		0.28	2.25	mg/Kg
Q2425-02	COMP-2	SOIL	Zinc	34.3		0.12	2.25	mg/Kg
Client ID : COMP-3								
Q2425-03	COMP-3	SOIL	Aluminum	11400		0.86	5.09	mg/Kg
Q2425-03	COMP-3	SOIL	Antimony	1.43	J	0.22	2.54	mg/Kg
Q2425-03	COMP-3	SOIL	Arsenic	7.98		0.19	1.02	mg/Kg
Q2425-03	COMP-3	SOIL	Barium	46.5		0.74	5.09	mg/Kg
Q2425-03	COMP-3	SOIL	Beryllium	0.86		0.025	0.31	mg/Kg
Q2425-03	COMP-3	SOIL	Chromium	17.8		0.048	0.51	mg/Kg
Q2425-03	COMP-3	SOIL	Cobalt	7.69		0.10	1.53	mg/Kg
Q2425-03	COMP-3	SOIL	Copper	7.82		0.22	1.02	mg/Kg
Q2425-03	COMP-3	SOIL	Iron	19100		4.06	5.09	mg/Kg
Q2425-03	COMP-3	SOIL	Lead	15.8		0.13	0.61	mg/Kg
Q2425-03	COMP-3	SOIL	Manganese	180		0.14	1.02	mg/Kg
Q2425-03	COMP-3	SOIL	Mercury	0.10		0.0090	0.015	mg/Kg
Q2425-03	COMP-3	SOIL	Nickel	12.2		0.13	2.03	mg/Kg
Q2425-03	COMP-3	SOIL	Potassium	760		28.2	102	mg/Kg
Q2425-03	COMP-3	SOIL	Selenium	5.97		0.27	1.02	mg/Kg
Q2425-03	COMP-3	SOIL	Silver	0.18	J	0.12	0.51	mg/Kg
Q2425-03	COMP-3	SOIL	Sodium	132		18.1	102	mg/Kg
Q2425-03	COMP-3	SOIL	Vanadium	29.2		0.25	2.03	mg/Kg
Q2425-03	COMP-3	SOIL	Zinc	27.8		0.11	2.03	mg/Kg



SAMPLE DATA

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-1	SDG No.:	Q2425
Lab Sample ID:	Q2425-01	Matrix:	SOIL
Level (low/med):	low	% Solid:	76.9

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units(Dry Weight)	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	7700		1	0.99	5.88	mg/Kg	06/26/25 10:20	06/27/25 18:19	6010D	SW3050
7440-36-0	Antimony	1.37	JN	1	0.26	2.94	mg/Kg	06/26/25 10:20	06/27/25 18:19	6010D	SW3050
7440-38-2	Arsenic	6.92		1	0.22	1.18	mg/Kg	06/26/25 10:20	06/27/25 18:19	6010D	SW3050
7440-39-3	Barium	76.0		1	0.86	5.88	mg/Kg	06/26/25 10:20	06/27/25 18:19	6010D	SW3050
7440-41-7	Beryllium	0.75		1	0.029	0.35	mg/Kg	06/26/25 10:20	06/27/25 18:19	6010D	SW3050
7440-43-9	Cadmium	0.028	U	1	0.028	0.35	mg/Kg	06/26/25 10:20	06/27/25 18:19	6010D	SW3050
7440-47-3	Chromium	12.1		1	0.055	0.59	mg/Kg	06/26/25 10:20	06/27/25 18:19	6010D	SW3050
7440-48-4	Cobalt	8.27		1	0.12	1.77	mg/Kg	06/26/25 10:20	06/27/25 18:19	6010D	SW3050
7440-50-8	Copper	71.1		1	0.26	1.18	mg/Kg	06/26/25 10:20	06/27/25 18:19	6010D	SW3050
7439-89-6	Iron	14900		1	4.70	5.88	mg/Kg	06/26/25 10:20	06/27/25 18:19	6010D	SW3050
7439-92-1	Lead	109		1	0.15	0.71	mg/Kg	06/26/25 10:20	06/27/25 18:19	6010D	SW3050
7439-96-5	Manganese	342		1	0.17	1.18	mg/Kg	06/26/25 10:20	06/27/25 18:19	6010D	SW3050
7439-97-6	Mercury	2.07	D	10	0.087	0.16	mg/Kg	06/26/25 12:15	06/26/25 16:06	7471B	
7440-02-0	Nickel	11.4		1	0.15	2.35	mg/Kg	06/26/25 10:20	06/27/25 18:19	6010D	SW3050
7440-09-7	Potassium	611	N	1	32.6	118	mg/Kg	06/26/25 10:20	06/27/25 18:19	6010D	SW3050
7782-49-2	Selenium	4.44		1	0.31	1.18	mg/Kg	06/26/25 10:20	06/27/25 18:19	6010D	SW3050
7440-22-4	Silver	0.45	J	1	0.14	0.59	mg/Kg	06/26/25 10:20	06/27/25 18:19	6010D	SW3050
7440-23-5	Sodium	55.9	J	1	20.9	118	mg/Kg	06/26/25 10:20	06/27/25 18:19	6010D	SW3050
7440-28-0	Thallium	0.27	U	1	0.27	2.35	mg/Kg	06/26/25 10:20	06/27/25 18:19	6010D	SW3050
7440-62-2	Vanadium	19.9	N	1	0.29	2.35	mg/Kg	06/26/25 10:20	06/27/25 18:19	6010D	SW3050
7440-66-6	Zinc	68.6	N	1	0.13	2.35	mg/Kg	06/26/25 10:20	06/27/25 18:19	6010D	SW3050

Color Before:	Brown	Clarity Before:	Texture:	Medium
Color After:	Yellow	Clarity After:	Artifacts:	
Comments:	Metals Group1			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-2	SDG No.:	Q2425
Lab Sample ID:	Q2425-02	Matrix:	SOIL
Level (low/med):	low	% Solid:	79.7

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units(Dry Weight)	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	5300		1	0.95	5.63	mg/Kg	06/26/25 10:20	06/27/25 18:24	6010D	SW3050
7440-36-0	Antimony	0.72	JN	1	0.25	2.81	mg/Kg	06/26/25 10:20	06/27/25 18:24	6010D	SW3050
7440-38-2	Arsenic	6.56		1	0.21	1.13	mg/Kg	06/26/25 10:20	06/27/25 18:24	6010D	SW3050
7440-39-3	Barium	84.9		1	0.82	5.63	mg/Kg	06/26/25 10:20	06/27/25 18:24	6010D	SW3050
7440-41-7	Beryllium	0.70		1	0.028	0.34	mg/Kg	06/26/25 10:20	06/27/25 18:24	6010D	SW3050
7440-43-9	Cadmium	0.047	J	1	0.027	0.34	mg/Kg	06/26/25 10:20	06/27/25 18:24	6010D	SW3050
7440-47-3	Chromium	8.04		1	0.053	0.56	mg/Kg	06/26/25 10:20	06/27/25 18:24	6010D	SW3050
7440-48-4	Cobalt	7.63		1	0.11	1.69	mg/Kg	06/26/25 10:20	06/27/25 18:24	6010D	SW3050
7440-50-8	Copper	15.1		1	0.25	1.13	mg/Kg	06/26/25 10:20	06/27/25 18:24	6010D	SW3050
7439-89-6	Iron	8290		1	4.49	5.63	mg/Kg	06/26/25 10:20	06/27/25 18:24	6010D	SW3050
7439-92-1	Lead	115		1	0.15	0.68	mg/Kg	06/26/25 10:20	06/27/25 18:24	6010D	SW3050
7439-96-5	Manganese	429		1	0.16	1.13	mg/Kg	06/26/25 10:20	06/27/25 18:24	6010D	SW3050
7439-97-6	Mercury	2.14	D	10	0.087	0.16	mg/Kg	06/26/25 12:15	06/26/25 16:11	7471B	
7440-02-0	Nickel	9.81		1	0.15	2.25	mg/Kg	06/26/25 10:20	06/27/25 18:24	6010D	SW3050
7440-09-7	Potassium	649	N	1	31.2	113	mg/Kg	06/26/25 10:20	06/27/25 18:24	6010D	SW3050
7782-49-2	Selenium	2.43		1	0.29	1.13	mg/Kg	06/26/25 10:20	06/27/25 18:24	6010D	SW3050
7440-22-4	Silver	0.40	J	1	0.14	0.56	mg/Kg	06/26/25 10:20	06/27/25 18:24	6010D	SW3050
7440-23-5	Sodium	60.9	J	1	20.0	113	mg/Kg	06/26/25 10:20	06/27/25 18:24	6010D	SW3050
7440-28-0	Thallium	0.29	J	1	0.26	2.25	mg/Kg	06/26/25 10:20	06/27/25 18:24	6010D	SW3050
7440-62-2	Vanadium	11.3	N	1	0.28	2.25	mg/Kg	06/26/25 10:20	06/27/25 18:24	6010D	SW3050
7440-66-6	Zinc	34.3	N	1	0.12	2.25	mg/Kg	06/26/25 10:20	06/27/25 18:24	6010D	SW3050

Color Before:	Brown	Clarity Before:	Texture:	Medium
Color After:	Yellow	Clarity After:	Artifacts:	
Comments:	Metals Group1			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-3	SDG No.:	Q2425
Lab Sample ID:	Q2425-03	Matrix:	SOIL
Level (low/med):	low	% Solid:	82.6

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units(Dry Weight)	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	11400		1	0.86	5.09	mg/Kg	06/26/25 10:20	06/27/25 19:03	6010D	SW3050
7440-36-0	Antimony	1.43	JN	1	0.22	2.54	mg/Kg	06/26/25 10:20	06/27/25 19:03	6010D	SW3050
7440-38-2	Arsenic	7.98		1	0.19	1.02	mg/Kg	06/26/25 10:20	06/27/25 19:03	6010D	SW3050
7440-39-3	Barium	46.5		1	0.74	5.09	mg/Kg	06/26/25 10:20	06/27/25 19:03	6010D	SW3050
7440-41-7	Beryllium	0.86		1	0.025	0.31	mg/Kg	06/26/25 10:20	06/27/25 19:03	6010D	SW3050
7440-43-9	Cadmium	0.024	U	1	0.024	0.31	mg/Kg	06/26/25 10:20	06/27/25 19:03	6010D	SW3050
7440-47-3	Chromium	17.8		1	0.048	0.51	mg/Kg	06/26/25 10:20	06/27/25 19:03	6010D	SW3050
7440-48-4	Cobalt	7.69		1	0.10	1.53	mg/Kg	06/26/25 10:20	06/27/25 19:03	6010D	SW3050
7440-50-8	Copper	7.82		1	0.22	1.02	mg/Kg	06/26/25 10:20	06/27/25 19:03	6010D	SW3050
7439-89-6	Iron	19100		1	4.06	5.09	mg/Kg	06/26/25 10:20	06/27/25 19:03	6010D	SW3050
7439-92-1	Lead	15.8		1	0.13	0.61	mg/Kg	06/26/25 10:20	06/27/25 19:03	6010D	SW3050
7439-96-5	Manganese	180		1	0.14	1.02	mg/Kg	06/26/25 10:20	06/27/25 19:03	6010D	SW3050
7439-97-6	Mercury	0.10		1	0.0090	0.015	mg/Kg	06/26/25 12:15	06/26/25 15:45	7471B	
7440-02-0	Nickel	12.2		1	0.13	2.03	mg/Kg	06/26/25 10:20	06/27/25 19:03	6010D	SW3050
7440-09-7	Potassium	760	N	1	28.2	102	mg/Kg	06/26/25 10:20	06/27/25 19:03	6010D	SW3050
7782-49-2	Selenium	5.97		1	0.27	1.02	mg/Kg	06/26/25 10:20	06/27/25 19:03	6010D	SW3050
7440-22-4	Silver	0.18	J	1	0.12	0.51	mg/Kg	06/26/25 10:20	06/27/25 19:03	6010D	SW3050
7440-23-5	Sodium	132		1	18.1	102	mg/Kg	06/26/25 10:20	06/27/25 19:03	6010D	SW3050
7440-28-0	Thallium	0.23	U	1	0.23	2.03	mg/Kg	06/26/25 10:20	06/27/25 19:03	6010D	SW3050
7440-62-2	Vanadium	29.2	N	1	0.25	2.03	mg/Kg	06/26/25 10:20	06/27/25 19:03	6010D	SW3050
7440-66-6	Zinc	27.8	N	1	0.11	2.03	mg/Kg	06/26/25 10:20	06/27/25 19:03	6010D	SW3050

Color Before:	Brown	Clarity Before:	Texture:	Medium
Color After:	Yellow	Clarity After:	Artifacts:	
Comments:	Metals Group1			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder **SDG No.:** Q2425
Contract: POWE02 **Lab Code:** CHEM **Case No.:** Q2425 **SAS No.:** Q2425
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV08	Mercury	3.92	4.0	98	90 - 110	CV	06/26/2025	14:34	1b136298

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder **SDG No.:** Q2425
Contract: POWE02 **Lab Code:** CHEM **Case No.:** Q2425 **SAS No.:** Q2425
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV23	Mercury	4.94	5.0	99	90 - 110	CV	06/26/2025	14:39	1b136298
CCV24	Mercury	4.79	5.0	96	90 - 110	CV	06/26/2025	15:12	1b136298
CCV25	Mercury	4.69	5.0	94	90 - 110	CV	06/26/2025	15:40	1b136298
CCV26	Mercury	4.76	5.0	95	90 - 110	CV	06/26/2025	16:20	1b136298

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder SDG No.: Q2425
Contract: POWE02 Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7840	8000	98	90 - 110	P	06/27/2025	11:45	LB136319
	Antimony	4280	4000	107	90 - 110	P	06/27/2025	11:45	LB136319
	Arsenic	3910	4000	98	90 - 110	P	06/27/2025	11:45	LB136319
	Barium	8020	8000	100	90 - 110	P	06/27/2025	11:45	LB136319
	Beryllium	204	200	102	90 - 110	P	06/27/2025	11:45	LB136319
	Cadmium	1990	2000	100	90 - 110	P	06/27/2025	11:45	LB136319
	Chromium	757	800	95	90 - 110	P	06/27/2025	11:45	LB136319
	Cobalt	1970	2000	99	90 - 110	P	06/27/2025	11:45	LB136319
	Copper	1040	1000	104	90 - 110	P	06/27/2025	11:45	LB136319
	Iron	3850	4000	96	90 - 110	P	06/27/2025	11:45	LB136319
	Lead	3950	4000	99	90 - 110	P	06/27/2025	11:45	LB136319
	Manganese	1990	2000	100	90 - 110	P	06/27/2025	11:45	LB136319
	Nickel	1950	2000	98	90 - 110	P	06/27/2025	11:45	LB136319
	Potassium	19500	20000	97	90 - 110	P	06/27/2025	11:45	LB136319
	Selenium	4050	4000	101	90 - 110	P	06/27/2025	11:45	LB136319
	Silver	941	1000	94	90 - 110	P	06/27/2025	11:45	LB136319
	Sodium	20500	20000	102	90 - 110	P	06/27/2025	11:45	LB136319
	Thallium	3950	4000	99	90 - 110	P	06/27/2025	11:45	LB136319
	Vanadium	2020	2000	101	90 - 110	P	06/27/2025	11:45	LB136319
	Zinc	2030	2000	101	90 - 110	P	06/27/2025	11:45	LB136319

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder SDG No.: Q2425
Contract: POWE02 Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	94.7	100	95	80 - 120	P	06/27/2025	12:03	LB136319
	Antimony	57.4	50.0	115	80 - 120	P	06/27/2025	12:03	LB136319
	Arsenic	16.7	20.0	84	80 - 120	P	06/27/2025	12:03	LB136319
	Barium	99.9	100	100	80 - 120	P	06/27/2025	12:03	LB136319
	Beryllium	6.43	6.0	107	80 - 120	P	06/27/2025	12:03	LB136319
	Cadmium	5.74	6.0	96	80 - 120	P	06/27/2025	12:03	LB136319
	Chromium	11.1	10.0	112	80 - 120	P	06/27/2025	12:03	LB136319
	Cobalt	30.3	30.0	101	80 - 120	P	06/27/2025	12:03	LB136319
	Copper	22.3	20.0	112	80 - 120	P	06/27/2025	12:03	LB136319
	Iron	111	100	111	80 - 120	P	06/27/2025	12:03	LB136319
	Lead	10.3	12.0	86	80 - 120	P	06/27/2025	12:03	LB136319
	Manganese	22.8	20.0	114	80 - 120	P	06/27/2025	12:03	LB136319
	Nickel	39.8	40.0	100	80 - 120	P	06/27/2025	12:03	LB136319
	Potassium	1800	2000	90	80 - 120	P	06/27/2025	12:03	LB136319
	Selenium	19.5	20.0	98	80 - 120	P	06/27/2025	12:03	LB136319
	Silver	10.7	10.0	107	80 - 120	P	06/27/2025	12:03	LB136319
	Sodium	1900	2000	95	80 - 120	P	06/27/2025	12:03	LB136319
	Thallium	40.7	40.0	102	80 - 120	P	06/27/2025	12:03	LB136319
	Vanadium	47.8	40.0	119	80 - 120	P	06/27/2025	12:03	LB136319
	Zinc	34.2	40.0	85	80 - 120	P	06/27/2025	12:03	LB136319

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder SDG No.: Q2425
 Contract: POWE02 Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425
 Initial Calibration Source: EPA
 Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9460	10000	95	90 - 110	P	06/27/2025	13:22	LB136319
	Antimony	5000	5000	100	90 - 110	P	06/27/2025	13:22	LB136319
	Arsenic	4830	5000	97	90 - 110	P	06/27/2025	13:22	LB136319
	Barium	10100	10000	101	90 - 110	P	06/27/2025	13:22	LB136319
	Beryllium	250	250	100	90 - 110	P	06/27/2025	13:22	LB136319
	Cadmium	2460	2500	99	90 - 110	P	06/27/2025	13:22	LB136319
	Chromium	1030	1000	103	90 - 110	P	06/27/2025	13:22	LB136319
	Cobalt	2470	2500	99	90 - 110	P	06/27/2025	13:22	LB136319
	Copper	1250	1250	100	90 - 110	P	06/27/2025	13:22	LB136319
	Iron	5370	5000	108	90 - 110	P	06/27/2025	13:22	LB136319
	Lead	4980	5000	100	90 - 110	P	06/27/2025	13:22	LB136319
	Manganese	2520	2500	101	90 - 110	P	06/27/2025	13:22	LB136319
	Nickel	2510	2500	100	90 - 110	P	06/27/2025	13:22	LB136319
	Potassium	24900	25000	99	90 - 110	P	06/27/2025	13:22	LB136319
	Selenium	5040	5000	101	90 - 110	P	06/27/2025	13:22	LB136319
	Silver	1230	1250	99	90 - 110	P	06/27/2025	13:22	LB136319
	Sodium	25000	25000	100	90 - 110	P	06/27/2025	13:22	LB136319
	Thallium	5000	5000	100	90 - 110	P	06/27/2025	13:22	LB136319
	Vanadium	2480	2500	99	90 - 110	P	06/27/2025	13:22	LB136319
	Zinc	2510	2500	101	90 - 110	P	06/27/2025	13:22	LB136319
CCV02	Aluminum	9650	10000	96	90 - 110	P	06/27/2025	14:19	LB136319
	Antimony	5050	5000	101	90 - 110	P	06/27/2025	14:19	LB136319
	Arsenic	4900	5000	98	90 - 110	P	06/27/2025	14:19	LB136319
	Barium	10200	10000	102	90 - 110	P	06/27/2025	14:19	LB136319
	Beryllium	256	250	102	90 - 110	P	06/27/2025	14:19	LB136319
	Cadmium	2500	2500	100	90 - 110	P	06/27/2025	14:19	LB136319
	Chromium	1010	1000	101	90 - 110	P	06/27/2025	14:19	LB136319
	Cobalt	2510	2500	100	90 - 110	P	06/27/2025	14:19	LB136319
	Copper	1290	1250	103	90 - 110	P	06/27/2025	14:19	LB136319
	Iron	5250	5000	105	90 - 110	P	06/27/2025	14:19	LB136319
	Lead	5070	5000	101	90 - 110	P	06/27/2025	14:19	LB136319
	Manganese	2540	2500	102	90 - 110	P	06/27/2025	14:19	LB136319
	Nickel	2530	2500	101	90 - 110	P	06/27/2025	14:19	LB136319
	Potassium	25600	25000	102	90 - 110	P	06/27/2025	14:19	LB136319

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder SDG No.: Q2425
 Contract: POWE02 Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425
 Initial Calibration Source: EPA
 Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Selenium	5100	5000	102	90 - 110	P	06/27/2025	14:19	LB136319
	Silver	1260	1250	101	90 - 110	P	06/27/2025	14:19	LB136319
	Sodium	25800	25000	103	90 - 110	P	06/27/2025	14:19	LB136319
	Thallium	5080	5000	102	90 - 110	P	06/27/2025	14:19	LB136319
	Vanadium	2500	2500	100	90 - 110	P	06/27/2025	14:19	LB136319
	Zinc	2550	2500	102	90 - 110	P	06/27/2025	14:19	LB136319
CCV03	Aluminum	9690	10000	97	90 - 110	P	06/27/2025	15:12	LB136319
	Antimony	5100	5000	102	90 - 110	P	06/27/2025	15:12	LB136319
	Arsenic	4930	5000	99	90 - 110	P	06/27/2025	15:12	LB136319
	Barium	10300	10000	103	90 - 110	P	06/27/2025	15:12	LB136319
	Beryllium	257	250	103	90 - 110	P	06/27/2025	15:12	LB136319
	Cadmium	2530	2500	101	90 - 110	P	06/27/2025	15:12	LB136319
	Chromium	1040	1000	104	90 - 110	P	06/27/2025	15:12	LB136319
	Cobalt	2540	2500	102	90 - 110	P	06/27/2025	15:12	LB136319
	Copper	1290	1250	103	90 - 110	P	06/27/2025	15:12	LB136319
	Iron	5340	5000	107	90 - 110	P	06/27/2025	15:12	LB136319
	Lead	5120	5000	102	90 - 110	P	06/27/2025	15:12	LB136319
	Manganese	2560	2500	103	90 - 110	P	06/27/2025	15:12	LB136319
	Nickel	2560	2500	103	90 - 110	P	06/27/2025	15:12	LB136319
	Potassium	25800	25000	103	90 - 110	P	06/27/2025	15:12	LB136319
	Selenium	5170	5000	103	90 - 110	P	06/27/2025	15:12	LB136319
	Silver	1270	1250	102	90 - 110	P	06/27/2025	15:12	LB136319
	Sodium	26000	25000	104	90 - 110	P	06/27/2025	15:12	LB136319
	Thallium	5140	5000	103	90 - 110	P	06/27/2025	15:12	LB136319
	Vanadium	2550	2500	102	90 - 110	P	06/27/2025	15:12	LB136319
	Zinc	2560	2500	102	90 - 110	P	06/27/2025	15:12	LB136319
CCV04	Aluminum	10300	10000	103	90 - 110	P	06/27/2025	16:04	LB136319
	Antimony	5110	5000	102	90 - 110	P	06/27/2025	16:04	LB136319
	Arsenic	4940	5000	99	90 - 110	P	06/27/2025	16:04	LB136319
	Barium	10300	10000	103	90 - 110	P	06/27/2025	16:04	LB136319
	Beryllium	256	250	102	90 - 110	P	06/27/2025	16:04	LB136319
	Cadmium	2530	2500	101	90 - 110	P	06/27/2025	16:04	LB136319
	Chromium	1010	1000	101	90 - 110	P	06/27/2025	16:04	LB136319
	Cobalt	2540	2500	102	90 - 110	P	06/27/2025	16:04	LB136319

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder SDG No.: Q2425
Contract: POWE02 Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Copper	1290	1250	103	90 - 110	P	06/27/2025	16:04	LB136319
	Iron	5250	5000	105	90 - 110	P	06/27/2025	16:04	LB136319
	Lead	5120	5000	102	90 - 110	P	06/27/2025	16:04	LB136319
	Manganese	2580	2500	103	90 - 110	P	06/27/2025	16:04	LB136319
	Nickel	2550	2500	102	90 - 110	P	06/27/2025	16:04	LB136319
	Potassium	25800	25000	103	90 - 110	P	06/27/2025	16:04	LB136319
	Selenium	5160	5000	103	90 - 110	P	06/27/2025	16:04	LB136319
	Silver	1260	1250	101	90 - 110	P	06/27/2025	16:04	LB136319
	Sodium	26000	25000	104	90 - 110	P	06/27/2025	16:04	LB136319
	Thallium	5130	5000	102	90 - 110	P	06/27/2025	16:04	LB136319
	Vanadium	2550	2500	102	90 - 110	P	06/27/2025	16:04	LB136319
	Zinc	2570	2500	103	90 - 110	P	06/27/2025	16:04	LB136319
CCV05	Aluminum	9590	10000	96	90 - 110	P	06/27/2025	16:56	LB136319
	Antimony	4950	5000	99	90 - 110	P	06/27/2025	16:56	LB136319
	Arsenic	4820	5000	96	90 - 110	P	06/27/2025	16:56	LB136319
	Barium	10200	10000	102	90 - 110	P	06/27/2025	16:56	LB136319
	Beryllium	255	250	102	90 - 110	P	06/27/2025	16:56	LB136319
	Cadmium	2460	2500	98	90 - 110	P	06/27/2025	16:56	LB136319
	Chromium	1040	1000	104	90 - 110	P	06/27/2025	16:56	LB136319
	Cobalt	2460	2500	98	90 - 110	P	06/27/2025	16:56	LB136319
	Copper	1280	1250	102	90 - 110	P	06/27/2025	16:56	LB136319
	Iron	5380	5000	108	90 - 110	P	06/27/2025	16:56	LB136319
	Lead	4960	5000	99	90 - 110	P	06/27/2025	16:56	LB136319
	Manganese	2490	2500	100	90 - 110	P	06/27/2025	16:56	LB136319
	Nickel	2500	2500	100	90 - 110	P	06/27/2025	16:56	LB136319
	Potassium	25700	25000	103	90 - 110	P	06/27/2025	16:56	LB136319
	Selenium	4980	5000	100	90 - 110	P	06/27/2025	16:56	LB136319
	Silver	1230	1250	98	90 - 110	P	06/27/2025	16:56	LB136319
	Sodium	25900	25000	104	90 - 110	P	06/27/2025	16:56	LB136319
	Thallium	4980	5000	100	90 - 110	P	06/27/2025	16:56	LB136319
	Vanadium	2470	2500	99	90 - 110	P	06/27/2025	16:56	LB136319
	Zinc	2480	2500	99	90 - 110	P	06/27/2025	16:56	LB136319
CCV06	Aluminum	9580	10000	96	90 - 110	P	06/27/2025	17:49	LB136319
	Antimony	5020	5000	100	90 - 110	P	06/27/2025	17:49	LB136319

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder SDG No.: Q2425
Contract: POWE02 Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV06	Arsenic	4880	5000	98	90 - 110	P	06/27/2025	17:49	LB136319
	Barium	10400	10000	104	90 - 110	P	06/27/2025	17:49	LB136319
	Beryllium	254	250	102	90 - 110	P	06/27/2025	17:49	LB136319
	Cadmium	2480	2500	99	90 - 110	P	06/27/2025	17:49	LB136319
	Chromium	1030	1000	103	90 - 110	P	06/27/2025	17:49	LB136319
	Cobalt	2480	2500	99	90 - 110	P	06/27/2025	17:49	LB136319
	Copper	1280	1250	102	90 - 110	P	06/27/2025	17:49	LB136319
	Iron	5430	5000	109	90 - 110	P	06/27/2025	17:49	LB136319
	Lead	5000	5000	100	90 - 110	P	06/27/2025	17:49	LB136319
	Manganese	2540	2500	102	90 - 110	P	06/27/2025	17:49	LB136319
	Nickel	2530	2500	101	90 - 110	P	06/27/2025	17:49	LB136319
	Potassium	25700	25000	103	90 - 110	P	06/27/2025	17:49	LB136319
	Selenium	5030	5000	101	90 - 110	P	06/27/2025	17:49	LB136319
	Silver	1220	1250	97	90 - 110	P	06/27/2025	17:49	LB136319
	Sodium	26100	25000	105	90 - 110	P	06/27/2025	17:49	LB136319
	Thallium	5060	5000	101	90 - 110	P	06/27/2025	17:49	LB136319
	Vanadium	2520	2500	101	90 - 110	P	06/27/2025	17:49	LB136319
	Zinc	2560	2500	102	90 - 110	P	06/27/2025	17:49	LB136319
CCV07	Aluminum	9940	10000	99	90 - 110	P	06/27/2025	18:41	LB136319
	Antimony	5140	5000	103	90 - 110	P	06/27/2025	18:41	LB136319
	Arsenic	5050	5000	101	90 - 110	P	06/27/2025	18:41	LB136319
	Barium	10600	10000	106	90 - 110	P	06/27/2025	18:41	LB136319
	Beryllium	260	250	104	90 - 110	P	06/27/2025	18:41	LB136319
	Cadmium	2550	2500	102	90 - 110	P	06/27/2025	18:41	LB136319
	Chromium	1030	1000	103	90 - 110	P	06/27/2025	18:41	LB136319
	Cobalt	2540	2500	102	90 - 110	P	06/27/2025	18:41	LB136319
	Copper	1310	1250	105	90 - 110	P	06/27/2025	18:41	LB136319
	Iron	5380	5000	108	90 - 110	P	06/27/2025	18:41	LB136319
	Lead	5120	5000	102	90 - 110	P	06/27/2025	18:41	LB136319
	Manganese	2590	2500	103	90 - 110	P	06/27/2025	18:41	LB136319
	Nickel	2570	2500	103	90 - 110	P	06/27/2025	18:41	LB136319
	Potassium	26400	25000	106	90 - 110	P	06/27/2025	18:41	LB136319
	Selenium	5140	5000	103	90 - 110	P	06/27/2025	18:41	LB136319
	Silver	1240	1250	99	90 - 110	P	06/27/2025	18:41	LB136319

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder **SDG No.:** Q2425
Contract: POWE02 **Lab Code:** CHEM **Case No.:** Q2425 **SAS No.:** Q2425
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Sodium	26800	25000	107	90 - 110	P	06/27/2025	18:41	LB136319
	Thallium	5190	5000	104	90 - 110	P	06/27/2025	18:41	LB136319
	Vanadium	2610	2500	104	90 - 110	P	06/27/2025	18:41	LB136319
	Zinc	2630	2500	105	90 - 110	P	06/27/2025	18:41	LB136319
CCV08	Aluminum	9930	10000	99	90 - 110	P	06/27/2025	19:29	LB136319
	Antimony	5100	5000	102	90 - 110	P	06/27/2025	19:29	LB136319
	Arsenic	5000	5000	100	90 - 110	P	06/27/2025	19:29	LB136319
	Barium	10500	10000	105	90 - 110	P	06/27/2025	19:29	LB136319
	Beryllium	256	250	102	90 - 110	P	06/27/2025	19:29	LB136319
	Cadmium	2520	2500	101	90 - 110	P	06/27/2025	19:29	LB136319
	Chromium	1010	1000	100	90 - 110	P	06/27/2025	19:29	LB136319
	Cobalt	2520	2500	101	90 - 110	P	06/27/2025	19:29	LB136319
	Copper	1290	1250	104	90 - 110	P	06/27/2025	19:29	LB136319
	Iron	5200	5000	104	90 - 110	P	06/27/2025	19:29	LB136319
	Lead	5040	5000	101	90 - 110	P	06/27/2025	19:29	LB136319
	Manganese	2530	2500	101	90 - 110	P	06/27/2025	19:29	LB136319
	Nickel	2530	2500	101	90 - 110	P	06/27/2025	19:29	LB136319
	Potassium	26300	25000	105	90 - 110	P	06/27/2025	19:29	LB136319
	Selenium	5060	5000	101	90 - 110	P	06/27/2025	19:29	LB136319
	Silver	1220	1250	98	90 - 110	P	06/27/2025	19:29	LB136319
	Sodium	26500	25000	106	90 - 110	P	06/27/2025	19:29	LB136319
	Thallium	5130	5000	103	90 - 110	P	06/27/2025	19:29	LB136319
	Vanadium	2550	2500	102	90 - 110	P	06/27/2025	19:29	LB136319
	Zinc	2610	2500	104	90 - 110	P	06/27/2025	19:29	LB136319

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Kleinfelder **SDG No.:** Q2425
Contract: POWE02 **Lab Code:** CHEM **Case No.:** Q2425 **SAS No.:** Q2425
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	105	70 - 130	CV	06/26/2025	14:44	lb136298
CRI01	Aluminum	92.6	100	93	65 - 135	P	06/27/2025	12:58	LB136319
	Antimony	56.3	50.0	113	65 - 135	P	06/27/2025	12:58	LB136319
	Arsenic	16.9	20.0	84	65 - 135	P	06/27/2025	12:58	LB136319
	Barium	106	100	106	65 - 135	P	06/27/2025	12:58	LB136319
	Beryllium	6.39	6.0	106	65 - 135	P	06/27/2025	12:58	LB136319
	Cadmium	5.61	6.0	94	65 - 135	P	06/27/2025	12:58	LB136319
	Chromium	12.6	10.0	126	65 - 135	P	06/27/2025	12:58	LB136319
	Cobalt	30.4	30.0	101	65 - 135	P	06/27/2025	12:58	LB136319
	Copper	23.3	20.0	116	65 - 135	P	06/27/2025	12:58	LB136319
	Iron	109	100	108	65 - 135	P	06/27/2025	12:58	LB136319
	Lead	12.2	12.0	101	65 - 135	P	06/27/2025	12:58	LB136319
	Manganese	22.7	20.0	114	65 - 135	P	06/27/2025	12:58	LB136319
	Nickel	40.0	40.0	100	65 - 135	P	06/27/2025	12:58	LB136319
	Potassium	1640	2000	82	65 - 135	P	06/27/2025	12:58	LB136319
	Selenium	20.1	20.0	100	65 - 135	P	06/27/2025	12:58	LB136319
	Silver	11.2	10.0	112	65 - 135	P	06/27/2025	12:58	LB136319
	Sodium	1900	2000	95	65 - 135	P	06/27/2025	12:58	LB136319
	Thallium	40.8	40.0	102	65 - 135	P	06/27/2025	12:58	LB136319
	Vanadium	39.3	40.0	98	65 - 135	P	06/27/2025	12:58	LB136319
	Zinc	34.2	40.0	86	65 - 135	P	06/27/2025	12:58	LB136319



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Kleinfelder</u>				SDG No.:	<u>Q2425</u>			
Contract:	<u>POWE02</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2425</u>	SAS No.:	<u>Q2425</u>		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB08	Mercury	0.076	+/-0.2	U	0.20	CV	06/26/2025	14:37	lb136298

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder SDG No.: Q2425
Contract: POWE02 Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB23	Mercury	0.076	+/-0.2	U	0.20	CV	06/26/2025	14:41	lb136298
CCB24	Mercury	0.076	+/-0.2	U	0.20	CV	06/26/2025	15:14	lb136298
CCB25	Mercury	0.076	+/-0.2	U	0.20	CV	06/26/2025	15:43	lb136298
CCB26	Mercury	0.076	+/-0.2	U	0.20	CV	06/26/2025	16:22	lb136298

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder **SDG No.:** Q2425
Contract: POWE02 **Lab Code:** CHEM **Case No.:** Q2425 **SAS No.:** Q2425

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	06/27/2025	12:54	LB136319
	Antimony	6.76	+/-25	U	50.0	P	06/27/2025	12:54	LB136319
	Arsenic	5.12	+/-10	U	20.0	P	06/27/2025	12:54	LB136319
	Barium	14.6	+/-50	U	100	P	06/27/2025	12:54	LB136319
	Beryllium	0.56	+/-3	U	6.00	P	06/27/2025	12:54	LB136319
	Cadmium	0.50	+/-3	U	6.00	P	06/27/2025	12:54	LB136319
	Chromium	2.12	+/-5	U	10.0	P	06/27/2025	12:54	LB136319
	Cobalt	2.26	+/-15	U	30.0	P	06/27/2025	12:54	LB136319
	Copper	4.60	+/-10	U	20.0	P	06/27/2025	12:54	LB136319
	Iron	23.4	+/-50	U	100	P	06/27/2025	12:54	LB136319
	Lead	2.30	+/-6	U	12.0	P	06/27/2025	12:54	LB136319
	Manganese	5.94	+/-10	U	20.0	P	06/27/2025	12:54	LB136319
	Nickel	3.06	+/-20	U	40.0	P	06/27/2025	12:54	LB136319
	Potassium	918	+/-1000	U	2000	P	06/27/2025	12:54	LB136319
	Selenium	9.64	+/-10	U	20.0	P	06/27/2025	12:54	LB136319
	Silver	1.62	+/-5	U	10.0	P	06/27/2025	12:54	LB136319
	Sodium	868	+/-1000	U	2000	P	06/27/2025	12:54	LB136319
	Thallium	4.38	+/-20	U	40.0	P	06/27/2025	12:54	LB136319
	Vanadium	6.26	+/-20	U	40.0	P	06/27/2025	12:54	LB136319
	Zinc	3.50	+/-20	U	40.0	P	06/27/2025	12:54	LB136319

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Kleinfelder</u>		SDG No.: <u>Q2425</u>							
Contract: <u>POWE02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2425</u>	SAS No.: <u>Q2425</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	06/27/2025	13:31	LB136319
	Antimony	6.76	+/-25	U	50.0	P	06/27/2025	13:31	LB136319
	Arsenic	5.12	+/-10	U	20.0	P	06/27/2025	13:31	LB136319
	Barium	14.6	+/-50	U	100	P	06/27/2025	13:31	LB136319
	Beryllium	0.56	+/-3	U	6.00	P	06/27/2025	13:31	LB136319
	Cadmium	0.50	+/-3	U	6.00	P	06/27/2025	13:31	LB136319
	Chromium	2.12	+/-5	U	10.0	P	06/27/2025	13:31	LB136319
	Cobalt	2.26	+/-15	U	30.0	P	06/27/2025	13:31	LB136319
	Copper	4.60	+/-10	U	20.0	P	06/27/2025	13:31	LB136319
	Iron	23.4	+/-50	U	100	P	06/27/2025	13:31	LB136319
	Lead	2.30	+/-6	U	12.0	P	06/27/2025	13:31	LB136319
	Manganese	5.94	+/-10	U	20.0	P	06/27/2025	13:31	LB136319
	Nickel	3.06	+/-20	U	40.0	P	06/27/2025	13:31	LB136319
	Potassium	918	+/-1000	U	2000	P	06/27/2025	13:31	LB136319
	Selenium	9.64	+/-10	U	20.0	P	06/27/2025	13:31	LB136319
	Silver	1.62	+/-5	U	10.0	P	06/27/2025	13:31	LB136319
	Sodium	868	+/-1000	U	2000	P	06/27/2025	13:31	LB136319
	Thallium	4.38	+/-20	U	40.0	P	06/27/2025	13:31	LB136319
	Vanadium	6.26	+/-20	U	40.0	P	06/27/2025	13:31	LB136319
	Zinc	3.50	+/-20	U	40.0	P	06/27/2025	13:31	LB136319
CCB02	Aluminum	11.3	+/-50	U	100	P	06/27/2025	14:24	LB136319
	Antimony	6.76	+/-25	U	50.0	P	06/27/2025	14:24	LB136319
	Arsenic	5.12	+/-10	U	20.0	P	06/27/2025	14:24	LB136319
	Barium	14.6	+/-50	U	100	P	06/27/2025	14:24	LB136319
	Beryllium	0.56	+/-3	U	6.00	P	06/27/2025	14:24	LB136319
	Cadmium	0.50	+/-3	U	6.00	P	06/27/2025	14:24	LB136319
	Chromium	2.12	+/-5	U	10.0	P	06/27/2025	14:24	LB136319
	Cobalt	2.26	+/-15	U	30.0	P	06/27/2025	14:24	LB136319
	Copper	4.60	+/-10	U	20.0	P	06/27/2025	14:24	LB136319
	Iron	23.4	+/-50	U	100	P	06/27/2025	14:24	LB136319
	Lead	2.30	+/-6	U	12.0	P	06/27/2025	14:24	LB136319
	Manganese	5.94	+/-10	U	20.0	P	06/27/2025	14:24	LB136319
	Nickel	3.06	+/-20	U	40.0	P	06/27/2025	14:24	LB136319
	Potassium	918	+/-1000	U	2000	P	06/27/2025	14:24	LB136319
	Selenium	9.64	+/-10	U	20.0	P	06/27/2025	14:24	LB136319
	Silver	1.62	+/-5	U	10.0	P	06/27/2025	14:24	LB136319
	Sodium	868	+/-1000	U	2000	P	06/27/2025	14:24	LB136319
	Thallium	4.38	+/-20	U	40.0	P	06/27/2025	14:24	LB136319
	Vanadium	6.26	+/-20	U	40.0	P	06/27/2025	14:24	LB136319

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Kleinfelder</u>		SDG No.: <u>Q2425</u>							
Contract: <u>POWE02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2425</u>	SAS No.: <u>Q2425</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Zinc	3.50	+/-20	U	40.0	P	06/27/2025	14:24	LB136319
CCB03	Aluminum	11.3	+/-50	U	100	P	06/27/2025	15:16	LB136319
	Antimony	6.76	+/-25	U	50.0	P	06/27/2025	15:16	LB136319
	Arsenic	5.12	+/-10	U	20.0	P	06/27/2025	15:16	LB136319
	Barium	14.6	+/-50	U	100	P	06/27/2025	15:16	LB136319
	Beryllium	0.56	+/-3	U	6.00	P	06/27/2025	15:16	LB136319
	Cadmium	0.50	+/-3	U	6.00	P	06/27/2025	15:16	LB136319
	Chromium	2.12	+/-5	U	10.0	P	06/27/2025	15:16	LB136319
	Cobalt	2.26	+/-15	U	30.0	P	06/27/2025	15:16	LB136319
	Copper	4.60	+/-10	U	20.0	P	06/27/2025	15:16	LB136319
	Iron	23.4	+/-50	U	100	P	06/27/2025	15:16	LB136319
	Lead	2.30	+/-6	U	12.0	P	06/27/2025	15:16	LB136319
	Manganese	5.94	+/-10	U	20.0	P	06/27/2025	15:16	LB136319
	Nickel	3.06	+/-20	U	40.0	P	06/27/2025	15:16	LB136319
	Potassium	918	+/-1000	U	2000	P	06/27/2025	15:16	LB136319
	Selenium	9.64	+/-10	U	20.0	P	06/27/2025	15:16	LB136319
	Silver	1.62	+/-5	U	10.0	P	06/27/2025	15:16	LB136319
	Sodium	868	+/-1000	U	2000	P	06/27/2025	15:16	LB136319
	Thallium	4.38	+/-20	U	40.0	P	06/27/2025	15:16	LB136319
	Vanadium	6.26	+/-20	U	40.0	P	06/27/2025	15:16	LB136319
	Zinc	3.50	+/-20	U	40.0	P	06/27/2025	15:16	LB136319
CCB04	Aluminum	11.3	+/-50	U	100	P	06/27/2025	16:09	LB136319
	Antimony	6.76	+/-25	U	50.0	P	06/27/2025	16:09	LB136319
	Arsenic	5.12	+/-10	U	20.0	P	06/27/2025	16:09	LB136319
	Barium	14.6	+/-50	U	100	P	06/27/2025	16:09	LB136319
	Beryllium	0.56	+/-3	U	6.00	P	06/27/2025	16:09	LB136319
	Cadmium	0.50	+/-3	U	6.00	P	06/27/2025	16:09	LB136319
	Chromium	2.12	+/-5	U	10.0	P	06/27/2025	16:09	LB136319
	Cobalt	2.26	+/-15	U	30.0	P	06/27/2025	16:09	LB136319
	Copper	4.60	+/-10	U	20.0	P	06/27/2025	16:09	LB136319
	Iron	23.4	+/-50	U	100	P	06/27/2025	16:09	LB136319
	Lead	2.30	+/-6	U	12.0	P	06/27/2025	16:09	LB136319
	Manganese	5.94	+/-10	U	20.0	P	06/27/2025	16:09	LB136319
	Nickel	3.06	+/-20	U	40.0	P	06/27/2025	16:09	LB136319
	Potassium	918	+/-1000	U	2000	P	06/27/2025	16:09	LB136319
	Selenium	9.64	+/-10	U	20.0	P	06/27/2025	16:09	LB136319
	Silver	1.62	+/-5	U	10.0	P	06/27/2025	16:09	LB136319
	Sodium	868	+/-1000	U	2000	P	06/27/2025	16:09	LB136319
	Thallium	4.38	+/-20	U	40.0	P	06/27/2025	16:09	LB136319

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Kleinfelder</u>		SDG No.: <u>Q2425</u>							
Contract: <u>POWE02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2425</u>	SAS No.: <u>Q2425</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Vanadium	6.26	+/-20	U	40.0	P	06/27/2025	16:09	LB136319
	Zinc	3.50	+/-20	U	40.0	P	06/27/2025	16:09	LB136319
CCB05	Aluminum	11.3	+/-50	U	100	P	06/27/2025	17:01	LB136319
	Antimony	6.76	+/-25	U	50.0	P	06/27/2025	17:01	LB136319
	Arsenic	5.12	+/-10	U	20.0	P	06/27/2025	17:01	LB136319
	Barium	14.6	+/-50	U	100	P	06/27/2025	17:01	LB136319
	Beryllium	0.56	+/-3	U	6.00	P	06/27/2025	17:01	LB136319
	Cadmium	0.50	+/-3	U	6.00	P	06/27/2025	17:01	LB136319
	Chromium	2.12	+/-5	U	10.0	P	06/27/2025	17:01	LB136319
	Cobalt	2.26	+/-15	U	30.0	P	06/27/2025	17:01	LB136319
	Copper	4.60	+/-10	U	20.0	P	06/27/2025	17:01	LB136319
	Iron	23.4	+/-50	U	100	P	06/27/2025	17:01	LB136319
	Lead	2.30	+/-6	U	12.0	P	06/27/2025	17:01	LB136319
	Manganese	5.94	+/-10	U	20.0	P	06/27/2025	17:01	LB136319
	Nickel	3.06	+/-20	U	40.0	P	06/27/2025	17:01	LB136319
	Potassium	918	+/-1000	U	2000	P	06/27/2025	17:01	LB136319
	Selenium	9.64	+/-10	U	20.0	P	06/27/2025	17:01	LB136319
	Silver	1.62	+/-5	U	10.0	P	06/27/2025	17:01	LB136319
	Sodium	868	+/-1000	U	2000	P	06/27/2025	17:01	LB136319
	Thallium	4.38	+/-20	U	40.0	P	06/27/2025	17:01	LB136319
CCB06	Vanadium	6.26	+/-20	U	40.0	P	06/27/2025	17:01	LB136319
	Zinc	3.50	+/-20	U	40.0	P	06/27/2025	17:01	LB136319
	Aluminum	11.3	+/-50	U	100	P	06/27/2025	17:53	LB136319
	Antimony	6.76	+/-25	U	50.0	P	06/27/2025	17:53	LB136319
	Arsenic	5.12	+/-10	U	20.0	P	06/27/2025	17:53	LB136319
	Barium	14.6	+/-50	U	100	P	06/27/2025	17:53	LB136319
	Beryllium	0.56	+/-3	U	6.00	P	06/27/2025	17:53	LB136319
	Cadmium	0.50	+/-3	U	6.00	P	06/27/2025	17:53	LB136319
	Chromium	2.12	+/-5	U	10.0	P	06/27/2025	17:53	LB136319
	Cobalt	2.26	+/-15	U	30.0	P	06/27/2025	17:53	LB136319
	Copper	4.60	+/-10	U	20.0	P	06/27/2025	17:53	LB136319
	Iron	23.4	+/-50	U	100	P	06/27/2025	17:53	LB136319
	Lead	2.30	+/-6	U	12.0	P	06/27/2025	17:53	LB136319
	Manganese	5.94	+/-10	U	20.0	P	06/27/2025	17:53	LB136319
	Nickel	3.06	+/-20	U	40.0	P	06/27/2025	17:53	LB136319
	Potassium	918	+/-1000	U	2000	P	06/27/2025	17:53	LB136319
	Selenium	9.64	+/-10	U	20.0	P	06/27/2025	17:53	LB136319
	Silver	1.62	+/-5	U	10.0	P	06/27/2025	17:53	LB136319
	Sodium	868	+/-1000	U	2000	P	06/27/2025	17:53	LB136319

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Kleinfelder</u>		SDG No.: <u>Q2425</u>							
Contract: <u>POWE02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2425</u>	SAS No.: <u>Q2425</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Thallium	4.38	+/-20	U	40.0	P	06/27/2025	17:53	LB136319
	Vanadium	6.26	+/-20	U	40.0	P	06/27/2025	17:53	LB136319
	Zinc	3.50	+/-20	U	40.0	P	06/27/2025	17:53	LB136319
CCB07	Aluminum	11.3	+/-50	U	100	P	06/27/2025	18:46	LB136319
	Antimony	6.76	+/-25	U	50.0	P	06/27/2025	18:46	LB136319
	Arsenic	5.12	+/-10	U	20.0	P	06/27/2025	18:46	LB136319
	Barium	14.6	+/-50	U	100	P	06/27/2025	18:46	LB136319
	Beryllium	0.56	+/-3	U	6.00	P	06/27/2025	18:46	LB136319
	Cadmium	0.50	+/-3	U	6.00	P	06/27/2025	18:46	LB136319
	Chromium	2.12	+/-5	U	10.0	P	06/27/2025	18:46	LB136319
	Cobalt	2.26	+/-15	U	30.0	P	06/27/2025	18:46	LB136319
	Copper	4.60	+/-10	U	20.0	P	06/27/2025	18:46	LB136319
	Iron	23.4	+/-50	U	100	P	06/27/2025	18:46	LB136319
	Lead	2.30	+/-6	U	12.0	P	06/27/2025	18:46	LB136319
	Manganese	5.94	+/-10	U	20.0	P	06/27/2025	18:46	LB136319
	Nickel	3.06	+/-20	U	40.0	P	06/27/2025	18:46	LB136319
	Potassium	918	+/-1000	U	2000	P	06/27/2025	18:46	LB136319
	Selenium	9.64	+/-10	U	20.0	P	06/27/2025	18:46	LB136319
	Silver	1.62	+/-5	U	10.0	P	06/27/2025	18:46	LB136319
	Sodium	868	+/-1000	U	2000	P	06/27/2025	18:46	LB136319
	Thallium	4.38	+/-20	U	40.0	P	06/27/2025	18:46	LB136319
	Vanadium	6.26	+/-20	U	40.0	P	06/27/2025	18:46	LB136319
	Zinc	3.50	+/-20	U	40.0	P	06/27/2025	18:46	LB136319
CCB08	Aluminum	11.3	+/-50	U	100	P	06/27/2025	19:33	LB136319
	Antimony	6.76	+/-25	U	50.0	P	06/27/2025	19:33	LB136319
	Arsenic	5.12	+/-10	U	20.0	P	06/27/2025	19:33	LB136319
	Barium	14.6	+/-50	U	100	P	06/27/2025	19:33	LB136319
	Beryllium	0.56	+/-3	U	6.00	P	06/27/2025	19:33	LB136319
	Cadmium	0.50	+/-3	U	6.00	P	06/27/2025	19:33	LB136319
	Chromium	2.12	+/-5	U	10.0	P	06/27/2025	19:33	LB136319
	Cobalt	2.26	+/-15	U	30.0	P	06/27/2025	19:33	LB136319
	Copper	4.60	+/-10	U	20.0	P	06/27/2025	19:33	LB136319
	Iron	23.4	+/-50	U	100	P	06/27/2025	19:33	LB136319
	Lead	2.30	+/-6	U	12.0	P	06/27/2025	19:33	LB136319
	Manganese	5.94	+/-10	U	20.0	P	06/27/2025	19:33	LB136319
	Nickel	3.06	+/-20	U	40.0	P	06/27/2025	19:33	LB136319
	Potassium	918	+/-1000	U	2000	P	06/27/2025	19:33	LB136319
	Selenium	9.64	+/-10	U	20.0	P	06/27/2025	19:33	LB136319
	Silver	1.62	+/-5	U	10.0	P	06/27/2025	19:33	LB136319

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder **SDG No.:** Q2425
Contract: POWE02 **Lab Code:** CHEM **Case No.:** Q2425 **SAS No.:** Q2425

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Sodium	868	+/-1000	U	2000	P	06/27/2025	19:33	LB136319
	Thallium	4.38	+/-20	U	40.0	P	06/27/2025	19:33	LB136319
	Vanadium	6.26	+/-20	U	40.0	P	06/27/2025	19:33	LB136319
	Zinc	3.50	+/-20	U	40.0	P	06/27/2025	19:33	LB136319

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q2425

Instrument: CV1

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB168633BL		SOLID		Batch Number:	PB168633		Prep Date:	06/26/2025	
	Mercury	0.0080	<0.013	U	0.013	CV	06/26/2025	14:51	lb136298

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q2425

Instrument: P5

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB168628BL	SOLID			Batch Number:	PB168628		Prep Date:	06/26/2025	
	Aluminum	0.84	<2.5	U	5.00	P	06/27/2025	17:45	LB136319
	Antimony	0.22	<1.25	U	2.50	P	06/27/2025	17:45	LB136319
	Arsenic	0.19	<0.5	U	1.00	P	06/27/2025	17:45	LB136319
	Barium	0.73	<2.5	U	5.00	P	06/27/2025	17:45	LB136319
	Beryllium	0.025	<0.15	U	0.30	P	06/27/2025	17:45	LB136319
	Cadmium	0.024	<0.15	U	0.30	P	06/27/2025	17:45	LB136319
	Chromium	0.047	<0.25	U	0.50	P	06/27/2025	17:45	LB136319
	Cobalt	0.10	<0.75	U	1.50	P	06/27/2025	17:45	LB136319
	Copper	0.22	<0.5	U	1.00	P	06/27/2025	17:45	LB136319
	Iron	3.99	<2.5	U	5.00	P	06/27/2025	17:45	LB136319
	Lead	0.13	<0.3	U	0.60	P	06/27/2025	17:45	LB136319
	Manganese	0.14	<0.5	U	1.00	P	06/27/2025	17:45	LB136319
	Nickel	0.13	<1	U	2.00	P	06/27/2025	17:45	LB136319
	Potassium	27.7	<50	U	100	P	06/27/2025	17:45	LB136319
	Selenium	0.26	<0.5	U	1.00	P	06/27/2025	17:45	LB136319
	Silver	0.12	<0.25	U	0.50	P	06/27/2025	17:45	LB136319
	Sodium	17.8	<50	U	100	P	06/27/2025	17:45	LB136319
	Thallium	0.23	<1	U	2.00	P	06/27/2025	17:45	LB136319
	Vanadium	0.28	<1	J	2.00	P	06/27/2025	17:45	LB136319
	Zinc	0.11	<1	U	2.00	P	06/27/2025	17:45	LB136319

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Kleinfelder **SDG No.:** Q2425
Contract: POWE02 **Lab Code:** CHEM **Case No.:** Q2425 **SAS No.:** Q2425
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	240000	250000	96	216000	294000	06/27/2025	13:04	LB136319
	Antimony	21.2			-50	50	06/27/2025	13:04	LB136319
	Arsenic	1.80			-20	20	06/27/2025	13:04	LB136319
	Barium	5.72	6.0	95	-94	106	06/27/2025	13:04	LB136319
	Beryllium	1.16			-6	6	06/27/2025	13:04	LB136319
	Cadmium	-1.37	1.0	137	-5	7	06/27/2025	13:04	LB136319
	Chromium	49.2	52.0	95	42	62	06/27/2025	13:04	LB136319
	Cobalt	29.0			-30	30	06/27/2025	13:04	LB136319
	Copper	-5.04	2.0	252	-18	22	06/27/2025	13:04	LB136319
	Iron	101000	100000	101	85600	116500	06/27/2025	13:04	LB136319
	Lead	-0.17			-12	12	06/27/2025	13:04	LB136319
	Manganese	5.41	7.0	77	-13	27	06/27/2025	13:04	LB136319
	Nickel	6.13	2.0	306	-38	42	06/27/2025	13:04	LB136319
	Potassium	10.4			0	0	06/27/2025	13:04	LB136319
	Selenium	4.19			-20	20	06/27/2025	13:04	LB136319
	Silver	4.58			-10	10	06/27/2025	13:04	LB136319
	Sodium	137			0	0	06/27/2025	13:04	LB136319
	Thallium	17.3			-40	40	06/27/2025	13:04	LB136319
	Vanadium	5.71			-40	40	06/27/2025	13:04	LB136319
	Zinc	3.40			-40	40	06/27/2025	13:04	LB136319
ICSAB01	Aluminum	236000	250000	94	209000	285000	06/27/2025	13:09	LB136319
	Antimony	633	620	102	525	711	06/27/2025	13:09	LB136319
	Arsenic	100	100	100	88.4	120	06/27/2025	13:09	LB136319
	Barium	518	540	96	437	637	06/27/2025	13:09	LB136319
	Beryllium	504	500	101	420	570	06/27/2025	13:09	LB136319
	Cadmium	994	970	102	826	1120	06/27/2025	13:09	LB136319
	Chromium	521	540	96	460	624	06/27/2025	13:09	LB136319
	Cobalt	519	480	108	404	548	06/27/2025	13:09	LB136319
	Copper	479	510	94	434	588	06/27/2025	13:09	LB136319
	Iron	99300	99000	100	84400	114500	06/27/2025	13:09	LB136319
	Lead	47.5	49.0	97	37	61	06/27/2025	13:09	LB136319
	Manganese	502	510	98	430	584	06/27/2025	13:09	LB136319
	Nickel	975	950	103	810	1100	06/27/2025	13:09	LB136319
	Potassium	6.48			0	0	06/27/2025	13:09	LB136319
	Selenium	55.8	46.0	121	26	66	06/27/2025	13:09	LB136319
	Silver	195	200	98	170	232	06/27/2025	13:09	LB136319
	Sodium	135			0	0	06/27/2025	13:09	LB136319
	Thallium	102	110	93	68	148	06/27/2025	13:09	LB136319
	Vanadium	487	490	99	417	565	06/27/2025	13:09	LB136319
	Zinc	1030	950	108	809	1095	06/27/2025	13:09	LB136319



METAL QC DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Kleinfelder</u>	level: <u>low</u>	sdg no.: <u>Q2425</u>
contract: <u>POWE02</u>	lab code: <u>CHEM</u>	case no.: <u>Q2425</u> sas no.: <u>Q2425</u>
matrix: <u>Solid</u>	sample id: <u>Q2425-03</u>	client id: <u>COMP-3MS</u>
Percent Solids for Sample: 82.6	Spiked ID: Q2425-03MS	Percent Solids for Spike Sample: 82.6

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	mg/Kg	75 - 125	13400		11400		110	1972		P
Antimony	mg/Kg	75 - 125	17.5		1.43	J	42.1	38	N	P
Arsenic	mg/Kg	75 - 125	49.4		7.98		42.1	98		P
Barium	mg/Kg	75 - 125	57.9		46.5		10.5	109		P
Beryllium	mg/Kg	75 - 125	9.94		0.86		10.5	86		P
Cadmium	mg/Kg	75 - 125	10.2		0.31	U	10.5	97		P
Chromium	mg/Kg	75 - 125	37.4		17.8		21.1	93		P
Cobalt	mg/Kg	75 - 125	18.7		7.69		10.5	104		P
Copper	mg/Kg	75 - 125	22.2		7.82		15.8	91		P
Iron	mg/Kg	75 - 125	21600		19100		160	1569		P
Lead	mg/Kg	75 - 125	65.8		15.8		52.6	95		P
Manganese	mg/Kg	75 - 125	178		180		10.5	-15		P
Mercury	mg/Kg	80 - 120	0.42		0.10		0.3	107		CV
Nickel	mg/Kg	75 - 125	39.7		12.2		26.3	104		P
Potassium	mg/Kg	75 - 125	1400		760		530	122		P
Selenium	mg/Kg	75 - 125	86.2		5.97		110	76		P
Silver	mg/Kg	75 - 125	3.45		0.18	J	3.9	83		P
Sodium	mg/Kg	75 - 125	266		132		160	84		P
Thallium	mg/Kg	75 - 125	101		2.03	U	110	96		P
Vanadium	mg/Kg	75 - 125	46.5		29.2		15.8	109		P
Zinc	mg/Kg	75 - 125	39.8		27.8		10.5	114		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Kleinfelder</u>	level: <u>low</u>	sdg no.: <u>Q2425</u>
contract: <u>POWE02</u>	lab code: <u>CHEM</u>	case no.: <u>Q2425</u> sas no.: <u>Q2425</u>
matrix: <u>Solid</u>	sample id: <u>Q2425-03</u>	client id: <u>COMP-3MSD</u>
Percent Solids for Sample: 82.6	Spiked ID: Q2425-03MSD	Percent Solids for Spike Sample: 82.6

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	mg/Kg	75 - 125	14800		11400		110	3258		P
Antimony	mg/Kg	75 - 125	18.4		1.43	J	42.5	40	N	P
Arsenic	mg/Kg	75 - 125	52.6		7.98		42.5	105		P
Barium	mg/Kg	75 - 125	58.6		46.5		10.6	114		P
Beryllium	mg/Kg	75 - 125	10.7		0.86		10.6	93		P
Cadmium	mg/Kg	75 - 125	10.7		0.31	U	10.6	101		P
Chromium	mg/Kg	75 - 125	41.1		17.8		21.2	110		P
Cobalt	mg/Kg	75 - 125	19.5		7.69		10.6	112		P
Copper	mg/Kg	75 - 125	24.8		7.82		15.9	106		P
Iron	mg/Kg	75 - 125	24100		19100		160	3127		P
Lead	mg/Kg	75 - 125	71.0		15.8		53.1	104		P
Manganese	mg/Kg	75 - 125	162		180		10.6	-166		P
Mercury	mg/Kg	80 - 120	0.45		0.10		0.32	109		CV
Nickel	mg/Kg	75 - 125	42.0		12.2		26.5	112		P
Potassium	mg/Kg	75 - 125	1560		760		530	150	N	P
Selenium	mg/Kg	75 - 125	92.8		5.97		110	82		P
Silver	mg/Kg	75 - 125	3.51		0.18	J	4.0	84		P
Sodium	mg/Kg	75 - 125	285		132		160	96		P
Thallium	mg/Kg	75 - 125	107		2.03	U	110	101		P
Vanadium	mg/Kg	75 - 125	50.3		29.2		15.9	132	N	P
Zinc	mg/Kg	75 - 125	43.5		27.8		10.6	148	N	P

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: <u>Kleinfelder</u>	SDG No.: <u>Q2425</u>	
Contract: <u>POWE02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2425</u> SAS No.: <u>Q2425</u>
Matrix: <u>Solid</u>	Level: <u>LOW</u>	Client ID: <u>COMP-3A</u>
Sample ID: <u>Q2425-03</u>	Spiked ID: <u>Q2425-03A</u>	

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Antimony	mg/Kg	75 - 125	32.2		1.43	J	40.7	76		P
Potassium	mg/Kg	75 - 125	1210		760		510	87		P
Vanadium	mg/Kg	75 - 125	39.9		29.2		15.3	70	N	P
Zinc	mg/Kg	75 - 125	35.2		27.8		10.2	73	N	P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Kleinfelder **Level:** LOW **SDG No.:** Q2425
Contract: POWE02 **Lab Code:** CHEM **Case No.:** Q2425 **SAS No.:** Q2425
Matrix: Solid **Sample ID:** Q2425-03 **Client ID:** COMP-3DUP
Percent Solids for Sample: 82.6 **Duplicate ID** Q2425-03DUP **Percent Solids for Spike Sample:** 82.6

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	mg/Kg	20	11400		12300		8		P
Antimony	mg/Kg	20	1.43	J	1.72	J	18		P
Arsenic	mg/Kg	20	7.98		8.70		9		P
Barium	mg/Kg	20	46.5		47.7		3		P
Beryllium	mg/Kg	20	0.86		0.89		3		P
Cadmium	mg/Kg	20	0.31	U	0.34	U			P
Chromium	mg/Kg	20	17.8		19.7		10		P
Cobalt	mg/Kg	20	7.69		7.96		3		P
Copper	mg/Kg	20	7.82		8.61		10		P
Iron	mg/Kg	20	19100		21200		10		P
Lead	mg/Kg	20	15.8		17.1		8		P
Manganese	mg/Kg	20	180		162		11		P
Mercury	mg/Kg	20	0.10		0.10		4		CV
Nickel	mg/Kg	20	12.2		12.7		4		P
Potassium	mg/Kg	20	760		833		9		P
Selenium	mg/Kg	20	5.97		6.45		8		P
Silver	mg/Kg	20	0.18	J	0.17	J	2		P
Sodium	mg/Kg	20	132		138		4		P
Thallium	mg/Kg	20	2.03	U	2.25	U			P
Vanadium	mg/Kg	20	29.2		32.3		10		P
Zinc	mg/Kg	20	27.8		29.9		7		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Kleinfelder **Level:** LOW **SDG No.:** Q2425
Contract: POWE02 **Lab Code:** CHEM **Case No.:** Q2425 **SAS No.:** Q2425
Matrix: Solid **Sample ID:** Q2425-03MS **Client ID:** COMP-3MSD
Percent Solids for Sample: 82.6 **Duplicate ID** Q2425-03MSD **Percent Solids for Spike Sample:** 82.6

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	mg/Kg	20	13400		14800		10		P
Antimony	mg/Kg	20	17.5		18.4		5		P
Arsenic	mg/Kg	20	49.4		52.6		6		P
Barium	mg/Kg	20	57.9		58.6		1		P
Beryllium	mg/Kg	20	9.94		10.7		7		P
Cadmium	mg/Kg	20	10.2		10.7		5		P
Chromium	mg/Kg	20	37.4		41.1		9		P
Cobalt	mg/Kg	20	18.7		19.5		4		P
Copper	mg/Kg	20	22.2		24.8		11		P
Iron	mg/Kg	20	21600		24100		11		P
Lead	mg/Kg	20	65.8		71.0		8		P
Manganese	mg/Kg	20	178		162		9		P
Mercury	mg/Kg	20	0.42		0.45		6		CV
Nickel	mg/Kg	20	39.7		42.0		6		P
Potassium	mg/Kg	20	1400		1560		11		P
Selenium	mg/Kg	20	86.2		92.8		7		P
Silver	mg/Kg	20	3.45		3.51		2		P
Sodium	mg/Kg	20	266		285		7		P
Thallium	mg/Kg	20	101		107		6		P
Vanadium	mg/Kg	20	46.5		50.3		8		P
Zinc	mg/Kg	20	39.8		43.5		9		P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Kleinfelder **SDG No.:** Q2425
Contract: POWE02 **Lab Code:** CHEM **Case No.:** Q2425 **SAS No.:** Q2425

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168628BS							
Aluminum	mg/Kg	100	92.8		93	80 - 120	P
Antimony	mg/Kg	40.0	39.5		99	80 - 120	P
Arsenic	mg/Kg	40.0	34.6		86	80 - 120	P
Barium	mg/Kg	10.0	10.1		101	80 - 120	P
Beryllium	mg/Kg	10.0	10.1		101	80 - 120	P
Cadmium	mg/Kg	10.0	9.23		92	80 - 120	P
Chromium	mg/Kg	20.0	18.4		92	80 - 120	P
Cobalt	mg/Kg	10.0	9.42		94	80 - 120	P
Copper	mg/Kg	15.0	15.8		105	80 - 120	P
Iron	mg/Kg	150	153		102	80 - 120	P
Lead	mg/Kg	50.0	46.6		93	80 - 120	P
Manganese	mg/Kg	10.0	10.1		100	80 - 120	P
Nickel	mg/Kg	25.0	23.6		94	80 - 120	P
Potassium	mg/Kg	500	441		88	80 - 120	P
Selenium	mg/Kg	100	96.9		97	80 - 120	P
Silver	mg/Kg	3.8	3.45		91	80 - 120	P
Sodium	mg/Kg	150	147		98	80 - 120	P
Thallium	mg/Kg	100	98.4		98	80 - 120	P
Vanadium	mg/Kg	15.0	14.8		99	80 - 120	P
Zinc	mg/Kg	10.0	9.81		98	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Kleinfelder **SDG No.:** Q2425
Contract: POWE02 **Lab Code:** CHEM **Case No.:** Q2425 **SAS No.:** Q2425

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168633BS Mercury	mg/Kg	0.26	0.25		95	80 - 120	CV

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

COMP-3L

Lab Name: Chemtech Consulting Group

Contract: POWE02

Lab Code: CHEM Lb No.: lb136319

Lab Sample ID : Q2425-03L

SDG No.: Q2425

Matrix (soil/water): Solid

Level (low/med): LOW

Concentration Units: mg/Kg

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Differ-ence	Q	M
Aluminum	11400		13700		20		P
Antimony	1.43	J	1.96	J	37		P
Arsenic	7.98		8.15		2		P
Barium	46.5		57.5		24		P
Beryllium	0.86		1.06	J	23		P
Cadmium	0.31	U	1.53	U			P
Chromium	17.8		21.3		20		P
Cobalt	7.69		7.88		3		P
Copper	7.82		9.83		26		P
Iron	19100		23800		25		P
Lead	15.8		16.6		5		P
Manganese	180		223		24		P
Mercury	0.10		0.090		9		CV
Nickel	12.2		12.0		2		P
Potassium	760		763		0		P
Selenium	5.97		6.78		14		P
Silver	0.18	J	2.54	U	100.0		P
Sodium	132		159	J	20		P
Thallium	2.03	U	10.2	U			P
Vanadium	29.2		34.5		18		P
Zinc	27.8		32.2		16		P



METAL PREPARATION & INSTRUMENT DATA

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Kleinfelder

SDG No.: Q2425

Contract: POWE02

Lab Code: CHEM

Case No.: Q2425

SAS No.: Q2425

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Al	Ca	Fe	Mg	Ag
Aluminum	396.100	0.0000000	-0.0002060	0.0000000	0.0000000	0.0000000
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	-0.0000440	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000930	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0007850	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	-0.0000920	0.0000000	0.0000380	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	-0.0001440	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	-0.0001490	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0001050	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Kleinfelder

SDG No.: Q2425

Contract: POWE02

Lab Code: CHEM

Case No.: Q2425

SAS No.: Q2425

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		As	Ba	Be	Cd	Co
Aluminum	396.100	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0002870
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0000000	0.0009530
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	-0.0039600
Lead	220.353	0.0000000	0.0003170	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	-0.0003570
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0054900
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Kleinfelder

SDG No.: Q2425

Contract: POWE02

Lab Code: CHEM

Case No.: Q2425

SAS No.: Q2425

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Cr	Cu	K	Mn	Mo
Aluminum	396.100	0.0000000	0.0000000	0.0000590	0.0000000	0.0396900
Antimony	206.833	0.0122000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	-0.0029000	0.0000000	0.0000000	0.0000000	0.0004900
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	-0.0000710	-0.0003400
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000070	0.0002200	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	-0.0007860
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0006510	0.0020500
Iron	240.488	0.0000000	0.0000000	0.0000730	0.0000000	-0.0015250
Lead	220.353	0.0000000	0.0000000	0.0000000	0.0001400	-0.0008600
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0007460	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	-0.0000120
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0017400	-0.0100400
Vanadium	292.402	-0.0025100	0.0000000	0.0000000	0.0000000	-0.0072000
Zinc	213.800	0.0000000	0.0009010	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Kleinfelder

SDG No.: Q2425

Contract: POWE02

Lab Code: CHEM

Case No.: Q2425

SAS No.: Q2425

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Na	Ni	Pb	Sb	Se
Aluminum	396.100	0.0000000	0.0000000	0.0012800	0.0000000	0.0000000
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	-0.0047000	0.0036100	0.0000000	0.0000000
Iron	240.488	0.0000000	-0.0017000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	0.0006580	0.0000000	0.0000000	0.0001290
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0003330	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0067600	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Kleinfelder

SDG No.: Q2425

Contract: POWE02

Lab Code: CHEM

Case No.: Q2425

SAS No.: Q2425

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Sn	Ti	Tl	V	Zn
Aluminum	396.100	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Antimony	206.833	-0.0035600	-0.0007970	0.0000000	-0.0018900	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000630	0.0001280	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0001110	0.0000000
Cobalt	228.616	0.0000000	0.0018800	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0003840	0.0000000	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	-0.0003610	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Silver	328.068	0.0000000	-0.0007420	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	-0.0039700	0.0000000	-0.0115600	0.0000000
Vanadium	292.402	0.0000000	0.0005320	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

METAL PREPARATION & ANALYICAL SUMMARY

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client:	<u>Kleinfelder</u>	SDG No.:	<u>Q2425</u>
Contract:	<u>POWE02</u>	Lab Code:	<u>CHEM</u>
		Method:	<u></u>
		Case No.:	<u>Q2425</u>
		SAS No.:	<u>Q2425</u>

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(g)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB168628							
PB168628BL	PB168628BL	MB	SOLID	06/26/2025	2.00	100.0	100.00
PB168628BS	PB168628BS	LCS	SOLID	06/26/2025	2.00	100.0	100.00
Q2425-01	COMP-1	SAM	SOLID	06/26/2025	2.21	100.0	76.90
Q2425-02	COMP-2	SAM	SOLID	06/26/2025	2.23	100.0	79.70
Q2425-03	COMP-3	SAM	SOLID	06/26/2025	2.38	100.0	82.60
Q2425-03DUP	COMP-3DUP	DUP	SOLID	06/26/2025	2.15	100.0	82.60
Q2425-03MS	COMP-3MS	MS	SOLID	06/26/2025	2.30	100.0	82.60
Q2425-03MSD	COMP-3MSD	MSD	SOLID	06/26/2025	2.28	100.0	82.60

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client: Kleinfelder **SDG No.:** Q2425
Contract: POWE02 **Lab Code:** CHEM **Method:**
Case No.: Q2425 **SAS No.:** Q2425

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(g)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB168633							
PB168633BL	PB168633BL	MB	SOLID	06/26/2025	0.52	35.0	100.00
PB168633BS	PB168633BS	LCS	SOLID	06/26/2025	0.54	35.0	100.00
Q2425-01	COMP-1	SAM	SOLID	06/26/2025	0.58	35.0	76.90
Q2425-02	COMP-2	SAM	SOLID	06/26/2025	0.56	35.0	79.70
Q2425-03	COMP-3	SAM	SOLID	06/26/2025	0.55	35.0	82.60
Q2425-03DUP	COMP-3DUP	DUP	SOLID	06/26/2025	0.53	35.0	82.60
Q2425-03MS	COMP-3MS	MS	SOLID	06/26/2025	0.56	35.0	82.60
Q2425-03MSD	COMP-3MSD	MSD	SOLID	06/26/2025	0.53	35.0	82.60

metals
- 14 -
ANALYSIS RUN LOG

Client: Kleinfelder **Contract:** POWE02

Lab code: CHEM **Case no.:** Q2425 **Sas no.:** Q2425 **Sdg no.:** Q2425

Instrument id number: **Method:** **Run number:** lb136298

Start date: 06/26/2025 **End date:** 06/26/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1411	HG
S0.2	S0.2	1	1413	HG
S2.5	S2.5	1	1415	HG
S5	S5	1	1418	HG
S7.5	S7.5	1	1420	HG
S10	S10	1	1431	HG
ICV08	ICV08	1	1434	HG
ICB08	ICB08	1	1437	HG
CCV23	CCV23	1	1439	HG
CCB23	CCB23	1	1441	HG
CRA	CRA	1	1444	HG
PB168633BL	PB168633BL	1	1451	HG
PB168633BS	PB168633BS	1	1458	HG
CCV24	CCV24	1	1512	HG
CCB24	CCB24	1	1514	HG
CCV25	CCV25	1	1540	HG
CCB25	CCB25	1	1543	HG
Q2425-03	COMP-3	1	1545	HG
Q2425-03DUP	COMP-3DUP	1	1547	HG
Q2425-03MS	COMP-3MS	1	1554	HG
Q2425-03MSD	COMP-3MSD	1	1556	HG
Q2425-03L	COMP-3L	5	1559	HG
Q2425-01	COMP-1	10	1606	HG
Q2425-02	COMP-2	10	1611	HG
CCV26	CCV26	1	1620	HG
CCB26	CCB26	1	1622	HG

metals
- 14 -
ANALYSIS RUN LOG

Client: Kleinfelder **Contract:** POWE02

Lab code: CHEM **Case no.:** Q2425 **Sas no.:** Q2425 **Sdg no.:** Q2425

Instrument id number: **Method:** **Run number:** LB136319

Start date: 06/27/2025 **End date:** 06/27/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1111	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S1	S1	1	1116	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S2	S2	1	1120	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S3	S3	1	1124	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S4	S4	1	1129	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S5	S5	1	1133	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICV01	ICV01	1	1145	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
LLICV01	LLICV01	1	1203	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICB01	ICB01	1	1254	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CRI01	CRI01	1	1258	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSA01	ICSA01	1	1304	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSAB01	ICSAB01	1	1309	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV01	CCV01	1	1322	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB01	CCB01	1	1331	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV02	CCV02	1	1419	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB02	CCB02	1	1424	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV03	CCV03	1	1512	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB03	CCB03	1	1516	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV04	CCV04	1	1604	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB04	CCB04	1	1609	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV05	CCV05	1	1656	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB05	CCB05	1	1701	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
PB168628BL	PB168628BL	1	1745	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV06	CCV06	1	1749	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB06	CCB06	1	1753	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
PB168628BS	PB168628BS	1	1758	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2425-01	COMP-1	1	1819	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2425-02	COMP-2	1	1824	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV07	CCV07	1	1841	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB07	CCB07	1	1846	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2425-03	COMP-3	1	1903	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2425-03DUP	COMP-3DUP	1	1908	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2425-03L	COMP-3L	5	1912	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2425-03MS	COMP-3MS	1	1916	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2425-03MSD	COMP-3MSD	1	1920	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2425-03A	COMP-3A	1	1925	K,Sb,V,Zn
CCV08	CCV08	1	1929	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB08	CCB08	1	1933	Ag,Al,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn

LAB CHRONICLE

OrderID:	Q2425	OrderDate:	6/25/2025 2:03:00 PM
Client:	Kleinfelder	Project:	AS Jenks School
Contact:	Mark Warchol	Location:	D51,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2425-01	COMP-1	SOIL			06/24/25 09:05			06/25/25
			Ammonia	SM4500-NH3		06/26/25	06/27/25 11:35	
			Anions Group1	9056A			06/26/25 10:27	
			Hexavalent Chromium	7196A		06/26/25	06/26/25 12:17	
			Trivalent Chromium	6010D			06/27/25 18:19	
Q2425-02	COMP-2	SOIL			06/24/25 09:30			06/25/25
			Ammonia	SM4500-NH3		06/26/25	06/27/25 11:43	
			Anions Group1	9056A			06/26/25 11:32	
			Hexavalent Chromium	7196A		06/26/25	06/26/25 12:18	
			Trivalent Chromium	6010D			06/27/25 18:24	
Q2425-03	COMP-3	SOIL			06/24/25 09:45			06/25/25
			Ammonia	SM4500-NH3		06/26/25	06/27/25 11:43	
			Anions Group1	9056A			06/26/25 11:53	
			Hexavalent Chromium	7196A		06/26/25	06/26/25 12:19	
			Trivalent Chromium	6010D			06/27/25 19:03	



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25 09:05
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-1	SDG No.:	Q2425
Lab Sample ID:	Q2425-01	Matrix:	SOIL
		% Solid:	76.9

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units(Dry Weight)	Prep Date	Date Ana.	Ana Met.
Ammonia as N	2.80	U	1	2.80	6.30	mg/Kg	06/26/25 10:45	06/27/25 11:35	SM 4500-NH3 B plus G-21
Chloride	6.20	J	1	4.50	15.5	mg/Kg		06/26/25 10:27	9056A
Fluoride	11.5		1	2.30	10.4	mg/Kg		06/26/25 10:27	9056A
Sulfate	19.5	J	1	11.4	77.7	mg/Kg		06/26/25 10:27	9056A
Hexavalent Chromium	0.090	U	1	0.090	0.52	mg/Kg	06/26/25 08:35	06/26/25 12:17	7196A
Trivalent Chromium	12.1		1	0.65	0.65	mg/Kg		06/27/25 18:19	6010D

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25 09:30
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-2	SDG No.:	Q2425
Lab Sample ID:	Q2425-02	Matrix:	SOIL
		% Solid:	79.7

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units(Dry Weight)	Prep Date	Date Ana.	Ana Met.
Ammonia as N	2.70	U	1	2.70	6.20	mg/Kg	06/26/25 10:45	06/27/25 11:43	SM 4500-NH3 B plus G-21
Chloride	9.40	J	1	4.30	14.9	mg/Kg		06/26/25 11:32	9056A
Fluoride	4.80	J	1	2.20	9.90	mg/Kg		06/26/25 11:32	9056A
Sulfate	21.8	J	1	10.9	74.4	mg/Kg		06/26/25 11:32	9056A
Hexavalent Chromium	0.086	U	1	0.086	0.49	mg/Kg	06/26/25 08:35	06/26/25 12:18	7196A
Trivalent Chromium	8.04		1	0.63	0.63	mg/Kg		06/27/25 18:24	6010D

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25 09:45
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-3	SDG No.:	Q2425
Lab Sample ID:	Q2425-03	Matrix:	SOIL
		% Solid:	82.6

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units(Dry Weight)	Prep Date	Date Ana.	Ana Met.
Ammonia as N	2.60	J	1	2.60	5.80	mg/Kg	06/26/25 10:45	06/27/25 11:43	SM 4500-NH3 B plus G-21
Chloride	36.6		1	4.20	14.5	mg/Kg		06/26/25 11:53	9056A
Fluoride	2.90	J	1	2.10	9.70	mg/Kg		06/26/25 11:53	9056A
Sulfate	17.6	J	1	10.6	72.6	mg/Kg		06/26/25 11:53	9056A
Hexavalent Chromium	0.083	U	1	0.083	0.47	mg/Kg	06/26/25 08:35	06/26/25 12:19	7196A
Trivalent Chromium	17.8		1	0.61	0.61	mg/Kg		06/27/25 19:03	6010D

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements
H = Sample Analysis Out Of Hold Time

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits

QC RESULT SUMMARY

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q2425

Project: AS Jenks School

RunNo.: LB136282

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	9.7	10	97	90-110	06/23/2025
Chloride	mg/L	2.9	3	97	90-110	06/23/2025
Fluoride	mg/L	2	2	100	90-110	06/23/2025
Nitrite	mg/L	2.9	3	97	90-110	06/23/2025
Nitrate	mg/L	2.4	2.5	96	90-110	06/23/2025
Sulfate	mg/L	14.3	15	95	90-110	06/23/2025
Orthophosphate as P	mg/L	4.9	5	98	90-110	06/23/2025
Sample ID: CCV1						
Bromide	mg/L	10.1	10	101	90-110	06/26/2025
Chloride	mg/L	3	3	100	90-110	06/26/2025
Fluoride	mg/L	2	2	100	90-110	06/26/2025
Nitrite	mg/L	3	3	100	90-110	06/26/2025
Nitrate	mg/L	2.5	2.5	100	90-110	06/26/2025
Sulfate	mg/L	14.9	15	99	90-110	06/26/2025
Orthophosphate as P	mg/L	5.1	5	102	90-110	06/26/2025
Sample ID: CCV2						
Bromide	mg/L	10.1	10	101	90-110	06/26/2025
Chloride	mg/L	3	3	100	90-110	06/26/2025
Fluoride	mg/L	2.1	2	105	90-110	06/26/2025
Nitrite	mg/L	3	3	100	90-110	06/26/2025
Nitrate	mg/L	2.5	2.5	100	90-110	06/26/2025
Sulfate	mg/L	15	15	100	90-110	06/26/2025
Orthophosphate as P	mg/L	5.1	5	102	90-110	06/26/2025

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q2425

Project: AS Jenks School

RunNo.: LB136293

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.492	0.5	98	90-110	06/26/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.497	0.5	99	90-110	06/26/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.493	0.5	99	90-110	06/26/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.500	0.5	100	90-110	06/26/2025

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q2425

Project: AS Jenks School

RunNo.: LB136312

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.99	1	99	90-110	06/27/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.98	1	98	90-110	06/27/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	06/27/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	06/27/2025

Initial and Continuing Calibration Blank Summary

Client: Kleinfelder

SDG No.: Q2425

Project: AS Jenks School

RunNo.: LB136282

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	06/23/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	06/23/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	06/23/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	06/23/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	06/23/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	06/23/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	06/23/2025
Sample ID: CCB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	06/26/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	06/26/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	06/26/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	06/26/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	06/26/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	06/26/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	06/26/2025
Sample ID: CCB2							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	06/26/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	06/26/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	06/26/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	06/26/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	06/26/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	06/26/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	06/26/2025

Initial and Continuing Calibration Blank Summary

Client: Kleinfelder

SDG No.: Q2425

Project: AS Jenks School

RunNo.: LB136293

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	06/26/2025
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	06/26/2025
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	06/26/2025
Sample ID: CCB3 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	06/26/2025

Initial and Continuing Calibration Blank Summary

Client: Kleinfelder

SDG No.: Q2425

Project: AS Jenks School

RunNo.: LB136312

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	06/27/2025
Sample ID: CCB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	06/27/2025
Sample ID: CCB2 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	06/27/2025
Sample ID: CCB3 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	06/27/2025

Preparation Blank Summary

Client: Kleinfelder

SDG No.: Q2425

Project: AS Jenks School

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB136282BLS							
Bromide	mg/Kg	< 20.0000	20.0000	U	7	40	06/26/2025
Chloride	mg/Kg	< 6.0000	6.0000	U	3.5	12	06/26/2025
Fluoride	mg/Kg	< 4.0000	4.0000	U	1.8	8	06/26/2025
Nitrite	mg/Kg	< 6.0000	6.0000	U	1.5	12	06/26/2025
Nitrate	mg/Kg	< 5.0000	5.0000	U	1.8	10	06/26/2025
Sulfate	mg/Kg	< 30.0000	30.0000	U	8.8	60	06/26/2025
Orthophosphate as P	mg/Kg	< 10.0000	10.0000	U	6.7	20	06/26/2025
Sample ID: PB168617BL							
Hexavalent Chromium	mg/Kg	< 0.2000	0.2000	U	0.07	0.4	06/26/2025
Sample ID: PB168620BL							
Ammonia as N	mg/Kg	< 2.5000	2.5000	U	2.2	5	06/27/2025

Matrix Spike Summary

Client:	Kleinfelder	SDG No.:	Q2425
Project:	AS Jenks School	Sample ID:	Q2416-01
Client ID:	MH-G/HMS	Percent Solids for Spike Sample:	89.7

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	75-125	1420		0.078	U	1430	40	99		06/26/2025

Matrix Spike Summary

Client:	Kleinfelder	SDG No.:	Q2425
Project:	AS Jenks School	Sample ID:	Q2416-01
Client ID:	MH-G/HMS	Percent Solids for Spike Sample:	89.7

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	85-115	42.7		0.078	U	44.6	2	96		06/26/2025

Matrix Spike Summary

Client:	Kleinfelder	SDG No.:	Q2425
Project:	AS Jenks School	Sample ID:	Q2416-01
Client ID:	MH-G/HMS	Percent Solids for Spike Sample:	89.7

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	75-125	36.9		0.078	U	44.6	2	83		06/26/2025

Matrix Spike Summary

Client:	Kleinfelder	SDG No.:	Q2425
Project:	AS Jenks School	Sample ID:	Q2425-01
Client ID:	COMP-1MS	Percent Solids for Spike Sample:	76.9

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/Kg	75-125	63.5		2.80	U	62.5	1	102		06/27/2025
Bromide	mg/Kg	80-120	254		9.10	U	260	1	98		06/26/2025
Chloride	mg/Kg	80-120	81.6		6.20	J	77.3	1	98		06/26/2025
Fluoride	mg/Kg	80-120	54.5		11.5		51.5	1	83		06/26/2025
Nitrite	mg/Kg	80-120	75.6		2.00	U	77.3	1	98		06/26/2025
Nitrate	mg/Kg	80-120	62.4		2.30	U	64.4	1	97		06/26/2025
Sulfate	mg/Kg	80-120	386		19.5	J	390	1	94		06/26/2025
Orthophosphate as P	mg/Kg	80-120	97.0		9.70	J	130	1	67	*	06/26/2025

Matrix Spike Summary

Client:	Kleinfelder	SDG No.:	Q2425
Project:	AS Jenks School	Sample ID:	Q2425-01
Client ID:	COMP-1MSD	Percent Solids for Spike Sample:	76.9

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/Kg	75-125	62.9		2.80	U	63.1	1	100		06/27/2025
Bromide	mg/Kg	80-120	257		9.10	U	260	1	99		06/26/2025
Chloride	mg/Kg	80-120	81.5		6.20	J	78	1	97		06/26/2025
Fluoride	mg/Kg	80-120	55.0		11.5		52	1	84		06/26/2025
Nitrite	mg/Kg	80-120	76.5		2.00	U	78	1	98		06/26/2025
Nitrate	mg/Kg	80-120	63.0		2.30	U	65	1	97		06/26/2025
Sulfate	mg/Kg	80-120	386		19.5	J	390	1	94		06/26/2025
Orthophosphate as P	mg/Kg	80-120	118		9.70	J	130	1	83		06/26/2025

Duplicate Sample Summary

Client:	Kleinfelder	SDG No.:	Q2425
Project:	AS Jenks School	Sample ID:	Q2416-01
Client ID:	MH-G/HDUP	Percent Solids for Spike Sample:	89.7

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	+/-20	0.078	U	0.078	U	1	0		06/26/2025

Duplicate Sample Summary

Client:	Kleinfelder	SDG No.:	Q2425
Project:	AS Jenks School	Sample ID:	Q2425-01
Client ID:	COMP-1DUP	Percent Solids for Spike Sample:	76.9

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Ammonia as N	mg/Kg	+/-20	2.80	U	2.80	U	1	0		06/27/2025

Duplicate Sample Summary

Client:	Kleinfelder	SDG No.:	Q2425
Project:	AS Jenks School	Sample ID:	Q2425-01
Client ID:	COMP-1MSD	Percent Solids for Spike Sample:	76.9

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Chloride	mg/Kg	+/-15	81.6		81.5		1	0		06/26/2025
Sulfate	mg/Kg	+/-15	386		386		1	0		06/26/2025
Bromide	mg/Kg	+/-15	254		257		1	1		06/26/2025
Fluoride	mg/Kg	+/-15	54.5		55.0		1	1		06/26/2025
Nitrate	mg/Kg	+/-15	62.4		63.0		1	1		06/26/2025
Nitrite	mg/Kg	+/-15	75.6		76.5		1	1		06/26/2025
Orthophosphate as P	mg/Kg	+/-15	97.0		118		1	20		06/26/2025
Ammonia as N	mg/Kg	+/-20	63.5		62.9		1	1		06/27/2025

Laboratory Control Sample Summary

Client: Kleinfelder

SDG No.: Q2425

Project: AS Jenks School

Run No.: LB136282

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136282BSS							
Bromide	mg/Kg	200	203		102	1	90-110	06/26/2025
Chloride	mg/Kg	60	59.8		100	1	90-110	06/26/2025
Fluoride	mg/Kg	40	41.0		102	1	90-110	06/26/2025
Nitrite	mg/Kg	60	60.2		100	1	90-110	06/26/2025
Nitrate	mg/Kg	50	49.0		98	1	90-110	06/26/2025
Sulfate	mg/Kg	300	299		100	1	90-110	06/26/2025
Orthophosphate as P	mg/Kg	100	102		102	1	90-110	06/26/2025

Laboratory Control Sample Summary

Client: Kleinfelder

SDG No.: Q2425

Project: AS Jenks School

Run No.: LB136293

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168617BS							
Hexavalent Chromium	mg/Kg	20	20.1		101	1	84-110	06/26/2025

Laboratory Control Sample Summary

Client: Kleinfelder

SDG No.: Q2425

Project: AS Jenks School

Run No.: LB136312

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168620BS							
Ammonia as N	mg/Kg	50	48.9		98	1	90-110	06/27/2025