

SOP ID : MSM4500-NH3 B,G-Ammonia-18

SDG No : N/A

Start Digest Date: 06/26/2025 Time : 10:45 Temp : 150 °C

Matrix : SOIL

End Digest Date: 06/26/2025 Time : 11:45 Temp : 160 °C

Pipette ID : WC

Balance ID : WC SC-7

Hood ID : HOOD#2

Digestion tube ID : M5595

Block Thermometer ID : WC CYANIDE

Block ID : WC-DIST-BLOCK-1

Filter paper ID : N/A

Prep Technician Signature:

Weigh By : RM

pH Meter ID : N/A

Supervisor Signature:

Standard Name	MLS USED	STD REF. # FROM LOG
LCSS	1.0ML	WP113449
MS/MSD SPIKE SOL.	1.0ML	WP113450
PBS003	50.0ML	W3112
RL CHECK	N/A	AS PER PB168619
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
BORATE BUFFER	2.5ML	WP111325
NAOH 6N	0.5-2.0ML	WP111318
H2SO4 0.04N	5.0ML	WP112828
pH strip-Ammonia	N/A	W3133
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

## Extraction Conformance/Non-Conformance Comments:

ALL GLASSWEAR ARE STEAMED OUT AND THERE WERE NO TRACE OF AMMONIA USING NESLER REAGENT WP111604,

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
06/26/2025 11:55	RM CWG	RM CWG
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Weight (g)	Final Vol (ml)	pH	Sulfide	Oxidizing	Nitrate/ Nitrite	Comment	Prep Pos
PB168620BL	PBS620	1.00	50	N/A	N/A	N/A	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
PB168620BS	LCS620	1.00	50	N/A	N/A	N/A	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2425-01DUP	COMP-1DUP	1.03	50	N/A	N/A	N/A	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2425-01MS	COMP-1MS	1.04	50	N/A	N/A	N/A	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2425-01MSD	COMP-1MSD	1.03	50	N/A	N/A	N/A	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2425-01	COMP-1	1.03	50	N/A	N/A	N/A	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2425-02	COMP-2	1.01	50	N/A	N/A	N/A	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2425-03	COMP-3	1.04	50	N/A	N/A	N/A	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A