

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q2425

Project: AS Jenks School

RunNo.: LB136282

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	9.7	10	97	90-110	06/23/2025
Chloride	mg/L	2.9	3	97	90-110	06/23/2025
Fluoride	mg/L	2	2	100	90-110	06/23/2025
Nitrite	mg/L	2.9	3	97	90-110	06/23/2025
Nitrate	mg/L	2.4	2.5	96	90-110	06/23/2025
Sulfate	mg/L	14.3	15	95	90-110	06/23/2025
Orthophosphate as P	mg/L	4.9	5	98	90-110	06/23/2025
Sample ID: CCV1						
Bromide	mg/L	10.1	10	101	90-110	06/26/2025
Chloride	mg/L	3	3	100	90-110	06/26/2025
Fluoride	mg/L	2	2	100	90-110	06/26/2025
Nitrite	mg/L	3	3	100	90-110	06/26/2025
Nitrate	mg/L	2.5	2.5	100	90-110	06/26/2025
Sulfate	mg/L	14.9	15	99	90-110	06/26/2025
Orthophosphate as P	mg/L	5.1	5	102	90-110	06/26/2025
Sample ID: CCV2						
Bromide	mg/L	10.1	10	101	90-110	06/26/2025
Chloride	mg/L	3	3	100	90-110	06/26/2025
Fluoride	mg/L	2.1	2	105	90-110	06/26/2025
Nitrite	mg/L	3	3	100	90-110	06/26/2025
Nitrate	mg/L	2.5	2.5	100	90-110	06/26/2025
Sulfate	mg/L	15	15	100	90-110	06/26/2025
Orthophosphate as P	mg/L	5.1	5	102	90-110	06/26/2025

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q2425

Project: AS Jenks School

RunNo.: LB136293

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.492	0.5	98	90-110	06/26/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.497	0.5	99	90-110	06/26/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.493	0.5	99	90-110	06/26/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.500	0.5	100	90-110	06/26/2025

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q2425

Project: AS Jenks School

RunNo.: LB136312

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.99	1	99	90-110	06/27/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.98	1	98	90-110	06/27/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	06/27/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	06/27/2025