

Duplicate Sample Summary

Client:	Kleinfelder	SDG No.:	Q2425
Project:	AS Jenks School	Sample ID:	Q2416-01
Client ID:	MH-G/HDUP	Percent Solids for Spike Sample:	89.7

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	+/-20	0.078	U	0.078	U	1	0		06/26/2025

Duplicate Sample Summary

Client: Kleinfelder Project: AS Jenks School Client ID: COMP-1DUP	SDG No.: Q2425 Sample ID: Q2425-01 Percent Solids for Spike Sample: 76.9
--	---

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/Kg	+/-20	2.80	U	2.80	U	1	0		06/27/2025

Duplicate Sample Summary

Client: Kleinfelder Project: AS Jenks School Client ID: COMP-1MSD	SDG No.: Q2425 Sample ID: Q2425-01 Percent Solids for Spike Sample: 76.9
--	---

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Chloride	mg/Kg	+/-15	81.6		81.5		1	0		06/26/2025
Sulfate	mg/Kg	+/-15	386		386		1	0		06/26/2025
Bromide	mg/Kg	+/-15	254		257		1	1		06/26/2025
Fluoride	mg/Kg	+/-15	54.5		55.0		1	1		06/26/2025
Nitrate	mg/Kg	+/-15	62.4		63.0		1	1		06/26/2025
Nitrite	mg/Kg	+/-15	75.6		76.5		1	1		06/26/2025
Orthophosphate as P	mg/Kg	+/-15	97.0		118		1	20		06/26/2025
Ammonia as N	mg/Kg	+/-20	63.5		62.9		1	1		06/27/2025