

Laboratory Control Sample Summary

Client: Kleinfelder

SDG No.: Q2425

Project: AS Jenks School

Run No.: LB136282

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136282BSS							
Bromide	mg/Kg	200	203		102	1	90-110	06/26/2025
Chloride	mg/Kg	60	59.8		100	1	90-110	06/26/2025
Fluoride	mg/Kg	40	41.0		102	1	90-110	06/26/2025
Nitrite	mg/Kg	60	60.2		100	1	90-110	06/26/2025
Nitrate	mg/Kg	50	49.0		98	1	90-110	06/26/2025
Sulfate	mg/Kg	300	299		100	1	90-110	06/26/2025
Orthophosphate as P	mg/Kg	100	102		102	1	90-110	06/26/2025

Laboratory Control Sample Summary

Client: Kleinfelder

SDG No.: Q2425

Project: AS Jenks School

Run No.: LB136293

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168617BS							
Hexavalent Chromium	mg/Kg	20	20.1		101	1	84-110	06/26/2025

Laboratory Control Sample Summary

Client: Kleinfelder

SDG No.: Q2425

Project: AS Jenks School

Run No.: LB136312

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168620BS							
Ammonia as N	mg/Kg	50	48.9		98	1	90-110	06/27/2025