



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Kleinfelder				SDG No.:	Q2425				
Contract:	POWE02	Lab Code:	CHEM		Case No.:	Q2425	SAS No.:	Q2425		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB08	Mercury	0.076	+/-0.2	U	0.20	CV	06/26/2025	14:37	lb136298	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Kleinfelder</u>		SDG No.: <u>Q2425</u>							
Contract: <u>POWE02</u>		Lab Code: <u>CHEM</u>		Case No.: <u>Q2425</u>		SAS No.: <u>Q2425</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB23	Mercury	0.076	+/-0.2	U	0.20	CV	06/26/2025	14:41	lb136298
CCB24	Mercury	0.076	+/-0.2	U	0.20	CV	06/26/2025	15:14	lb136298
CCB25	Mercury	0.076	+/-0.2	U	0.20	CV	06/26/2025	15:43	lb136298
CCB26	Mercury	0.076	+/-0.2	U	0.20	CV	06/26/2025	16:22	lb136298

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Kleinfelder</u>		SDG No.: <u>Q2425</u>							
Contract: <u>POWE02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2425</u>	SAS No.: <u>Q2425</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	06/27/2025	12:54	LB136319
	Antimony	6.76	+/-25	U	50.0	P	06/27/2025	12:54	LB136319
	Arsenic	5.12	+/-10	U	20.0	P	06/27/2025	12:54	LB136319
	Barium	14.6	+/-50	U	100	P	06/27/2025	12:54	LB136319
	Beryllium	0.56	+/-3	U	6.00	P	06/27/2025	12:54	LB136319
	Cadmium	0.50	+/-3	U	6.00	P	06/27/2025	12:54	LB136319
	Chromium	2.12	+/-5	U	10.0	P	06/27/2025	12:54	LB136319
	Cobalt	2.26	+/-15	U	30.0	P	06/27/2025	12:54	LB136319
	Copper	4.60	+/-10	U	20.0	P	06/27/2025	12:54	LB136319
	Iron	23.4	+/-50	U	100	P	06/27/2025	12:54	LB136319
	Lead	2.30	+/-6	U	12.0	P	06/27/2025	12:54	LB136319
	Manganese	5.94	+/-10	U	20.0	P	06/27/2025	12:54	LB136319
	Nickel	3.06	+/-20	U	40.0	P	06/27/2025	12:54	LB136319
	Potassium	918	+/-1000	U	2000	P	06/27/2025	12:54	LB136319
	Selenium	9.64	+/-10	U	20.0	P	06/27/2025	12:54	LB136319
	Silver	1.62	+/-5	U	10.0	P	06/27/2025	12:54	LB136319
	Sodium	868	+/-1000	U	2000	P	06/27/2025	12:54	LB136319
	Thallium	4.38	+/-20	U	40.0	P	06/27/2025	12:54	LB136319
	Vanadium	6.26	+/-20	U	40.0	P	06/27/2025	12:54	LB136319
	Zinc	3.50	+/-20	U	40.0	P	06/27/2025	12:54	LB136319

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Kleinfelder</u>		SDG No.: <u>Q2425</u>							
Contract: <u>POWE02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2425</u>	SAS No.: <u>Q2425</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	06/27/2025	13:31	LB136319
	Antimony	6.76	+/-25	U	50.0	P	06/27/2025	13:31	LB136319
	Arsenic	5.12	+/-10	U	20.0	P	06/27/2025	13:31	LB136319
	Barium	14.6	+/-50	U	100	P	06/27/2025	13:31	LB136319
	Beryllium	0.56	+/-3	U	6.00	P	06/27/2025	13:31	LB136319
	Cadmium	0.50	+/-3	U	6.00	P	06/27/2025	13:31	LB136319
	Chromium	2.12	+/-5	U	10.0	P	06/27/2025	13:31	LB136319
	Cobalt	2.26	+/-15	U	30.0	P	06/27/2025	13:31	LB136319
	Copper	4.60	+/-10	U	20.0	P	06/27/2025	13:31	LB136319
	Iron	23.4	+/-50	U	100	P	06/27/2025	13:31	LB136319
	Lead	2.30	+/-6	U	12.0	P	06/27/2025	13:31	LB136319
	Manganese	5.94	+/-10	U	20.0	P	06/27/2025	13:31	LB136319
	Nickel	3.06	+/-20	U	40.0	P	06/27/2025	13:31	LB136319
	Potassium	918	+/-1000	U	2000	P	06/27/2025	13:31	LB136319
	Selenium	9.64	+/-10	U	20.0	P	06/27/2025	13:31	LB136319
	Silver	1.62	+/-5	U	10.0	P	06/27/2025	13:31	LB136319
	Sodium	868	+/-1000	U	2000	P	06/27/2025	13:31	LB136319
	Thallium	4.38	+/-20	U	40.0	P	06/27/2025	13:31	LB136319
	Vanadium	6.26	+/-20	U	40.0	P	06/27/2025	13:31	LB136319
	Zinc	3.50	+/-20	U	40.0	P	06/27/2025	13:31	LB136319
CCB02	Aluminum	11.3	+/-50	U	100	P	06/27/2025	14:24	LB136319
	Antimony	6.76	+/-25	U	50.0	P	06/27/2025	14:24	LB136319
	Arsenic	5.12	+/-10	U	20.0	P	06/27/2025	14:24	LB136319
	Barium	14.6	+/-50	U	100	P	06/27/2025	14:24	LB136319
	Beryllium	0.56	+/-3	U	6.00	P	06/27/2025	14:24	LB136319
	Cadmium	0.50	+/-3	U	6.00	P	06/27/2025	14:24	LB136319
	Chromium	2.12	+/-5	U	10.0	P	06/27/2025	14:24	LB136319
	Cobalt	2.26	+/-15	U	30.0	P	06/27/2025	14:24	LB136319
	Copper	4.60	+/-10	U	20.0	P	06/27/2025	14:24	LB136319
	Iron	23.4	+/-50	U	100	P	06/27/2025	14:24	LB136319
	Lead	2.30	+/-6	U	12.0	P	06/27/2025	14:24	LB136319
	Manganese	5.94	+/-10	U	20.0	P	06/27/2025	14:24	LB136319
	Nickel	3.06	+/-20	U	40.0	P	06/27/2025	14:24	LB136319
	Potassium	918	+/-1000	U	2000	P	06/27/2025	14:24	LB136319
	Selenium	9.64	+/-10	U	20.0	P	06/27/2025	14:24	LB136319
	Silver	1.62	+/-5	U	10.0	P	06/27/2025	14:24	LB136319
	Sodium	868	+/-1000	U	2000	P	06/27/2025	14:24	LB136319
	Thallium	4.38	+/-20	U	40.0	P	06/27/2025	14:24	LB136319
	Vanadium	6.26	+/-20	U	40.0	P	06/27/2025	14:24	LB136319

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Kleinfelder</u>		SDG No.: <u>Q2425</u>							
Contract: <u>POWE02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2425</u>	SAS No.: <u>Q2425</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Zinc	3.50	+/-20	U	40.0	P	06/27/2025	14:24	LB136319
CCB03	Aluminum	11.3	+/-50	U	100	P	06/27/2025	15:16	LB136319
	Antimony	6.76	+/-25	U	50.0	P	06/27/2025	15:16	LB136319
	Arsenic	5.12	+/-10	U	20.0	P	06/27/2025	15:16	LB136319
	Barium	14.6	+/-50	U	100	P	06/27/2025	15:16	LB136319
	Beryllium	0.56	+/-3	U	6.00	P	06/27/2025	15:16	LB136319
	Cadmium	0.50	+/-3	U	6.00	P	06/27/2025	15:16	LB136319
	Chromium	2.12	+/-5	U	10.0	P	06/27/2025	15:16	LB136319
	Cobalt	2.26	+/-15	U	30.0	P	06/27/2025	15:16	LB136319
	Copper	4.60	+/-10	U	20.0	P	06/27/2025	15:16	LB136319
	Iron	23.4	+/-50	U	100	P	06/27/2025	15:16	LB136319
	Lead	2.30	+/-6	U	12.0	P	06/27/2025	15:16	LB136319
	Manganese	5.94	+/-10	U	20.0	P	06/27/2025	15:16	LB136319
	Nickel	3.06	+/-20	U	40.0	P	06/27/2025	15:16	LB136319
	Potassium	918	+/-1000	U	2000	P	06/27/2025	15:16	LB136319
	Selenium	9.64	+/-10	U	20.0	P	06/27/2025	15:16	LB136319
	Silver	1.62	+/-5	U	10.0	P	06/27/2025	15:16	LB136319
	Sodium	868	+/-1000	U	2000	P	06/27/2025	15:16	LB136319
	Thallium	4.38	+/-20	U	40.0	P	06/27/2025	15:16	LB136319
	Vanadium	6.26	+/-20	U	40.0	P	06/27/2025	15:16	LB136319
	Zinc	3.50	+/-20	U	40.0	P	06/27/2025	15:16	LB136319
CCB04	Aluminum	11.3	+/-50	U	100	P	06/27/2025	16:09	LB136319
	Antimony	6.76	+/-25	U	50.0	P	06/27/2025	16:09	LB136319
	Arsenic	5.12	+/-10	U	20.0	P	06/27/2025	16:09	LB136319
	Barium	14.6	+/-50	U	100	P	06/27/2025	16:09	LB136319
	Beryllium	0.56	+/-3	U	6.00	P	06/27/2025	16:09	LB136319
	Cadmium	0.50	+/-3	U	6.00	P	06/27/2025	16:09	LB136319
	Chromium	2.12	+/-5	U	10.0	P	06/27/2025	16:09	LB136319
	Cobalt	2.26	+/-15	U	30.0	P	06/27/2025	16:09	LB136319
	Copper	4.60	+/-10	U	20.0	P	06/27/2025	16:09	LB136319
	Iron	23.4	+/-50	U	100	P	06/27/2025	16:09	LB136319
	Lead	2.30	+/-6	U	12.0	P	06/27/2025	16:09	LB136319
	Manganese	5.94	+/-10	U	20.0	P	06/27/2025	16:09	LB136319
	Nickel	3.06	+/-20	U	40.0	P	06/27/2025	16:09	LB136319
	Potassium	918	+/-1000	U	2000	P	06/27/2025	16:09	LB136319
	Selenium	9.64	+/-10	U	20.0	P	06/27/2025	16:09	LB136319
	Silver	1.62	+/-5	U	10.0	P	06/27/2025	16:09	LB136319
	Sodium	868	+/-1000	U	2000	P	06/27/2025	16:09	LB136319
	Thallium	4.38	+/-20	U	40.0	P	06/27/2025	16:09	LB136319

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Kleinfelder</u>		SDG No.: <u>Q2425</u>							
Contract: <u>POWE02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2425</u>	SAS No.: <u>Q2425</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Vanadium	6.26	+/-20	U	40.0	P	06/27/2025	16:09	LB136319
	Zinc	3.50	+/-20	U	40.0	P	06/27/2025	16:09	LB136319
CCB05	Aluminum	11.3	+/-50	U	100	P	06/27/2025	17:01	LB136319
	Antimony	6.76	+/-25	U	50.0	P	06/27/2025	17:01	LB136319
	Arsenic	5.12	+/-10	U	20.0	P	06/27/2025	17:01	LB136319
	Barium	14.6	+/-50	U	100	P	06/27/2025	17:01	LB136319
	Beryllium	0.56	+/-3	U	6.00	P	06/27/2025	17:01	LB136319
	Cadmium	0.50	+/-3	U	6.00	P	06/27/2025	17:01	LB136319
	Chromium	2.12	+/-5	U	10.0	P	06/27/2025	17:01	LB136319
	Cobalt	2.26	+/-15	U	30.0	P	06/27/2025	17:01	LB136319
	Copper	4.60	+/-10	U	20.0	P	06/27/2025	17:01	LB136319
	Iron	23.4	+/-50	U	100	P	06/27/2025	17:01	LB136319
	Lead	2.30	+/-6	U	12.0	P	06/27/2025	17:01	LB136319
	Manganese	5.94	+/-10	U	20.0	P	06/27/2025	17:01	LB136319
	Nickel	3.06	+/-20	U	40.0	P	06/27/2025	17:01	LB136319
	Potassium	918	+/-1000	U	2000	P	06/27/2025	17:01	LB136319
	Selenium	9.64	+/-10	U	20.0	P	06/27/2025	17:01	LB136319
	Silver	1.62	+/-5	U	10.0	P	06/27/2025	17:01	LB136319
	Sodium	868	+/-1000	U	2000	P	06/27/2025	17:01	LB136319
	Thallium	4.38	+/-20	U	40.0	P	06/27/2025	17:01	LB136319
	Vanadium	6.26	+/-20	U	40.0	P	06/27/2025	17:01	LB136319
	Zinc	3.50	+/-20	U	40.0	P	06/27/2025	17:01	LB136319
CCB06	Aluminum	11.3	+/-50	U	100	P	06/27/2025	17:53	LB136319
	Antimony	6.76	+/-25	U	50.0	P	06/27/2025	17:53	LB136319
	Arsenic	5.12	+/-10	U	20.0	P	06/27/2025	17:53	LB136319
	Barium	14.6	+/-50	U	100	P	06/27/2025	17:53	LB136319
	Beryllium	0.56	+/-3	U	6.00	P	06/27/2025	17:53	LB136319
	Cadmium	0.50	+/-3	U	6.00	P	06/27/2025	17:53	LB136319
	Chromium	2.12	+/-5	U	10.0	P	06/27/2025	17:53	LB136319
	Cobalt	2.26	+/-15	U	30.0	P	06/27/2025	17:53	LB136319
	Copper	4.60	+/-10	U	20.0	P	06/27/2025	17:53	LB136319
	Iron	23.4	+/-50	U	100	P	06/27/2025	17:53	LB136319
	Lead	2.30	+/-6	U	12.0	P	06/27/2025	17:53	LB136319
	Manganese	5.94	+/-10	U	20.0	P	06/27/2025	17:53	LB136319
	Nickel	3.06	+/-20	U	40.0	P	06/27/2025	17:53	LB136319
	Potassium	918	+/-1000	U	2000	P	06/27/2025	17:53	LB136319
	Selenium	9.64	+/-10	U	20.0	P	06/27/2025	17:53	LB136319
	Silver	1.62	+/-5	U	10.0	P	06/27/2025	17:53	LB136319
	Sodium	868	+/-1000	U	2000	P	06/27/2025	17:53	LB136319

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Kleinfelder</u>		SDG No.: <u>Q2425</u>							
Contract: <u>POWE02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2425</u>	SAS No.: <u>Q2425</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Thallium	4.38	+/-20	U	40.0	P	06/27/2025	17:53	LB136319
	Vanadium	6.26	+/-20	U	40.0	P	06/27/2025	17:53	LB136319
	Zinc	3.50	+/-20	U	40.0	P	06/27/2025	17:53	LB136319
CCB07	Aluminum	11.3	+/-50	U	100	P	06/27/2025	18:46	LB136319
	Antimony	6.76	+/-25	U	50.0	P	06/27/2025	18:46	LB136319
	Arsenic	5.12	+/-10	U	20.0	P	06/27/2025	18:46	LB136319
	Barium	14.6	+/-50	U	100	P	06/27/2025	18:46	LB136319
	Beryllium	0.56	+/-3	U	6.00	P	06/27/2025	18:46	LB136319
	Cadmium	0.50	+/-3	U	6.00	P	06/27/2025	18:46	LB136319
	Chromium	2.12	+/-5	U	10.0	P	06/27/2025	18:46	LB136319
	Cobalt	2.26	+/-15	U	30.0	P	06/27/2025	18:46	LB136319
	Copper	4.60	+/-10	U	20.0	P	06/27/2025	18:46	LB136319
	Iron	23.4	+/-50	U	100	P	06/27/2025	18:46	LB136319
	Lead	2.30	+/-6	U	12.0	P	06/27/2025	18:46	LB136319
	Manganese	5.94	+/-10	U	20.0	P	06/27/2025	18:46	LB136319
	Nickel	3.06	+/-20	U	40.0	P	06/27/2025	18:46	LB136319
	Potassium	918	+/-1000	U	2000	P	06/27/2025	18:46	LB136319
	Selenium	9.64	+/-10	U	20.0	P	06/27/2025	18:46	LB136319
	Silver	1.62	+/-5	U	10.0	P	06/27/2025	18:46	LB136319
	Sodium	868	+/-1000	U	2000	P	06/27/2025	18:46	LB136319
	Thallium	4.38	+/-20	U	40.0	P	06/27/2025	18:46	LB136319
	Vanadium	6.26	+/-20	U	40.0	P	06/27/2025	18:46	LB136319
	Zinc	3.50	+/-20	U	40.0	P	06/27/2025	18:46	LB136319
CCB08	Aluminum	11.3	+/-50	U	100	P	06/27/2025	19:33	LB136319
	Antimony	6.76	+/-25	U	50.0	P	06/27/2025	19:33	LB136319
	Arsenic	5.12	+/-10	U	20.0	P	06/27/2025	19:33	LB136319
	Barium	14.6	+/-50	U	100	P	06/27/2025	19:33	LB136319
	Beryllium	0.56	+/-3	U	6.00	P	06/27/2025	19:33	LB136319
	Cadmium	0.50	+/-3	U	6.00	P	06/27/2025	19:33	LB136319
	Chromium	2.12	+/-5	U	10.0	P	06/27/2025	19:33	LB136319
	Cobalt	2.26	+/-15	U	30.0	P	06/27/2025	19:33	LB136319
	Copper	4.60	+/-10	U	20.0	P	06/27/2025	19:33	LB136319
	Iron	23.4	+/-50	U	100	P	06/27/2025	19:33	LB136319
	Lead	2.30	+/-6	U	12.0	P	06/27/2025	19:33	LB136319
	Manganese	5.94	+/-10	U	20.0	P	06/27/2025	19:33	LB136319
	Nickel	3.06	+/-20	U	40.0	P	06/27/2025	19:33	LB136319
	Potassium	918	+/-1000	U	2000	P	06/27/2025	19:33	LB136319
	Selenium	9.64	+/-10	U	20.0	P	06/27/2025	19:33	LB136319
	Silver	1.62	+/-5	U	10.0	P	06/27/2025	19:33	LB136319

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Kleinfelder</u>		SDG No.: <u>Q2425</u>							
Contract: <u>POWE02</u>		Lab Code: <u>CHEM</u>		Case No.: <u>Q2425</u>		SAS No.: <u>Q2425</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Sodium	868	+/-1000	U	2000	P	06/27/2025	19:33	LB136319
	Thallium	4.38	+/-20	U	40.0	P	06/27/2025	19:33	LB136319
	Vanadium	6.26	+/-20	U	40.0	P	06/27/2025	19:33	LB136319
	Zinc	3.50	+/-20	U	40.0	P	06/27/2025	19:33	LB136319