

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL01 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PO111858.D Time Analyzed: 10:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.767	4.666	4.866	493.110	500.000	-1.4
Aroclor-1016-2	4.786	4.684	4.884	509.120	500.000	1.8
Aroclor-1016-3	4.842	4.741	4.941	505.140	500.000	1.0
Aroclor-1016-4	4.962	4.861	5.061	503.830	500.000	0.8
Aroclor-1016-5	5.220	5.118	5.318	530.310	500.000	6.1
Aroclor-1260-1	6.258	6.157	6.357	513.670	500.000	2.7
Aroclor-1260-2	6.448	6.346	6.546	553.480	500.000	10.7
Aroclor-1260-3	6.813	6.713	6.913	575.780	500.000	15.2
Aroclor-1260-4	7.074	6.972	7.172	564.630	500.000	12.9
Aroclor-1260-5	7.316	7.216	7.416	545.190	500.000	9.0
Decachlorobiphenyl	8.712	8.612	8.812	48.170	50.000	-3.7
Tetrachloro-m-xylene	3.679	3.577	3.777	53.720	50.000	7.4

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL01 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PO111858.D Time Analyzed: 10:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.747	4.651	4.851	495.750	500.000	-0.9
Aroclor-1016-2	4.766	4.669	4.869	504.470	500.000	0.9
Aroclor-1016-3	4.941	4.844	5.044	501.230	500.000	0.2
Aroclor-1016-4	4.984	4.887	5.087	498.050	500.000	-0.4
Aroclor-1016-5	5.196	5.099	5.299	502.020	500.000	0.4
Aroclor-1260-1	6.226	6.129	6.329	501.900	500.000	0.4
Aroclor-1260-2	6.414	6.316	6.516	526.020	500.000	5.2
Aroclor-1260-3	6.566	6.469	6.669	497.850	500.000	-0.4
Aroclor-1260-4	7.036	6.939	7.139	497.700	500.000	-0.5
Aroclor-1260-5	7.277	7.181	7.381	508.850	500.000	1.8
Decachlorobiphenyl	8.658	8.561	8.761	47.550	50.000	-4.9
Tetrachloro-m-xylene	3.671	3.573	3.773	49.570	50.000	-0.9